


Republic of the Philippines
WEST VISAYAS STATE UNIVERSITY
 La Paz, Iloilo City
 MAIN CAMPUS

REQUEST FOR PRICE QUOTATION

KAGER DENTAL HJ-698A (HAIQIU)	
Upholstery	P.L.U Color
<ul style="list-style-type: none"> • Thermostatic water heating system • Suction flushing and cup filling system • Electric lift dental Chair • Control panel integrated with tool holder (5 positions: 3 handpiece tubings, 1 three-way • Motor Ventilation • Non-function pedal • Chair movement pedal switch on bottom cover • Pure water bottle 600ml • Dental Stool 	

Date: 05/23/22
 Quotation No.: 22-05-0085
 Based on P.R. 22-05-1044/1082
 Enduser: Health Services
 ABC: 393,116.00
 Mode: SVP (Section 53.9)

Please quote your lowest price on the item/s listed below, subject to the General Conditions provided, stating the shortest time of delivery and submit your quotation duly signed by your representative not later than **MAY 30 2022** in a sealed envelope.


JULIUS B. UNДАР
 Chairperson, BAC

NOTE:

- All entries must be written in legible ink and if there are erasures the same should bear the initial of the authorized signatory.
- Delivery period within fifteen (15) calendar days
- Warranty shall be for a period of six (6) months for supplies & materials, one (1) year for equipment, from date of acceptance by the Procuring Entity
- Price Validity shall be for a period of 30 Calendar Days
- Phil-GEPS Registration Certificate or Registration Number shall be attached upon submission of the quotation**
- Bidders shall submit original brochures showing certifications of the product being offered
- Include in your submitted quotation the following documents:
 - Valid Current Mayor's Permit,
 - Omnibus Sworn Statement,
 - BIR Registration and
 - Latest ITR For above 500,000.00

Item No.	Articles/Description	Brand	Model	ABC / Unit Price	Unit	Qty.	Unit Price	Total Cost
Procurement of Various Dental Equipment								
1	Dental Chair			358,116.00	unit	1		
2	Aerosol Droplet UV Disinfection Fume Extractor			35,000.00	unit	1		
<i>(Please see attached specifications)</i>								

Voucher
 Acctg.

Price Validity: _____

After having carefully read and accepted your General Conditions, I/We quote you on the item at prices noted above.

 Printed Name/Signature of Supplier

 Date/Tel. No./Cellphone No. / ADDRESS

 TIN Number

 Philgeps Registration Number

CANVASSED BY:

**TECHNICAL SPECIFICATIONS
Dental Chair**

HAGER DENTAL HJ-698A (HAI0U)	
Upholstery	P.U Color
<ul style="list-style-type: none"> • Thermostatic water heating system • Spittoon flushing and cup filling system • Electric lift dental Chair • Control panel integrated with tool holder (5 positions: 3 handpiece tubings, 1 three-way syringe, 1 optional) • Rotable ceramic spittoon • Brightness adjustable dental light • Multi-function pedal • Chair movement pedal switch on bottom cover • Pure water bottle 600ml • Dental Stool 	
SPECIFICATIONS	
• Supply Voltage	AC110V220V 50Hz/60Hz
• Motor Voltage	DC24V
• Air Pressure	0.5MPa-0.8MPa
• Water Pressure	0.2MPa-0.4MPa
ACCESSORIES	
• Built in Intra oral camera	
• Built in monitor	
• Built in Ultrasonic scaler	
• Built in curing light	
• Air Turbine Handpiece (High Speed) 2 holes	
• Compressor 1HP	
Warranty: 1 Year	

George N. STONHA DMD



Dental Chair Specs



SUBMITTED BY:

[Handwritten signature]

TECHNICAL SPECIFICATION

Supply voltage: AC110V/220V 50Hz/60Hz

Motor voltage: DC24V

Air pressure: 0.5MPa – 0.8MPa

Water pressure: 0.2MPa – 0.4MPa

Handpiece Tubing: 2/4-hole optional,


Fiber optic tubing with extra charge

- **ACCESSORY INCLUDED**
- Built-in LED curing light
- Built-in ultrasonic scaler
- Intra-oral camera system with Monitor

REVIEWED BY:

[Handwritten signature]
HECTOR P. GARCIA, DMD

Aerosol Droplet UV Disinfection Suction Machine SPECS



SUBMITTED BY:

Specifications

1. Voltage: AC110V-AC240V, 50HZ/ 60HZ
2. Rated power: 1000W
3. Control mode: touch screen/ two gears adjustable
4. Suction flow: 4000L/min
5. Filtration efficiency: (0.3 μm) 99.9%
6. UV intensity (Distance=1M): 1210 μw /cm²
7. UV wavelength: 254nm
8. UV light working life: 10000 hours
9. Plasma sterilization working life: 6000 hours
10. Features: UV-C Irradiation+Plasma sterilization
11. Noise: \leq 60 dB
12. Pressure: -25Kpa
13. Product size: 32*33*75.5 cm (without arm)

REVIEW BY: