

WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City

"PhilHealth Accredited Health Care Provider"

Tel No.: (033) 320 2431 | Fax No.: (033) 3202623 | Email Address: medcenter@wvsu.edu.ph



Standard Form Number: SF-GOOD-59

Revised on: May 24, 2004

Standard Form Title: Request for Quotation

Date : June 7, 2021
Quotation No.: 21-05-144
ABC: P 600,000.00

(Name of Company)

(Address)

Please quote your lowest price on the item/s listed below, subject to the General Conditions stated below, stating the shortest time of delivery and submit your quotation duly signed by your representative not later than <u>June 11</u>, <u>2021</u>; <u>10:00AM</u> in the returned envelope.

Procurement Officer MC

Note:

- 1. ALL ENTRIES MUST BE CLEARLY WRITTEN.
- 2. DELIVERY PERIOD WITHIN 15 CALENDAR DAYS FROM RECEIPT OF PURCHASE ORDER.
- 3. WARRANTY SHALL BE FOR A PERIOD OF __MONTHS FOR SUPPLIES & MATERIALS, 6 MONTHS FOR EQUIPMENT FROM DATE OF ACCEPTANCE BY THE PROCURING ENTITY.
- 4. PRICE VALIDITY SHALL BE FOR A PERIOD OF __ CALENDAR DAYS.
- 5. PHILGEPS REGISTRATION CERTIFICATION SHALL BE ATTACHED UPON SUBMISSION OF THE QUOTATION
- BIDDERS SHALL SUBMIT ORIGINAL BROCHURES SHOWING CERTIFICATIONS OF THE PRODUCT BEING OFFERED (For Equipment)

7. ALTERNATIVE BID OFFER IS NOT ALLOWED. Page 1 of 2 BRAND NAME ITEM NO ITEM & DESCRIPTION QTY. UNIT TOTAL / MODEL NO. PRICE Suction Machine HD (For OR use) unit 1 Specifications: - Brand New - Power Source: 220V-240V, 60Hz Heavy Duty Mobile-Type with casters and Brake systems Sturdy and Durable Body Construct - Suction Range: 0 - 100kPa - Silent or Quiet Operation: not> 50dB - Suction Flow Rate: 40-60 LPM Dial-Type suction pressure regulator Dial-Type suction gauge With foot switch With overflow safety device Collection Container: 2 pieces of not<2L Silicon suction tubing (Reusable/Autoclavable) One suction container remains functional while the other is being cleaned/disinfected With Fluid Collection Systems to provide integrated protection. Operating and Service Manual should be supplied With compatible UPS(10-15 min backup time; with UPS brochure.) Compliant with IEC (60601) for Electrical Safety and EN Standard for Equipment ISO 10079-1:2015 or its equivalent - Regulatory Standard Compliant CE and ISO approved or its equivalent - Warranty: at least 2 years AFTER-SALES REQUIREMENTS: (Please provide certification for each) - Should be, or at least have, an authorized Service Center in the Technicians should comply with the required 72 hours response time for trouble shooting purposes. If the unit needs to be pulled out while being repaired, the company should have an available backup equipment of similar specifications for use at Perinatal Complex. ---- Continued on Page 2 ----