



West Visayas State University

(Formerly Iloilo Normal School)

Bids and Awards Committee Secretariat Office/ Procurement Section

Luna St., La Paz, Iloilo City 5000

Iloilo, Philippines

* Trunkline: (063) (033) 320-0870 loc1103/1104 * Telefax No.: (033) 320-0879

* Website: www.wvsu.edu.ph * Email Address: bac@wvsu.edu.ph



SUPPLEMENTAL/BID BULLETIN

Project Reference No.	IB No. 21-05-16
Name of the Project	Procurement of Various Dental Equipment (Re-bidding)
Location of the Project	WVSU - Main Campus

ADDENDUM NO. 2021-16

June 11, 2021

This **Addendum No. 2021-16** is issued to modify the **Section I. Invitation to Bid, Section III. Bid Data Sheet and Section VII. Technical Specifications.** These shall form an integral part of the Bidding Documents.

I. Section I. Invitation to Bid:

Item	Description	Corrections/ Amendments/ Instructions
2.	The <u>West Visayas State University</u> now invites bids for the above Procurement Project. Delivery of the Goods is required by <u>30 calendar days for each item</u> . Bidders should have completed, within <u>2 years</u> from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II (Instructions to Bidders).	The <u>West Visayas State University</u> now invites bids for the above Procurement Project. Delivery of the Goods is required by <u>30 calendar days for each item</u> . Bidders should have completed, within <u>5 years</u> from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II (Instructions to Bidders).

II. Section III. Bid Data Sheet:

Item	Description	Corrections/ Amendments/ Instructions
ITB Clause		
5.3	For this purpose, contracts similar to the Project shall be: a. <u>Procurement of Various Dental Equipment (Re-bidding)</u> . Completed within <u>2 years</u> prior to the deadline for the submission and receipt of bids.	For this purpose, contracts similar to the Project shall be: <u>a. Procurement of Various Dental Equipment</u> Completed within <u>5 years</u> prior to the deadline for the submission and receipt of bids.

III. Section VII. Technical Specifications:

Item	Description	Corrections/ Amendments/ Instructions
4.	Portable Digital Periapical Dental X-ray Machine	Portable Digital Periapical Dental X-ray Machine
	<ul style="list-style-type: none"> Has total Filtration; The design may allow for a one-handed shooting. Focal spot is 0.4mm- 0.8mm It is at least lighter by 1.64 kg, quick-charging: 1-2 minutes full charging Trouble-free charging system that will prevent battery explosion With laptop With battery charger and extra rechargeable battery pack 	<ul style="list-style-type: none"> Has total Filtration; The design may allow for a one-handed shooting. Focal spot is 0.4mm- 0.8mm It is at least lighter by 1.6 – 1.7 kg, quick-charging: 1-2 minutes full charging Trouble-free charging system that will prevent battery explosion With laptop With battery charger and extra rechargeable battery pack
5.	Ultrasonic Scaler	Ultrasonic Scaler
	<ul style="list-style-type: none"> Can be oriented vertically or horizontally to accommodate any operator space portable 25 Khz 220V Adjustable water supply Frequency of vibration can be controlled With foot control Water hose has valve for easy attachment Base has rubber matting/part for stability Insert or tip is removable & autoclavable Hand piece with LED Includes 5 insert or tip 	<ul style="list-style-type: none"> Can be oriented vertically or horizontally to accommodate any operator space portable 25 Khz 220V Adjustable water supply Frequency of vibration can be controlled With foot control Water hose has valve for easy attachment Base has rubber matting/part for stability Insert or tip is removable & autoclavable Hand piece with or without LED Includes 2 inserts or tips



West Visayas State University

(Formerly Iloilo Normal School)

Bids and Awards Committee Secretariat Office/ Procurement Section

Luna St., La Paz, Iloilo City 5000

Iloilo, Philippines

* Trunkline: (063) (033) 320-0870 loc1103/1104 * Telefax No.: (033) 320-0879

* Website: www.wvsu.edu.ph * Email Address: bac@wvsu.edu.ph



PHI-18939-2-QM

<i>Item</i>	<i>Description</i>	<i>Corrections/ Amendments/ Instructions</i>
6.	<i>Semi-Adjustable Articulator with Face Bow</i>	<i>Semi-Adjustable Articulator with Face Bow</i>
	<p><i>Performance</i> <i>An articulator assists in the accurate fabrication of the biting surfaces of dentures, dental restorations and orthodontic appliances. A Hanau articulator is a semi adjustable type which simulates vertical and horizontal movement of the jaws. It is more accurate in recording jaw movements compared to other semi adjustable articulators. The facebow records the spatial relationship of the maxillary arch to some anterior reference point or points and then transfer this relationship to an articulator. It orients the dental cast in the same relationship to the opening axis or the articulator.</i></p>	<p><i>Performance</i> <i>An articulator assists in the accurate fabrication of the biting surfaces of dentures, dental restorations and orthodontic appliances. A semi-adjustable type simulates vertical and horizontal movement of the jaws. It is accurate in recording jaw movements. The facebow records the spatial relationship of the maxillary arch to some anterior reference point or points and then transfer this relationship to an articulator. It orients the dental cast in the same relationship to the opening axis or the articulator.</i></p>

For guidance and information of all concerned.

JULIUS B. UNDAR

Chairperson, Bids and Awards Committee