



WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City

"PhilHealth Accredited Health Care Provider"

Tel No.: (033) 320 2431 | Fax No.: (033) 3202623 | Email Address: modcenter@wvsu.edu.ph



SUPPLEMENTAL/BID BULLETIN

Project Reference No. **IB No. 22-23**
 Name of the Project **Procurement of Various Accessories for Medical Oxygen**
 Location of the Project **WVSU Medical Center**

ADDENDUM NO. 01 - 2022


May 18, 2022

This Addendum No. 01 is issued to modify or amend items in the Bidding Documents. Attached are the revised Section III. Bid Data Sheet (page 12), Section VI. Schedule of Requirement (page 17), Section VII. Technical Specifications (page 18) and Price Schedule for Goods (pages 21 - 22). These shall form an integral part of the Bidding Documents.

A. Section VII. Technical Specifications

Item No.	Item Description	Qty	Unit	Corrections/Modifications/Amendments
1	Medical Oxygen Cylinder Tank Pigtail High pressure copper tube 5/16" x 1.2mm dia.	12	assy	Medical Oxygen Cylinder Tank Pigtail High pressure copper tube 5/16" x 1.2mm dia. x at least 1.5m (length)- see attached picture
2	Medical Oxygen Flowmeter with Humidifier 0-15 LPM capacity, wall type, DISS Connection Calibrated at 0.35 Mpa, 50 Psi, 20°C. Australian/US Standard	212	assy	
3	Tee Connector for Medical Oxygen Ventilator Compatible for Medical Oxygen Flowmeter and Ventilator Machine	120	pcs	
4	Starting and Running Capacitor for 5.0 Hp Electric Motor for medical vacuum pump 400-552µF, 300-330VAC, 50/60 Hz, Starting Capacitor 40µF, 450 VAC, 50/60Hz, Running Capacitor	4	sets	
5	O-Ring for Medical Oxygen Flowmeter and Medical Vacuum Regulator Varries on the brand/model of the oxygen flowmeter offered O-Ring for CONAIRCO/MEDIARE Vacuum Regulator	100	sets	
Terms and Conditions:				
Submit Calibration Certificate of Oxygen Flowmeter with humidifier				
Submit Certificate of warranty for one year of Oxygen Flowmeter with Humidifier				
Warranty: One year on parts and services				

For guidance and information of all concerned.


MARY EUGENE C. ORILLO
 Chairperson
 Bids and Awards Committee

Received by the Bidder:

Printed Name & Signature

Name of Company/Bidder

Date: _____



Section III. Bid Data Sheet

IB No. 22-23 Procurement of Various Accessories for Medical Oxygen (Revised as indicated in the Bid Bulletin)

ITB Clause					
5.3	For this purpose, contracts similar to the Project shall be: <ol style="list-style-type: none"> a. <u>Procurement of Various Accessories for Medical Oxygen.</u> b. Completed within <u>2 years</u> prior to the deadline for the submission and receipt of bids. 				
7.1	<i>N/A</i>				
12	The price of the Goods shall be quoted DDP <u>in Philippines</u> or the applicable International Commercial Terms (INCOTERMS) for this Project.				
14.1	The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts: <ol style="list-style-type: none"> a. The amount of not less than <u>PhP 23,360.00</u> [Indicate the amount equivalent to two percent (2%) of ABC], if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or b. The amount of not less than <u>PhP 58,400.00</u> [Indicate the amount equivalent to five percent (5%) of ABC] if bid security is in Surety Bond. 				
19.3	Procurement of Various Accessories for Medical Oxygen				
	Item No.	Description	Qty.	Unit	Total ABC
	1	Medical Oxygen Cylinder Tank Pigtail	12	assy	60,000.00
		High pressure copper tube 5/16" x 1.2mm dia. x at least 1.5m (length)			
	2	Medical Oxygen Flowmeter with Humidifier	212	assy	742,000.00
		0-15 LPM capacity, wall type, DISS Connection			
		Calibrated at 0.35 Mpa, 50 Psi, 20°C.			
		Australian/US Standard			
	3	Tee Connector for Medical Oxygen Ventilator	120	pcs	336,000.00
		Compatible for Medical Oxygen Flowmeter and Ventilator Machine			
	4	Starting and Running Capacitor for 5.0 Hp Electric Motor for medical vacuum pump	4	sets	10,000.00
		400-552µF, 300-330VAC, 50/60 Hz, Starting Capacitor			
		40µF, 450 VAC, 50/60Hz, Running Capacitor			
	5	O-Ring for Medical Oxygen Flowmeter and Medical Vacuum Regulator	100	sets	20,000.00
		Varries on the brand/model of the oxygen flowmeter offered			
		O-Ring for CONAIRCO/MEDIARE Vacuum Regulator			
20.2	<i>No additional Requirements.</i>				
21.2	<i>No additional contract documents</i>				

Section VI. Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

IB No. UMC 22-23 Procurement of Various Accessories for Medical Oxygen
(Revised as indicated in the Bid Bulletin)

Item No.	Description	Quantity		Total	Delivered, Weeks/Months
1	Medical Oxygen Cylinder Tank Pigtail	12	assy		
	High pressure copper tube 5/16" x 1.2mm dia. x at least 1.5m (length)				
2	Medical Oxygen Flowmeter with Humidifier	212	assy		
	0-15 LPM capacity, wall type, DISS Connection				
	Calibrated at 0.35 Mpa, 50 Psi, 20°C.				
	Australian/US Standard				
3	Tee Connector for Medical Oxygen Ventilator	120	pcs		
	Compatible for Medical Oxygen Flowmeter and Ventilator Machine				
4	Starting and Running Capacitor for 5.0 Hp Electric Motor for medical vacuum pump	4	sets		
	400-552µF, 300-330VAC, 50/60 Hz, Starting Capacitor				
	40µF, 450 VAC, 50/60Hz, Running Capacitor				
5	O-Ring for Medical Oxygen Flowmeter and Medical Vacuum Regulator	100	sets		
	Varries on the brand/model of the oxygen flowmeter offered				
	O-Ring for CONAIRCO/MEDIARE Vacuum Regulator				

Name of Bidder: _____

Submitted by: _____
(Printed Name and Signature)

Designation: _____

Date: _____

Section VII. Technical Specifications

IB No. UMC 22-23 Procurement of Various Accessories for Medical Oxygen
 (Revised as indicated in the Bid Bulletin)

Item No.	Specification	Statement of Compliance			
		Qty	Unit	Brand	Statement of Compliance
	Procurement of Various Accessories for Medical Oxygen				
1	Medical Oxygen Cylinder Tank Pigtail	12	assy		
	High pressure copper tube 5/16" x 1.2mm dia. x at least 1.5m (length)				
2	Medical Oxygen Flowmeter with Humidifier	212	assy		
	0-15 LPM capacity, wall type, DISS Connection				
	Calibrated at 0.35 Mpa, 50 Psi, 20°C.				
	Australian/US Standard				
3	Tee Connector for Medical Oxygen Ventilator	120	pcs		
	Compatible for Medical Oxygen Flowmeter and Ventilator Machine				
4	Starting and Running Capacitor for 5.0 Hp Electric Motor for medical vacuum pump	4	sets		
	400-552µF, 300-330VAC, 50/60 Hz, Starting Capacitor				
	40µF, 450 VAC, 50/60Hz, Running Capacitor				
5	O-Ring for Medical Oxygen Flowmeter and Medical Vacuum Regulator	100	sets		
	Varries on the brand/model of the oxygen flowmeter offered				
	O-Ring for CONAIRCO/MEDIARE Vacuum Regulator				
	Terms and Conditions:				
	Submit Calibration Certificate of Oxygen Flowmeter with humidifier				
	Submit Certificate of warranty for one year of Oxygen Flowmeter with Humidifier				
	Warranty: One year on parts and services				

Name of Bidder: _____

Submitted by: _____
 (Printed Name and Signature)

Designation: _____

Date: _____

Price Schedule for Goods Offered from Abroad
[shall be submitted with the Bid if bidder is offering goods from Abroad]

For Goods Offered from Abroad

Name of Bidder _____

Project ID No. UMC 22-23 Page 1 of 1

Revised as indicated in the Bid Bulletin

1	2	3	4	5	6	7	8	9
Item	Description	Country of origin	Quantity	Unit Price CIF port of entry (specify port) or CIP named place (specify border point or place of destination)	Total CIF or CIP price per item (col. 4 x 5)	Unit Price Delivered Duty Unpaid (DDU)	Unit Price Delivered Duty Paid (DDP)	Total Price Delivered DDP (col 4 x 8)
1	Medical Oxygen Cylinder Tank Pigtail		12 assy					
	High pressure copper tube 5/16" x 1.2mm dia. x at least 1.5m (length)							
2	Medical Oxygen Flowmeter with Humidifier		212 assy					
	0-15 LPM capacity, wall type, DISS Connection							
	Calibrated at 0.35 Mpa, 50 Psi, 20°C.							
	Australian/US Standard							
3	Tee Connector for Medical Oxygen Ventilator		120 pcs					
	Compatible for Medical Oxygen Flowmeter and Ventilator Machine							
4	Starting and Running Capacitor for 5.0 Hp Electric Motor for medical vacuum pump		4 sets					
	400-552µF, 300-330VAC, 50/60 Hz, Starting Capacitor							
	40µF, 450 VAC, 50/60Hz, Running Capacitor							
5	O-Ring for Medical Oxygen Flowmeter and Medical Vacuum Regulator		100 sets					
	Varries on the brand/model of the oxygen flowmeter offered							
	O-Ring for CONAIRCO/ MEDIARE Vacuum Regulator							
	TOTAL BID							

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____

Price Schedule for Goods Offered from Within the Philippines
[shall be submitted with the Bid if bidder is offering goods from within the Philippines]

For Goods Offered from Within the Philippines

Name of Bidder _____ *Project ID No. UMC 22-23 Page 1 of 1*

Revised as indicated in the Bid Bulletin

1	2	3	4	5	6	7	8	9	10
Item	Description	Country of origin	Quantity	Unit price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if Contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 9) x (col 4)
1	Medical Oxygen Cylinder Tank Pigtail		12 assy						
	High pressure copper tube 5/16" x 1.2mm dia. x at least 1.5m (length)								
2	Medical Oxygen Flowmeter with Humidifier		212 assy						
	0-15 LPM capacity, wall type, DISS Connection								
	Calibrated at 0.35 Mpa, 50 Psi, 20°C.								
	Australian/US Standard								
3	Tee Connector for Medical Oxygen Ventilator		120 pcs						
	Compatible for Medical Oxygen Flowmeter and Ventilator Machine								
4	Starting and Running Capacitor for 5.0 Hp Electric Motor for medical vacuum pump		4 sets						
	400-552µF, 300-330VAC, 50/60 Hz, Starting Capacitor								
	40µF, 450 VAC, 50/60Hz, Running Capacitor								
5	O-Ring for Medical Oxygen Flowmeter and Medical Vacuum Regulator		100 sets						
	Varries on the brand/ model of the oxygen flowmeter offered								
	O-Ring for CONAIRCO/ MEDIARE Vacuum Regulator								
	TOTAL BID								

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____