



**SUPPLEMENTAL/BID BULLETIN**

Project Reference No. **IB No. 22-29**  
 Name of the Project **Procurement of Various Medical Equipment**  
 Location of the Project **WVSU Medical Center**

**ADDENDUM NO. 01 - 2022**

July 8, 2022

This Addendum No. 01 is issued to modify or amend items in the Bidding Documents. Attached are the revised Section VII. Technical Specifications (pages 18-24). These shall form an integral part of the Bidding Documents.

**A. Delivery Period**

As Indicated in the Bidding Documents	Corrections/Modifications/Amendments
Delivery of the Goods is required by Sixty (60) Calendar Days	Delivery of the Goods is required by <b>Ninety (90) Calendar Days</b>

**B. Section VII. Technical Specifications**

Item No.	Item Description	Qty	Unit	Corrections/Modifications/Amendments
<b>1</b>	<b>Resuscitation Warmer</b>	<b>1</b>	<b>unit</b>	<b>Resuscitation Warmer</b>
	Brand New			
	Open Care Unit with Baby Tray and Mattress Sheet			
	Thermoregulation System with Manual and Automatic Controls			
	Temperature monitoring/ sensors with Alarm Settings			
	Patient Gas Supply and Oxygen Blender			
	T-Piece with Provisions for CPAP and PEEP			
	Built-in Suction Device			
				<b>Add:</b>
				<b>IV Pole</b>
				<b>X-ray cassette</b>
	Power Supply Main Line: 220V/240 V			
	<b>Terms and Condition:</b>			<b>Terms and Conditions:</b> <b>(Provide Certification)</b>
	Covers a 12-month warranty with quarterly and service warranty good for 1 year. Iloilo-based technician			
				<b>Add:</b>
				<b>After Sales: Technicians called upon for troubleshooting should comply with the required 72 hours response time</b>
				<b>The company should provide a back-up unit in the case the present unit will be pulled-out indefinitely for repairs.</b>
<b>2</b>	<b>Syringe Pump</b>	<b>4</b>	<b>unit</b>	<b>Syringe Pump</b>
	Brand New			
	Multiple Syringe Sizes			
	Multiple Infusion Settings with Flexible Delivery Rates			
	Alarm Settings (Sound, light or display)			
	Occlusion Detection System			
	Handle and Pole Clamp			
	Power Supply: Main Line: 220V			



# WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City

"PhilHealth Accredited Health Care Provider"

Tel No.: (033) 320 2431 | Fax No.: (033) 3202623 | Email Address: modocenter@wvsu.edu.ph



## SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 22-29  
 Name of the Project Procurement of Various Medical Equipment  
 Location of the Project WVSU Medical Center

### ADDENDUM NO. 01 - 2022

July 8, 2022

Item No.	Item Description	Qty	Unit	Corrections/Modifications/Amendments
	Alternative Battery Supply: 1500mAh, at least 5 hours operating time			
	<b>Terms and Condition:</b>			<b>Terms and Conditions: (Provide certification)</b>
	1 year service warranty			
				<b>Add:</b>
				<b>After Sales: Technicians called upon for troubleshooting should comply with the required 72 hours response time</b>
				<b>The company should provide a back-up unit in the case the present unit will be pulled-out indefinitely for repairs.</b>
<b>3</b>	<b>Desk-type Pulse Oximeter</b>	<b>5</b>	<b>unit</b>	<b>Desk-type Pulse Oximeter</b>
	LCD/LED Display showing heart rate and pulse waveform display			
	O2 saturation monitoring with provisions for poor perfusion			
	Adult, pediatric, and neonatal settings and application, with probe locking design for ideal fit			Adult, pediatric, and neonatal settings and application, with probe locking design for ideal fit, <b>may be a Universal probe</b>
	Audible and visible alarm			
	Portable and stable design			<b>Stability in reading even if with patient movement (pediatric patients)</b>
	Power Supply: Main Line: 240V			
	Alternative Battery Supply: Built-in rechargeable			
	<b>Terms and Condition:</b>			<b>Terms and Conditions: (Provide certification)</b>
	1 year service warranty			
				<b>Add:</b>
				<b>After Sales: Technicians called upon for troubleshooting should comply with the required 72 hours response time</b>
				<b>The company should provide a back-up unit in the case the present unit will be pulled-out indefinitely for repairs.</b>
<b>4</b>	<b>Vein Illuminator/Vein Finder (with stand)</b>	<b>1</b>	<b>unit</b>	<b>Vein Illuminator/Vein Finder (with stand)</b>
	Brand New			
	Operation Mode-Handheld and may be hands-free when attached to wheeled stand			
	Brightness-adjustable within 3 levels			
	Light source-Near-Infrared Light, not an eye or skin hazard			
	Net weight - not more than 1500 grams			
	Dimension - Not bigger than 240mm x 65mm x 65mm			
	Battery Standby/Working Time - at least 2 hours			
	Optimal Imaging Distance - not less than 200mm			
	Alignment Accuracy- Not less than 0.25mm and not more than 0.5mm			



# WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City

"PhilHealth Accredited Health Care Provider"

Tel No. : (033) 320 2431 | Fax No. : (033) 3202623 | Email Address: mdocenter@wvsu.edu.ph



## SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 22-29  
 Name of the Project Procurement of Various Medical Equipment  
 Location of the Project WVSU Medical Center

### ADDENDUM NO. 01 - 2022

July 8, 2022

Item No.	Item Description	Qty	Unit	Corrections/Modifications/Amendments
	Stand Available - Wheeled Floor Stand (height-adjustable)			
	<b>Features:</b>			
	Must be applicable to body parts (hand, wrist, forearm, elbow, thigh, neck and face, able to visualize even on dark skin)			
	Clinical Application-Vein Evaluation/Venipuncture, Drawing Blood			
	Has ISO, FDA, and CE Certificates			
	Electrical Supply: 220V 60Hz			
	Warranty: 1 year on parts and services			
	Shipping: Within 60 days after the receipt of Purchase Order			
	Technical Support: 24/7 open line			Technical Support: 24/7 open line <i>(provide certification)</i>
	Training: On-site after installation			Training: On-site after installation <i>(provide certification)</i>
<b>6</b>	<b>Delivery Table / OB Examination Table</b>	<b>1</b>	<b>unit</b>	<b>Delivery Table / OB Examination Table</b>
	Brand new			
	Stainless steel frame			
	Up and down motion is hydraulic foot control			
	Mobile, with 4 - 5 castors with brake			
	Tilting Capabilities for:			
	Normal Tilting			
	Back Section (up & down)			
	Trendelenburg			
	Reverse Trend.			
	Load Capacity: 200kg-300kg			
	With Accessories for:			
	Arm Support			
	Hand Grips			
	Detachable Padded Stirrups			
	Detachable Padded Foot Stirrups			
	With Infusion Rods, with at least 2 hooks			
	Mattress: detachable; at least 4in; high density foam			
	Revolving adjustable stool			
	Pail, stainless; at least 10L capacity			
	Dimension:			
	Length: 1250mm-1600mm			
	Width: 500mm - 650mm			
	Height: 700-1000mm			Height: <b>620mm-1000mm</b>
	Warranty: 2 years			
	Regulatory Standard Compliant			
	ISO Standards Compliant: (ISO 13485:2016) or its equivalent.			



# WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City

"PhilHealth Accredited Health Care Provider"

Tel No. (033) 320 2431 | Fax No. (033) 3202623 | Email Address: medcenter@wvsu.edu.ph



## SUPPLEMENTAL/BID BULLETIN


Project Reference No. IB No. 22-29  
 Name of the Project Procurement of Various Medical Equipment  
 Location of the Project WVSU Medical Center

### ADDENDUM NO. 01 - 2022

July 8, 2022

Item No.	Item Description	Qty	Unit	Corrections/Modifications/Amendments
	<b>AFTER-SALES REQUIREMENTS: (Provide certification for each)</b>			
	1. Technicians should comply with the required 72 hours response time for trouble shooting purposes.			
	2. If the unit needs to be pulled-out while being repaired, company should have an available back-up equipment of similar specifications for use at unit.			
	3. Must provide Preventive Maintenance schedule within the warranty period.			
<b>8</b>	<b>Patient Transfer Board</b>	<b>1</b>	<b>unit</b>	<b>Patient Transfer Board</b>
	Brand new			Brand new
	Made from thick, semi-flexible thermoplastic material			Made from thick, semi-flexible thermoplastic material
	10-12 longitudinal ridges on base to reduce sideway movement of the board			10-12 longitudinal ridges on base to reduce sideway movement of the board
	Cut out on both sides to assist with transfers and and carrying			Cut out on both sides to assist with transfers <b>and</b> carrying
	Static insulation			Static insulation
	Radio translucent			Radio translucent
	Impact resistant			Impact resistant
	Dimension:			Dimension:
	Length: 1520mm - 1530mm			Length: 1520mm - 1530mm
	Width: 630mm - 640mm			Width: 630mm - 640mm
	Warranty: At least 1 year			Warranty: At least 1 year

For guidance and information of all concerned.

  
**MARY EUGENE C. ORILLO**  
 Chairperson  
 Bids and Awards Committee


Received by the Bidder:

Printed Name & Signature

Name of Bidder/Company

Date: \_\_\_\_\_

# Section VII. Technical Specifications

*IB No. UMC 22-29 Procurement of Various Medical Equipment (Revised as indicated in the Bid Bulletin)*

Item No.	Specification	Statement of Compliance			
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]			
Procurement of Various Medical Equipment		Qty	Unit	Brand	Statement of Compliance
<b>1</b>	<b>Resuscitation Warmer</b>	<b>1</b>	<b>unit</b>		
	Brand New				
	Open Care Unit with Baby Tray and Mattress Sheet				
	Thermoregulation System with Manual and Automatic Controls				
	Temperature monitoring/sensors with Alarm Settings				
	Patient Gas Supply and Oxygen Blender				
	T-Piece with Provisions for CPAP and PEEP				
	Built-in Suction Device				
	<b>IV Pole</b>				
	<b>X-ray cassette</b>				
	Power Supply Main Line: 220V/240 V				
	<b>Terms and Conditions: (Provide Certification)</b>				
	Covers a 12-month warranty with quarterly and service warranty good for 1 year. Iloilo-based technician				
	<b>After Sales: Technicians called upon for troubleshooting should comply with the required 72 hours response time</b>				
	<b>The company should provide a back-up unit in the case the present unit will be pulled-out indefinitely for repairs.</b>				
<b>2</b>	<b>Syringe Pump</b>	<b>4</b>	<b>unit</b>		
	Brand New				
	Multiple Syringe Sizes				
	Multiple Infusion Settings with Flexible Delivery Rates				
	Alarm Settings (Sound, light or display)				
	Occlusion Detection System				
	Handle and Pole Clamp				
	Power Supply: Main Line: 220V				
	Alternative Battery Supply: 1500mAh, at least 5 hours operating time				
	<b>Terms and Conditions: (Provide certification)</b>				
	1 year service warranty				
	<b>After Sales: Technicians called upon for troubleshooting should comply with the required 72 hours response time</b>				
	<b>The company should provide a back-up unit in the case the present unit will be pulled-out indefinitely for repairs.</b>				
<b>3</b>	<b>Desk-type Pulse Oximeter</b>	<b>5</b>	<b>unit</b>		
	LCD/LED Display showing heart rate and pulse waveform display				

# Section VII. Technical Specifications

*IB No. UMC 22-29 Procurement of Various Medical Equipment (Revised as indicated in the Bid Bulletin)*

Item No.	Specification	Statement of Compliance			
		<p><i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>			
Procurement of Various Medical Equipment		Qty	Unit	Brand	Statement of Compliance
	O2 saturation monitoring with provisions for poor perfusion				
	Adult, pediatric, and neonatal settings and application, with probe locking design for ideal fit, <b>may be a Universal probe</b>				
	Audible and visible alarm				
	<b>Stability in reading even if with patient movement (pediatric patients)</b>				
	Power Supply: Main Line: 240V				
	Alternative Battery Supply: Built-in rechargeable				
	<b>Terms and Conditions: (Provide Certification)</b>				
	1 year service warranty				
	<b>After Sales: Technicians called upon for troubleshooting should comply with the required 72 hours response time</b>				
	<b>The company should provide a back-up unit in the case the present unit will be pulled-out indefinitely for repairs.</b>				
<b>4</b>	<b>Vein Illuminator/Vein Finder (with stand)</b>	<b>1</b>	<b>unit</b>		
	Brand New				
	Operation Mode-Handheld and may be hands-free when attached to wheeled stand				
	Brightness-adjustable within 3 levels				
	Light source-Near-Infrared Light, not an eye or skin hazard				
	Net weight - not more than 1500 grams				
	Dimension - Not bigger than 240mm x 65mm x 65mm				
	Battery Standby/Working Time - at least 2 hours				
	Optimal Imaging Distance - not less than 200mm				
	Alignment Accuracy- Not less than 0.25mm and not more than 0.5mm				
	Stand Available - Wheeled Floor Stand (height-adjustable)				
	<b>Features:</b>				
	Must be applicable to body parts (hand, wrist, forearm, elbow, thigh, neck and face, able to visualize even on dark skin)				
	Clinical Application-Vein Evaluation/Venipuncture, Drawing Blood				
	Has ISO, FDA, and CE Certificates				
	Electrical Supply: 220V 60Hz				
	Warranty: 1 year on parts and services				
	Shipping: Within 60 days after the receipt of Purchase Order				
	Technical Support: 24/7 open line ( <b>provide certification</b> )				
	Training: On-site after installation ( <b>provide certification</b> )				

# Section VII. Technical Specifications

*IB No. UMC 22-29 Procurement of Various Medical Equipment (Revised as indicated in the Bid Bulletin)*

Item No.	Specification	Statement of Compliance			
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]			
	Procurement of Various Medical Equipment	Qty	Unit	Brand	Statement of Compliance
<b>5</b>	<b>Suction Machine HD (For OR use)</b>	<b>1</b>	<b>unit</b>		
	Brand New				
	Power Source: 220V-240V, 60Hz				
	Heavy duty				
	Mobile-type with casters and brake systems				
	Sturdy and durable body construct				
	Suction Range: 0 – 100kPa				
	Silent or quiet operation: not > 50 dB				
	Suction Flow Rate: 40-60 LPM				
	Dial-type suction pressure regulator				
	Dial-type suction gauge				
	With foot switch				
	With overflow safety device				
	Collection container: 2 pieces of not < 2L each.				
	Silicon suction tubing (Reusable / Autoclavable).				
	One suction container remains functional while the other is being cleaned/disinfected.				
	With Fluid Collection Systems to provide integrated protection.				
	Operating and service manual should be supplied				
	With compatible UPS (10-15min back-up time; with UPS brochure.)				
	Compliant with IEC (60601) for Electrical Safety and EN Standard for Equipment				
	ISO 10079-1:2015 or its equivalent				
	Regulatory Standard Compliant				
	CE and ISO approved or its equivalent				
	Warranty: at least 2 years				
	<b>AFTER-SALES REQUIREMENTS:</b>				
	(Please provide certification for each)				
	Should be, or at least have, an authorized Service Center in the Visayas.				
	Technicians should comply with the required 72 hours response time for trouble shooting purposes.				

# Section VII. Technical Specifications

*IB No. UMC 22-29 Procurement of Various Medical Equipment (Revised as indicated in the Bid Bulletin)*

Item No.	Specification	Statement of Compliance			
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]			
Procurement of Various Medical Equipment		Qty	Unit	Brand	Statement of Compliance
6	<b>Delivery Table / OB Examination Table</b>	1	unit		
	Brand new				
	Stainless steel frame				
	Up and down motion is hydraulic foot control				
	Mobile, with 4 - 5 castors with brake				
	Tilting Capabilities for:				
	Normal Tilting				
	Back Section (up & down)				
	Trendelenburg				
	Reverse Trend.				
	Load Capacity: 200kg-300kg				
	With Accessories for:				
	Arm Support				
	Hand Grips				
	Detachable Padded Stirrups				
	Detachable Padded Foot Stirrups				
	With Infusion Rods, with at least 2 hooks				
	Mattress: detachable; at least 4in; high density foam				
	Revolving adjustable stool				
	Pail, stainless; at least 10L capacity				
	Dimension:				
	Length: 1250mm-1600mm				
	Width: 500mm - 650mm				
	Height: <b>620mm</b> -1000mm				
	Warranty: 2 years				
	Regulatory Standard Compliant				
	ISO Standards Compliant: (ISO 13485:2016) or its equivalent.				



# Section VII. Technical Specifications

*IB No. UMC 22-29 Procurement of Various Medical Equipment (Revised as indicated in the Bid Bulletin)*

Item No.	Specification	Statement of Compliance			
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]			
Procurement of Various Medical Equipment		Qty	Unit	Brand	Statement of Compliance
	<b>AFTER-SALES REQUIREMENTS: (Provide certification for each)</b>				
	1. Technicians should comply with the required 72 hours response time for trouble shooting purposes.				
	2. If the unit needs to be pulled-out while being repaired, company should have an available back-up equipment of similar specifications for use at unit.				
	3. Must provide Preventive Maintenance schedule within the warranty period.				
<b>7</b>	<b>Electro-Surgical Unit</b>	<b>2</b>	<b>unit</b>		
	Brand New				
	Power Source: 220V, 60Hz				
	Heavy duty				
	Output Configuration: Isolated output				
	Display: At least 6.2 inch LCD touchscreen				
	Enclosure: Magnesium				
	Equipment Class - Class I Equipment, Type CF				
	Duty Cycle: Must be capable of operating at 25% duty cycle (10 seconds active and 30 seconds inactive in any mode for a period of 4 hours).				
	Must have internal memory with real time clock and memory capacity of at least 8GB.				
	Must have different tones for the following settings: Cut, Coag, Bipolar and Shared Coag				
	Must have at least seven (7) monopolar currents with the following settings:				
	a. For monopolar cut the following settings must be available:				
	i. Pure				
	ii. Blend				
	b. For monopolar coagulation the following settings must be available:				
	i. Soft				
	ii. Fulgurate				
	iii. Shared fulgurate				
	iv. Spray				
	v. Shared Spray				
	c. Must have at least six (6) Bipolar currents. The following settings must be available:				
	i. Precise				
	ii. Standard				

# Section VII. Technical Specifications

*IB No. UMC 22-29 Procurement of Various Medical Equipment (Revised as indicated in the Bid Bulletin)*

Item No.	Specification	Statement of Compliance			
		<p><i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>			
Procurement of Various Medical Equipment		Qty	Unit	Brand	Statement of Compliance
	iii. Macro				
	iv. Low				
	v. Medium				
	vi. High				
	Must have at least recall program from the last shutdown for power settings.				
	Must have Auto-bipolar function and Ammeter reading.				
	Must have a neutral electrode alarm that is activated in case contact with patient is broken for maximum safety.				
	Automatic adjustment, controls all modes and effects. As tissue resistance increases from zero, the energy platform outputs constant current, followed by constant power, followed by constant voltage.				
	The maximum output voltage is controlled to reduce capacitive coupling and video interference and to minimize sparking.				
	Read tissue changes at least 434,000 times per second and adjust to tissue changes at least 434,000 times per second.				
	Must have at least one (1) bipolar output plug and at least two (2) monopolar.				
	Must be capable of two (2) simultaneous monopolar coagulation.				
	Must be supplied with two (2) twin foot pedals: One (1) for monopolar and bipolar use, and one (1) for independent bipolar only use.				
	Software upgradeable.				
	Dimensions:				
	Weight: 19.5-20 lb      Width: 13.18-14 in				
	Height : 5.73-6 in      Length : 17.15-18 in				
	Compliant with IEC (60601) for Electrical Safety and EN Standard for Equipment.				
	Compliant with standard regulatory requirements for ISO or its equivalent.				
	<b>Others:</b>				
	1. Must provide accompanying stable, flat surface table or cart with casters, with brake.				
	2. With compatible UPS (10-15min back-up time; with UPS brochure.)				
	3. Must be compatible with existing smoke evacuator machine.				
	4. Must provide 1 bipolar cushioning forcep				
	5. 1 year warranty on service (preventive maintenance and calibration), parts and must provide warranty certificate.				

# Section VII. Technical Specifications

*IB No. UMC 22-29 Procurement of Various Medical Equipment (Revised as indicated in the Bid Bulletin)*

Item No.	Specification	Statement of Compliance			
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]			
	<b>Procurement of Various Medical Equipment</b>	<b>Qty</b>	<b>Unit</b>	<b>Brand</b>	<b>Statement of Compliance</b>
<b>8</b>	<b>Patient Transfer Board</b>	<b>1</b>	<b>unit</b>		
	Brand new				
	Made from thick, semi-flexible thermoplastic material				
	10-12 longitudinal ridges on base to reduce sideway movement of the board				
	Cut out on both sides to assist with transfers <b>and</b> carrying				
	Static insulation				
	Radio translucent				
	Impact resistant				
	Dimension:				
	Length: 1520mm - 1530mm				
	Width: 630mm - 640mm				
	Warranty: At least 1 year				

Name of Bidder: \_\_\_\_\_

Submitted by: \_\_\_\_\_  
*(Printed Name and Signature)*

Designation: \_\_\_\_\_

Date: \_\_\_\_\_