



# WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City

"PhilHealth Accredited Health Care Provider"

Tel No.: (033) 320 2431 | Fax No.: (033) 3202623 | Email Address: medcenter@wvsu.edu.ph



## SUPPLEMENTAL/BID BULLETIN

Project Reference No. **IB No. 22-37**  
 Name of the Project **Procurement of Brand New Laparoscopic System**  
 Location of the Project **WVSU Medical Center**

### ADDENDUM NO. 01 - 2022

August 22, 2022

This Addendum No. 01 is issued to modify or amend items in the Bidding Documents. Attached are the **Revised Section III. Bid Data Sheet (page 12), Section VI. Schedule of Requirements (page 17), Section VII. Technical Specifications (pages 18-22) and Price Schedule for Goods (pages 25-26)**. These shall form an integral part of the Bidding Documents.

#### A. Change of Project Title

As Specified in the Bidding Documents	Corrections/Modifications/Amendments
Procurement of Brand New Laparoscopic System	<b>Procurement of Brand New Hysteroscopy and Laparoscopy System</b>

#### B. Delivery Period

As Specified in the Bidding Documents	Corrections/Modifications/Amendments
Sixty (60) Calendar Days	<b>Seventy-Five (75) Calendar Days</b>

#### C. Section III. Bid Data Sheet

As Specified in the Bidding Documents		Corrections/Modifications/Amendments																					
ITB Clause																							
5.3	For this purpose, contracts similar to the Project shall be: a. <u>Procurement of Brand New Laparoscopic System.</u>	For this purpose, contracts similar to the Project shall be: a. <b><u>Procurement of Brand New Hysteroscopy and Laparoscopy System.</u></b>																					
19.3	Procurement of Brand New Laparoscopic System - Php 15,000,000.00	<b>Procurement of Brand New Hysteroscopy and Laparoscopy System - Php 15,000,000.00</b>																					
	<table border="1"> <thead> <tr> <th>Item No.</th> <th>Description</th> <th>Qty.</th> <th>Unit</th> <th>Total ABC</th> </tr> </thead> <tbody> <tr> <td>1</td> <td>Laparoscopic System, Brand New</td> <td>1</td> <td>set</td> <td>15,000,000.00</td> </tr> </tbody> </table>	Item No.	Description	Qty.	Unit	Total ABC	1	Laparoscopic System, Brand New	1	set	15,000,000.00	<table border="1"> <thead> <tr> <th>Item No.</th> <th>Description</th> <th>Qty.</th> <th>Unit</th> <th>Total ABC</th> </tr> </thead> <tbody> <tr> <td>1</td> <td><b>Hysteroscopy and Laparoscopy System, Brand New</b></td> <td>1</td> <td>set</td> <td>15,000,000.00</td> </tr> </tbody> </table>	Item No.	Description	Qty.	Unit	Total ABC	1	<b>Hysteroscopy and Laparoscopy System, Brand New</b>	1	set	15,000,000.00	
Item No.	Description	Qty.	Unit	Total ABC																			
1	Laparoscopic System, Brand New	1	set	15,000,000.00																			
Item No.	Description	Qty.	Unit	Total ABC																			
1	<b>Hysteroscopy and Laparoscopy System, Brand New</b>	1	set	15,000,000.00																			

#### D. Section VI. Schedule of Requirements

As Specified in the Bidding Documents			Corrections/Modifications/Amendments	
Item Number	Description	Quantity	Description	Quantity
1	Laparoscopic System, Brand New	1 set	<b>Hysteroscopy and Laparoscopy System, Brand New</b>	



# WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City

"PhilHealth Accredited Health Care Provider"

Tel No.: (033) 320 2431 | Fax No.: (033) 3202623 | Email Address: medcenter@wvsu.edu.ph



## SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 22-37  
 Name of the Project Procurement of Brand New Laparoscopic System  
 Location of the Project WVSU Medical Center

### ADDENDUM NO. 01 - 2022

August 22, 2022

#### E. Section VII. Technical Specifications

As Specified in the Bidding Documents				Corrections/Modifications/Amendments		
Item No.	Item Description	Qty.	Unit	Item Description	Qty.	Unit
	Procurement of Brand New Laparoscopic System			<b>Procurement of Brand New Hysteroscopy and Laparoscopy System</b>		
1	Laparoscopic System, Brand New	1	set	<b>Hysteroscopy and Laparoscopy System, Brand New</b>		
	SPECIFICATIONS					
	3. 4K CAMERA HEAD					
	<ul style="list-style-type: none"> <li>Compatible with 3CCD Camera Control Unit</li> </ul>			<ul style="list-style-type: none"> <li>Compatible with <b>4k</b> Camera Control Unit</li> </ul>		
	<ul style="list-style-type: none"> <li>Full high definition</li> </ul>					
	<ul style="list-style-type: none"> <li>Autoclavable camera head</li> </ul>			<ul style="list-style-type: none"> <li>Autoclavable camera head <b>add: (/ Gas sterilization)</b></li> </ul>		
	<ul style="list-style-type: none"> <li>Can control image enhancement system via camera head</li> </ul>					
	<ul style="list-style-type: none"> <li>Use in NIR/ICG fluorescence endoscopy in standard white light endoscopy</li> </ul>					
	5. DOCUMENTATION SYSTEM					
	<ul style="list-style-type: none"> <li>Touch screen display</li> </ul>					
	<ul style="list-style-type: none"> <li>Documentation of still images and videos</li> </ul>					
	<ul style="list-style-type: none"> <li>With integrated documentation function, transferable to external storage media via USB</li> </ul>					
	<ul style="list-style-type: none"> <li>Full high definition, 2D/3D display</li> </ul>			<ul style="list-style-type: none"> <li>Full high definition, <b>3D display</b></li> </ul>		
	<ul style="list-style-type: none"> <li>Patient data input and storage via USB drive</li> </ul>					
	7. SCOPES					
	<ul style="list-style-type: none"> <li>Rigid Scope 10 mm 0°</li> </ul>					
	<ul style="list-style-type: none"> <li>Rigid Scope 10 mm 30°</li> </ul>					
	<ul style="list-style-type: none"> <li>Rigid Scope 5 mm 30°</li> </ul>			<ul style="list-style-type: none"> <li>Rigid Scope <b>5 – 5.5 mm 30°</b></li> </ul>		



**SUPPLEMENTAL/BID BULLETIN**

Project Reference No. **IB No. 22-37**  
 Name of the Project **Procurement of Brand New Laparoscopic System**  
 Location of the Project **WVSU Medical Center**

**ADDENDUM NO. 01 - 2022**

August 22, 2022

**E. Section VII. Technical Specifications**

As Specified in the Bidding Documents				Corrections/Modifications/Amendments		
Item No.	Item Description	Qty.	Unit	Item Description	Qty.	Unit
11.	PUMP AND SUCTION MACHINE					
	▪ Irrigation: 2 L/min					
	▪ Suction 2L/min					
	▪ Flow rate 0-1000 ml/min			▪ Flow rate <b>0 – 2 L/min.</b>		
	▪ Maximum irrigation pressure 0-1200mmHg					
	▪ With sterile tubing package					
	▪ Low noise system			▪ Low noise system <b>add: (in decibels)</b>		
	HYSTEROSCOPY SYSTEM	1	set			
1.	Diagnostic Hysteroscopy					
	▪ Compact Hysteroscope 3mm			▪ Compact Hysteroscope <b>2.9 – 3.8 mm</b>		
	▪ Compact Hysteroscope 5mm			▪ Compact Hysteroscope <b>4 – 5 mm</b>		
	▪ Grasping forceps					
	▪ Biopsy punch					
	▪ Scissors					
3.	Irrigation and Pump System					
	▪ Maximum irrigation pressure: 800 ml/min			▪ Maximum irrigation pressure: <b>20 – 150 mmHg</b>		
	▪ Maximum irrigation pressure: 200 mmHg			▪ <b>Suction flow: 100 -180 mmHg</b>		
	<b>OTHER TERMS AND CONDITIONS</b>					
	1. 24/7 for service concerns with trained technicians			1. 24/7 for service concerns with trained technicians <b>add: (provide certification)</b>		
	2. Should be, or at least have, an authorized Service Center in the Visayas					
	3. Technical operation and Maintenance trainings of the above system will be provided			3. Technical operation and Maintenance trainings of the above system will be provided <b>add: (provide certification)</b>		
	4. Compliant with IEC (60601) for Electrical Safety and EN Standard for Equipment					



# WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City

"PhilHealth Accredited Health Care Provider"

Tel No.: (033) 320 2431 | Fax No.: (033) 3202623 | Email Address: medcenter@wvsu.edu.ph



## SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 22-37  
 Name of the Project Procurement of Brand New Laparoscopic System  
 Location of the Project WVSU Medical Center

### ADDENDUM NO. 01 - 2022

August 22, 2022


#### E. Section VII. Technical Specifications

As Specified in the Bidding Documents				Corrections/Modifications/Amendments		
Item No.	Item Description	Qty.	Unit	Item Description	Qty.	Unit
	5. Certificate of guarantee that the equipment is brand new, unused and not a discontinued model or was in the product recall					
	6. Certificate of guarantee that the supplier shall be responsible for the delivery and installation					
	7. Certificate from the manufacturer to honor the terms and conditions stated in the contract throughout its duration even in the occurrence of a change of dealership or any similar circumstance.					
	8. Provide a demo unit post qualification			8. Provide a demo unit <i>for post qualification within the timeline set by R.A. 9184</i>		
	9. At least (2) two years warranty on parts and service including at least bi-annual preventive maintenance					

#### F. Price Schedule for Goods

As Specified in the Bidding Documents				Corrections/Modifications/Amendments		
Item	Description	Country of Origin	Quantity	Description	Country of Origin	Quantity
1	Laparoscopic System, Brand New		1 Set	<i>Hysteroscopy and Laparoscopy System, Brand New</i>		

For guidance and information of all concerned.

  
**MARY EUGENE C. ORILLO**  
 Chairperson  
 Bids and Awards Committee

Received by the Bidder:

\_\_\_\_\_  
Printed Name & Signature

\_\_\_\_\_  
Name of Bidder/Company

Date: \_\_\_\_\_

## Section III. Bid Data Sheet

(Revised as indicated in the Bid Bulletin No. 01-2022)

ITB Clause											
5.3	For this purpose, contracts similar to the Project shall be: <ol style="list-style-type: none"> <li>a. <b><u>Procurement of Brand New Hysteroscopy and Laparoscopy System.</u></b></li> <li>b. Completed within <b><u>2 years</u></b> prior to the deadline for the submission and receipt of bids.</li> </ol>										
7.1	<i>N/A</i>										
12	The price of the Goods shall be quoted DDP <b><i>in Philippines</i></b> or the applicable International Commercial Terms (INCOTERMS) for this Project.										
14.1	The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts: <ol style="list-style-type: none"> <li>a. The amount of not less than <b><u>PhP 300,000.00</u></b> [Indicate the amount equivalent to two percent (2%) of ABC], if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or</li> <li>b. The amount of not less than <b><u>PhP 750,000.00</u></b> [Indicate the amount equivalent to five percent (5%) of ABC] if bid security is in Surety Bond.</li> </ol>										
19.3	<p><b><i>Procurement of Brand New Hysteroscopy and Laparoscopy System - Php 15,000,000.00</i></b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="text-align: center;">Item No.</th> <th style="text-align: center;">Description</th> <th style="text-align: center;">Qty.</th> <th style="text-align: center;">Unit</th> <th style="text-align: center;">Total ABC</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">1</td> <td><b><i>Hysteroscopy and Laparoscopy System, Brand New</i></b></td> <td style="text-align: center;">1</td> <td style="text-align: center;">set</td> <td style="text-align: right;">15,000,000.00</td> </tr> </tbody> </table>	Item No.	Description	Qty.	Unit	Total ABC	1	<b><i>Hysteroscopy and Laparoscopy System, Brand New</i></b>	1	set	15,000,000.00
Item No.	Description	Qty.	Unit	Total ABC							
1	<b><i>Hysteroscopy and Laparoscopy System, Brand New</i></b>	1	set	15,000,000.00							
20.2	<b><i>No additional Requirements.</i></b>										
21.2	<b><i>No additional contract documents</i></b>										

## *Section VI. Schedule of Requirements*

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

**IB No. UMC 22-37 Procurement of Brand New Hysteroscopy and Laparoscopy System  
(Revised as indicated in the Bid Bulletin No. 01-2022)**

Item Number	Description	Quantity		Total	Delivered, Weeks/Months
1	<b>Hysteroscopy and Laparoscopy System, Brand New</b>	1	set		
	---nothing follows---				

Name of Bidder: \_\_\_\_\_

Submitted by: \_\_\_\_\_  
(Printed Name and Signature)

Designation: \_\_\_\_\_

Date: \_\_\_\_\_

# Section VII. Technical Specifications

**IB No. UMC 22-37 Procurement of Brand New Hysteroscopy and Laparoscopy System  
(Revised as indicated in the Bid Bulletin No. 01-2022)**

Item No.	Specification	Statement of Compliance				
		Qty	Unit	Brand	Model	Statement of Compliance
	<b>Procurement of Brand New Hysteroscopy and Laparoscopy System</b>					
1	<b>Hysteroscopy and Laparoscopy System, Brand New</b>	1	set			
	<b>SPECIFICATIONS</b>					
	<b>1. MONITOR</b>					
	▪ At least 27" full high definition 4K Monitor					
	▪ Screen Resolution 3840 x 2160					
	▪ Color Systems PAL/NTSC					
	▪ Image format 16:9					
	▪ Wall mounted					
	▪ Video/Image Input HDMI, DVI-D, 3G/HD-SDI,12G-SD					
	▪ Video/Image Output G-SDI, DVI-D, 3G/HD-SDI					
	<b>2. 4K CAMERA CONTROLLER</b>					
	▪ Screen resolution 3840 x 2160 pixels					
	▪ For use with full High Definition three-chip camera					
	▪ With Image Enhancement System					
	▪ Indocyanine green fluorescence imaging					
	▪ Image and Video Archive USB					
	<b>3. 4K CAMERA HEAD</b>					
	▪ Compatible with 4k Camera Control Unit					
	▪ Full high definition					
	▪ Autoclavable camera head / <b>Gas sterilization</b>					
	▪ Can control image enhancement system via camera head					
	▪ Use in NIR/ICG fluorescence endoscopy in standard white light endoscopy					
	<b>4. CARBON DIOXIDE INSUFFLATOR</b>					
	▪ 45-50L/min flow rate					
	▪ With smoke evacuator					
	▪ Touchscreen control for preselection of set values					
	▪ User-friendly and practical design					
	▪ CO2 gas heating function					

# Section VII. Technical Specifications

**IB No. UMC 22-37 Procurement of Brand New Hysteroscopy and Laparoscopy System  
(Revised as indicated in the Bid Bulletin No. 01-2022)**

Item No.	Specification	Statement of Compliance				
		<p><i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>				
	<b>Procurement of Brand New Hysteroscopy and Laparoscopy System</b>	Qty	Unit	Brand	Model	Statement of Compliance
<b>5.</b>	<b>DOCUMENTATION SYSTEM</b>					
	▪ Touch screen display					
	▪ Documentation of still images and videos					
	▪ With integrated documentation function, transferable to external storage media via USB					
	▪ Full high definition, <b>3D display</b>					
	▪ Patient data input and storage via USB drive					
<b>6.</b>	<b>LIGHT SOURCE</b>					
	▪ LED light source for endoscopic NIR/ICG application					
	▪ Switchable between NIR/ICG and white light endoscopy					
	▪ Compatible with rigid and flexible scope					
	▪ Light cable					
<b>7.</b>	<b>SCOPES</b>					
	▪ Rigid Scope 10 mm 0°					
	▪ Rigid Scope 10 mm 30°					
	▪ Rigid Scope 5 – 5.5 mm 30°					
<b>8.</b>	<b>MONOPOLAR &amp; BIPOLAR CABLE</b>					
<b>9.</b>	<b>INSTRUMENTS</b>					
	▪ Bowel Grasper					
	▪ Allis Grasper					
	▪ Maryland dissector					
	▪ Bipolar forceps					
	▪ Uterine manipulator					
	▪ Hook electrode					
	▪ Metzenbaum Scissors					
	▪ Suction and Irrigation tube and handle					
	▪ Macro needle Holder					
	▪ Cleaning Set					



# Section VII. Technical Specifications

**IB No. UMC 22-37 Procurement of Brand New Hysteroscopy and Laparoscopy System  
(Revised as indicated in the Bid Bulletin No. 01-2022)**

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
<b>Procurement of Brand New Hysteroscopy and Laparoscopy System</b>		Qty	Unit	Brand	Model	Statement of Compliance
<b>10.</b>	<b>TROCARS</b>					
	▪ Pyramidal tip					
	▪ Blunt, conical tip					
<b>11.</b>	<b>PUMP AND SUCTION MACHINE</b>					
	▪ Irrigation: 2 L/min					
	▪ Suction 2L/min					
	▪ Flow rate <b>0 – 2 L/min.</b>					
	▪ Maximum irrigation pressure 0-1200mmHg					
	▪ With sterile tubing package					
	▪ Low noise system <i>(In decibels)</i>					
<b>12.</b>	<b>ENERGY SOURCE</b>					
	▪ Compatible for interdisciplinary use (OB-Gyn)					
	▪ Compatible with laparoscopy and operative hysteroscopy					
	▪ Capable for Bipolar and Monopolar Coagulation and Cutting					
	▪ With two pedal footswitch					
	▪ Automatic activation via hand switch or footswitch					
<b>13.</b>	<b>EQUIPMENT TROLLEY</b>					
	▪ Mobile with locking brakes					
	▪ With swivel arm for slave monitor					
	▪ With shelves and drawer					
<b>14.</b>	<b>INSTRUMENT BOX</b>					
<b>15.</b>	<b>COMPATIBLE UPS</b>					
	▪ 220-230V output voltage					
<b>16.</b>	<b>POWER SUPPLY</b>					
	▪ Can Operate at 220-240 V, 60Hz					
	<b>HYSTEROSCOPY SYSTEM</b>	1	set			
<b>1.</b>	<b>Diagnostic Hysteroscopy</b>					
	▪ Compact Hysteroscope <b>2.9 – 3.8 mm</b>					
	▪ Compact Hysteroscope <b>4 - 5mm</b>					

# Section VII. Technical Specifications

**IB No. UMC 22-37 Procurement of Brand New Hysteroscopy and Laparoscopy System  
(Revised as indicated in the Bid Bulletin No. 01-2022)**

Item No.	Specification	Statement of Compliance				
		<p><i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>				
<b>Procurement of Brand New Hysteroscopy and Laparoscopy System</b>		Qty	Unit	Brand	Model	Statement of Compliance
	▪ Grasping forceps					
	▪ Biopsy punch					
	▪ Scissors					
<b>2.</b>	<b>Operative Hysteroscopy with Bipolar Resectoscope</b>					
	▪ Autoclavable 12 degree Telescope, 2.7-3mm					
	▪ Bipolar working element					
	▪ Inner sheath					
	▪ Outer sheath					
	▪ Obturator					
	▪ Bipolar cutting loop					
	▪ Bipolar hook electrode					
	▪ Bipolar roller electrode					
	▪ Bipolar cable					
<b>3.</b>	<b>Irrigation and Pump System</b>					
	▪ Maximum irrigation pressure: <b>20 – 150 mmHg</b>					
	▪ <b>Suction flow: 100 -180 mmHg</b>					
<b>OTHER TERMS AND CONDITIONS</b>						
	1. 24/7 for service concerns with trained technicians <b>(provide certification)</b>					
	2. Should be, or at least have, an authorized Service Center in the Visayas					
	3. Technical operation and Maintenance trainings of the above system will be provided <b>(provide certification)</b>					
	4. Compliant with IEC (60601) for Electrical Safety and EN Standard for Equipment					
	5. Certificate of guarantee that the equipment is brand new, unused and not a discontinued model or was in the product recall					
	6. Certificate of guarantee that the supplier shall be responsible for the delivery and installation					
	7. Certificate from the manufacturer to honor the terms and conditions stated in the contract throughout its duration even in the occurrence of a change of dealership or any similar circumstance.					

## Section VII. Technical Specifications

**IB No. UMC 22-37 Procurement of Brand New Hysteroscopy and Laparoscopy System  
(Revised as indicated in the Bid Bulletin No. 01-2022)**

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	<b>Procurement of Brand New Hysteroscopy and Laparoscopy System</b>	<b>Qty</b>	<b>Unit</b>	<b>Brand</b>	<b>Model</b>	<b>Statement of Compliance</b>
8.	Provide a demo unit <i>for post qualification within the timeline set by R.A. 9184</i>					
9.	At least (2) two years warranty on parts and service including at least bi-annual preventive maintenance					
	<b>TRAINING AND OTHER DOCUMENTS</b>					
10.	Application and clinical training for staff					
11.	Provision of Use, Technical Operation and Maintenance					
	<b>REGULATORY STANDARDS</b>					
12.	Regulatory Standard Compliant					
13.	CE and ISO approved or its equivalent					

Name of Bidder: \_\_\_\_\_

Submitted by: \_\_\_\_\_  
(Printed Name and Signature)

Designation: \_\_\_\_\_

Date: \_\_\_\_\_

**Price Schedule for Goods Offered from Abroad**  
*[shall be submitted with the Bid if bidder is offering goods from Abroad]*

**For Goods Offered from Abroad**

Name of Bidder \_\_\_\_\_

**Project ID No. UMC 22-37 Page 1 of 1**

*(Revised as indicated in the Bid Bulletin No. 01-2022)*

1	2	3	4	5	6	7	8	9
Item	Description	Country of origin	Quantity	Unit Price CIF port of entry (specify port) or CIP named place (specify border point or place of destination)	Total CIF or CIP price per item (col. 4 x 5)	Unit Price Delivered Duty Unpaid (DDU)	Unit Price Delivered Duty Paid (DDP)	Total Price Delivered DDP (col 4 x 8)
1	<b>PROCUREMENT OF BRAND NEW HYSTEROSCOPY AND LAPAROSCOPY SYSTEM</b> <i>Hysteroscopy and Laparoscopy System, Brand New</i>		1 Set					
	<b>TOTAL BID FOR THE PROJECT</b>							

Name: \_\_\_\_\_

Legal Capacity: \_\_\_\_\_

Signature: \_\_\_\_\_

Duly authorized to sign the Bid for and behalf of: \_\_\_\_\_

**Price Schedule for Goods Offered from Within the Philippines**  
*[shall be submitted with the Bid if bidder is offering goods from within the Philippines]*

**For Goods Offered from Within the Philippines**

Name of Bidder \_\_\_\_\_

**Project ID No. UMC 22-37 Page 1 of 1**

*(Revised as indicated in the Bid Bulletin No. 01-2022)*

1	2	3	4	5	6	7	8	9	10
Item	Description	Country of origin	Quantity	Unit price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if Contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 9) x (col 4)
1	<b>PROCUREMENT OF BRAND NEW HYSTEROSCOPY AND LAPAROSCOPY SYSTEM</b>								
	<i>Hysteroscopy and Laparoscopy System, Brand New</i>		1 Set						
	<b>TOTAL BID FOR THE PROJECT</b>								

Name: \_\_\_\_\_  
 Legal Capacity: \_\_\_\_\_  
 Signature: \_\_\_\_\_  
 Duly authorized to sign the Bid for and behalf of: \_\_\_\_\_