



SUPPLEMENTAL/BID BULLETIN

Project Reference No. **IB No. 22-43**
 Name of the Project **Procurement of Various Medical Equipment-A**
 Location of the Project **WVSU Medical Center**

ADDENDUM NO. 01 - 2022

September 30, 2022

This Addendum No. 01 is issued to modify or amend items in the Bidding Documents. Attached is the revised Section VII. Technical Specifications (pages 18-21). These shall form an integral part of the Bidding Documents.

A. Cost of Bidding Documents

COST OF BIDDING DOCUMENTS PER ITEM			
Item No.	Description	Total ABC	Cost of Bidding Documents
1	OPERATING ROOM TABLE, Brand New	4,000,000.00	5,000.00
2	SUCTION MACHINE HD (for OR use)	400,000.00	500.00
3	CRIB WITH BASSINET	80,000.00	500.00
4	SUCTION MACHINE	550,000.00	1,000.00

B. Section VIII. Checklist of Technical and Financial Documents

Class "B" Documents – if applicable, a duly signed joint venture agreement (JVA) in case the joint venture is already in existence; or duly notarized statements from all the potential joint venture partners stating that they will enter into and abide by the provisions of the JVA in the instance that the bid is successful. **If NOT applicable, please provide a certification or statement duly signed by the authorized representative that JVA is NOT APPLICABLE.**

C. Section VII. Technical Specifications

Item No.	Item Description	Qty	Unit	Corrections/Modifications/Amendments
1	OPERATING ROOM TABLE, Brand New	2	UNIT	
	• Table Structure			
	✓Table top: 4- 5			
	✓Frame: Made of stainless- steel structure			
	✓Base clearance that permits under the base space of the foot			
	✓Emergency Unlock- handle below the chassis, quick release handle in emergency case			
	✓Electoral conductive wheels			
	✓Electro-motorized floor locks with 4- 5 stamps, casters lock and unlock.			
	• Table Measurement			
	✓Table length: 1900-2100mm			
	✓Table width: 530mm- 590mm			
	✓Table height (without pad): 600mm- 1000mm			
	✓Weight capacity: 450- 500kg			
	✓Translucent, C-Arm compatible			
	• Table Movements			
	✓Trendelenburg/Reverse- 30°/30°			
	✓Tilt: ±25°			
	✓Longitudinal slide overall: 200-270mm			
	✓Joint back section: 75°-80°			
	✓Joint leg section: 85°-90°			
	✓Head section length: 300-340mm			
	✓Back & Seat Section Length: 1000-1010mm			
	✓Leg section length: 700-740mm			
	✓One button zero position for motorized leveling of all sections			
	✓Flex up and down for kidney break positioning can be adjusted viacable remote control and column keypad			
	• Power System			
	✓External power supply: 220-240V, 60Hz			
	✓Power input: max. 450 VA			
	✓Built-in battery, automatic battery charging			
	✓AC/DC auto switch			



WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City

"PhilHealth Accredited Health Care Provider"

Tel No.: (033) 320 2431 | Fax No.: (033) 3202623 | Email Address: medcenter@wvsu.edu.ph



SUPPLEMENTAL/BID BULLETIN

Project Reference No. **IB No. 22-43**
 Name of the Project **Procurement of Various Medical Equipment-A**
 Location of the Project **WVSU Medical Center**

ADDENDUM NO. 01 - 2022

September 30, 2022

Item No.	Item Description	Qty	Unit	Corrections/Modifications/Amendments
	✓ Internal power supply: IPS, 2 accumulators, 12V, 10Ah			
	✓ Degree of protection against electrical shock: user part type B, patient leakage current as per CF according to EN 60601-1			✓ Degree of protection against electrical shock: user part type B, patient leakage current as per CF according to EN 60601-1 or its equivalent
	✓ IP X4: degree of protection against water			✓ IP X4: degree of protection against water or its equivalent
	• Accessories (must be included)			
	✓ Anesthesia screen (1pc)			
	✓ Infusion pole with 4 hooks (1pc)			
	✓ Arm support (1 pair)			
	✓ Body strap (1pc)			
	✓ Leg holder (1pair)			
	Requirements:			
	• Must be compatible with existing fluoroscopy machine (C-Arm) and/or compatible with fracture table attachment.			
	• Aftersales warranty of 1-2 years on parts and labor.			
	• With available technical assistance within the Visayas Region that can respond within 24-48 hours upon request			• With available technical assistance that can respond within 24-48 hours upon request
2	SUCTION MACHINE HD (for OR use)	1	UNIT	
	Brand New			
	Power Source: 220V-240V, 60Hz			
	Heavy duty			
	Mobile-type with casters and brake systems			
	Sturdy and durable body construct			
	Suction Range: 0 – 100kPa			
	Silent or quiet operation: not > 50 dB			
	Suction Flow Rate:40-60 LPM			
	Dial-type suction pressure regulator			
	Dial-type suction gauge			
	With foot switch			
	With overflow safety device			
	Collection container: 2 pieces of not < 2L each.			
	Silicon suction tubing (Reusable / Autoclavable).			
	One suction container remains functional while the other is being cleaned/disinfected.			
	With Fluid Collection Systems to provide integrated protection.			
	Operating and service manual should be supplied			
	With compatible UPS (10-15min back-up time; with UPS brochure.)			
	Compliant with IEC (60601) for Electrical Safety and EN Standard for Equipment			
	ISO 10079-1:2015 or its equivalent			
	Regulatory Standard Compliant			
	CE and ISO approved or its equivalent			
	Warranty: at least 2 years			
	AFTER-SALES REQUIREMENTS:			
	<i>(Please provide certification for each)</i>			
	Should be, or at least have, an authorized Service Center in the Visayas.			Should be, or at least have, an authorized Service Center
	Technicians should comply with the required 72 hours response time for trouble shooting purposes.			



WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City

"PhilHealth Accredited Health Care Provider"

Tel No.: (033) 320 2431 | Fax No.: (033) 3202623 | Email Address: modcenter@wvsu.edu.ph



SUPPLEMENTAL/BID BULLETIN


Project Reference No. IB No. 22-43
 Name of the Project Procurement of Various Medical Equipment-A
 Location of the Project WVSU Medical Center

ADDENDUM NO. 01 - 2022

September 30, 2022

Item No.	Item Description	Qty	Unit	Corrections/Modifications/Amendments
3	CRIB WITH BASSINET	5	UNIT	
	Brand New			
	Stainless steel			
	Baby bassinet is made of Transparent Acrylic Plastic			
	Mattress of bassinet is covered with leather			
	With (4) four lockable wheels			
	With (4) four caster wheels			
	Size: 810 x 460 x 810mm			Size: 810-830 x 460 x 645-810mm
	Plastic basket: 430 x 760 x 255mm			Plastic basket: 220-430x755-760x255-400mm
	After Sales Requirements:			
	72 hours response time for technicians for trouble shooting purpose			
	Available equipment for use if the unit needs to be pulled out while being repaired			
	Provide preventive maintenance agreement after warranty			
	Quality Assurance (ISO, TUV, etc.): ISO 13485; IEC 60601-2-52 compliant			
	Warranty: 1-2 years warranty			
4	SUCTION MACHINE	10	UNIT	
	Portable			
	Power supply: AC 220V-240V 50/60Hz			
	Power: ≤ 90 VA			Power: 45-90 VA
	Max negative pressure value: ≥0.075MPa			
	Pumping Rate: ≥18L /min			
	Noise: ≤ 65 db(A)			
	Reservoir volume: 1000ml /1 liter, 1 piece			Reservoir volume: 1000ml-1200ml / 1-1.2liter, 1 piece
	Oil free lubrication pump			
	Used as an aspirator in the operating room or home patient care			
	Net weight: 3-4.5 kg			
	With initial calibration certificate			
	1-2 years warranty			

For guidance and information of all concerned.


MARY EUGENE C. ORILLO
 Chairperson
 Bids and Awards Committee

Received by the Bidder:

Printed Name & Signature

Name of Bidder/Company

Date: _____

Section VII. Technical Specifications

IB No. UMC 22-43 Procurement of Various Medical Equipment-A (Revised as indicated in the Bid Bulletin)

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	Procurement of Various Medical Equipment-A	Qty	Unit	Brand	Model	Statement of Compliance
1	OPERATING ROOM TABLE, Brand New	2	UNIT			
	<ul style="list-style-type: none"> • Table Structure ✓ Table top: 4- 5 ✓ Frame: Made of stainless- steel structure ✓ Base clearance that permits under the base space of the foot ✓ Emergency Unlock- handle below the chassis, quick release handle in emergency case ✓ Electrol conductive wheels ✓ Electro-motorized floor locks with 4- 5 stamps, casters lock and unlock. • Table Measurement ✓ Table length: 1900-2100mm ✓ Table width: 530mm- 590mm ✓ Table height (without pad): 600mm- 1000mm ✓ Weight capacity: 450- 500kg ✓ Translucent, C-Arm compatible • Table Movements ✓ Trendelenburg/Reverse- 30°/30° ✓ Tilt: ±25° ✓ Longitudinal slide overall: 200-270mm ✓ Joint back section: 75°-80° ✓ Joint leg section: 85°-90° ✓ Head section length: 300-340mm ✓ Back & Seat Section Length: 1000-1010mm ✓ Leg section length: 700-740mm ✓ One button zero position for motorized leveling of all sections ✓ Flex up and down for kidney break positioning can be adjusted via cable remote control and column keypad • Power System ✓ External power supply: 220-240V, 60Hz ✓ Power input: max. 450 VA ✓ Built-in battery, automatic battery charging ✓ AC/DC auto switch ✓ Internal power supply: IPS, 2 accumulators, 12V, 10Ah ✓ Degree of protection against electrical shock: user part type B, patient leakage current as per CF according to EN 60601-1 or its equivalent ✓ IP X4: degree of protection against water or its equivalent 					

Section VII. Technical Specifications

IB No. UMC 22-43 Procurement of Various Medical Equipment-A (Revised as indicated in the Bid Bulletin)

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	Procurement of Various Medical Equipment-A	Qty	Unit	Brand	Model	Statement of Compliance
	<ul style="list-style-type: none"> • Accessories (must be included) 					
	✓ Anesthesia screen (1pc)					
	✓ Infusion pole with 4 hooks (1pc)					
	✓ Arm support (1 pair)					
	✓ Body strap (1pc)					
	✓ Leg holder (1pair)					
	Requirements:					
	• Must be compatible with existing fluoroscopy machine (C-Arm) and/or compatible with fracture table attachment.					
	• Aftersales warranty of 1-2 years on parts and labor.					
	• With available technical assistance that can respond within 24-48 hours upon request					
2	SUCTION MACHINE HD (for OR use)	1	UNIT			
	Specifications:					
	Brand New					
	Power Source: 220V-240V, 60Hz					
	Heavy duty					
	Mobile-type with casters and brake systems					
	Sturdy and durable body construct					
	Suction Range: 0 – 100kPa					
	Silent or quiet operation: not > 50 dB					
	Suction Flow Rate:40-60 LPM					
	Dial-type suction pressure regulator					
	Dial-type suction gauge					
	With foot switch					
	With overflow safety device					
	Collection container: 2 pieces of not < 2L each.					
	Silicon suction tubing (Reusable / Autoclavable).					
	One suction container remains functional while the other is being cleaned/disinfected.					
	With Fluid Collection Systems to provide integrated protection.					

Section VII. Technical Specifications

IB No. UMC 22-43 Procurement of Various Medical Equipment-A (Revised as indicated in the Bid Bulletin)

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	Procurement of Various Medical Equipment-A	Qty	Unit	Brand	Model	Statement of Compliance
	Operating and service manual should be supplied					
	With compatible UPS (10-15min back-up time; with UPS brochure.)					
	Compliant with IEC (60601) for Electrical Safety and EN Standard for Equipment					
	ISO 10079-1:2015 or its equivalent					
	Regulatory Standard Compliant					
	CE and ISO approved or its equivalent					
	Warranty: at least 2 years					
	AFTER-SALES REQUIREMENTS:					
	<i>(Please provide certification for each)</i>					
	Should be, or at least have, an authorized Service Center					
	Technicians should comply with the required 72 hours response time for trouble shooting purposes.					
3	CRIB WITH BASSINET	5	UNIT			
	Brand New					
	Stainless steel					
	Baby bassinet is made of Transparent Acrylic Plastic					
	Mattress of bassinet is covered with leather					
	With (4) four lockable wheels					
	With (4) four caster wheels					
	Size: 810-830 x 460 x 645-810mm					
	Plastic basket: 220-430 x 755-760 x 255-400mm					
	After Sales Requirements:					
	72 hours response time for technicians for trouble shooting purpose					
	Available equipment for use if the unit needs to be pulled out while being repaired					
	Provide preventive maintenance agreement after warranty					
	Quality Assurance (ISO, TUV, etc.): ISO 13485: IEC 60601-2-52 compliant					
	Warranty: 1-2 years warranty					

Section VII. Technical Specifications

IB No. UMC 22-43 Procurement of Various Medical Equipment-A (Revised as indicated in the Bid Bulletin)

Item No.	Specification	Statement of Compliance				
		<p><i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>				
	Procurement of Various Medical Equipment-A	Qty	Unit	Brand	Model	Statement of Compliance
4	SUCTION MACHINE	10	UNIT			
	Portable					
	Power supply: AC 220V-240V 50/60Hz					
	Power: 45-90 VA					
	Max negative pressure value: $\geq 0.075\text{MPa}$					
	Pumping Rate: $\geq 18\text{L /min}$					
	Noise: $\leq 65\text{ db(A)}$					
	Reservoir volume: 1000ml-1200ml / 1-1.2 liter , 1 piece					
	Oil free lubrication pump					
	Used as an aspirator in the operating room or home patient care					
	Net weight: 3-4.5 kg					
	With initial calibration certificate					
	1-2 years warranty					

Name of Bidder: _____

Submitted by: _____
(Printed Name and Signature)

Designation: _____

Date: _____