



**SUPPLEMENTAL/BID BULLETIN**

Project Reference No. IB No. 22-45  
 Name of the Project Procurement of Various Medical Equipment for Molecular Laboratory  
 Location of the Project WVSU Medical Center

**ADDENDUM NO. 01 - 2022**

October 24, 2022

This Addendum No. 01 is issued to modify or amend items in the Bidding Documents. Attached is the **Revised Section VII. Technical Specifications (pages 18-21)**. These shall form an integral part of the Bidding Documents.

**A. Bid Opening**

As Specified in the Bidding Documents	Corrections/Modifications/Amendments
Bid opening shall be on <u>November 7, 2022 8:30 A.M.</u>	Bid opening shall be on <u>November 7, 2022 9:00 A.M.</u>

**B. Delivery Period**

As Specified in the Bidding Documents	Corrections/Modifications/Amendments
Thirty (30) Calendar Days	<b>Forty-Five (45) Calendar Days</b>

**C. Section VIII. Checklist of Technical and Financial Documents**

**Class "B" Documents** – if applicable, a duly signed joint venture agreement (JVA) in case the joint venture is already in existence; or duly notarized statements from all the potential joint venture partners stating that they will enter into and abide by the provisions of the JVA in the instance that the bid is successful. **If NOT applicable, please provide a certification or statement duly signed by the authorized representative that JVA is NOT APPLICABLE.**

**D. Section VII. Technical Specifications**

Item No.	Item Description	Qty	Unit	Corrections/Modifications/Amendments
2	<b>ULTRA LOW FREEZER BANKING / INVENTORY RACK ORGANIZER</b>	20	Pc	
	<ul style="list-style-type: none"> <li>Inventory rack</li> <li>External Dimension: 140x561x331 mm</li> <li>Box type capacity</li> <li>Quantity</li> </ul>			<ul style="list-style-type: none"> <li><b>Quantity (delete)</b></li> </ul>
3	<b>2-20 µL SINGLE CHANNEL MICROPIPETTOR</b>	1	Unit	
	<ul style="list-style-type: none"> <li>Volume range: 2-20µL</li> <li>Fully visible volume display</li> <li>Allows selecting desired volume and locking the indicator</li> <li>Finger hook</li> <li>Fully autoclavable body</li> </ul>			<ul style="list-style-type: none"> <li>Fully autoclavable body, <b>assembled</b></li> </ul>
4	<b>0.5-10 µL SINGLE CHANNEL MICROPIPETTOR</b>	1	Unit	
	<ul style="list-style-type: none"> <li>Volume range: 0.5-10µL</li> <li>Fully visible volume display</li> <li>Allows selecting desired volume and locking the indicator</li> <li>Finger hook</li> <li>Fully autoclavable body</li> </ul>			<ul style="list-style-type: none"> <li>Fully autoclavable body, <b>assembled</b></li> </ul>
5	<b>20-200 µL SINGLE CHANNEL MICROPIPETTOR</b>	2	Unit	
	<ul style="list-style-type: none"> <li>Volume range: 20-200µL</li> <li>Fully visible volume display</li> <li>Allows selecting desired volume and locking the indicator</li> <li>Finger hook</li> <li>Fully autoclavable body</li> </ul>			<ul style="list-style-type: none"> <li>Fully autoclavable body, <b>assembled</b></li> </ul>



SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 22-45
Name of the Project Procurement of Various Medical Equipment for Molecular Laboratory
Location of the Project WVSU Medical Center

ADDENDUM NO. 01 - 2022

October 24, 2022

D. Section VII. Technical Specifications

Table with 5 columns: Item No., Item Description, Qty, Unit, Corrections/Modifications/Amendments. Contains specifications for various micro pipettors (Items 6-10) and an After Sales / Warranty section.

For guidance and information of all concerned.

MARY EUGENE C. ORILLO
Chairperson
Bids and Awards Committee

Received by the Bidder:

Printed Name & Signature

Name of Bidder/Company

Date:

# Section VII. Technical Specifications

**IB No. UMC 22-45 Procurement of Various Medical Equipment for Molecular Laboratory  
(Revised as indicated in the Bid Bulletin No. 01-2022)**

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
Procurement of Various Medical Equipment for Molecular Laboratory		Qty	Unit	Brand	Model	Statement of Compliance
1	<b>UV LIGHT FOR ROOMS</b>	5	Unit			
	<ul style="list-style-type: none"> <li>• Kills and deactivates 99.9% of viruses, bacteria, molds, and dust mites through UVC (light) disinfection technology</li> <li>• For sterilizing rooms or enclosed areas at least 20sqm in 15 minutes</li> <li>• Comes with a remote and power-on delay for 0 exposure to UVC rays</li> <li>• With time options</li> <li>• Turns off automatically once timer is done</li> <li>• Eliminates foul odor caused by mold and bacteria</li> <li>• Equipped with UVC bulb that emits <b>medical-grade</b> UVC light rays</li> <li>• Input Voltage: 220-240V</li> <li>• Frequency: 50/60 Hz</li> </ul>					
2	<b>ULTRA LOW FREEZER BANKING / INVENTORY RACK ORGANIZER</b>	20	Pc			
	<ul style="list-style-type: none"> <li>• Inventory rack</li> <li>• External Dimension: 140x561x331 mm</li> <li>• Box type capacity</li> </ul>					
3	<b>2-20 µL SINGLE CHANNEL MICROPIPETTOR</b>	1	Unit			
	<ul style="list-style-type: none"> <li>• Volume range: 2-20µL</li> <li>• Fully visible volume display</li> <li>• Allows selecting desired volume and locking the indicator</li> <li>• Finger hook</li> <li>• Fully autoclavable body, <b>assembled</b></li> </ul>					
4	<b>0.5-10 µL SINGLE CHANNEL MICROPIPETTOR</b>	1	Unit			
	<ul style="list-style-type: none"> <li>• Volume range: 0.5-10µL</li> <li>• Fully visible volume display</li> <li>• Allows selecting desired volume and locking the indicator</li> </ul>					

# Section VII. Technical Specifications

**IB No. UMC 22-45 Procurement of Various Medical Equipment for Molecular Laboratory  
(Revised as indicated in the Bid Bulletin No. 01-2022)**

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
Procurement of Various Medical Equipment for Molecular Laboratory		Qty	Unit	Brand	Model	Statement of Compliance
	• Finger hook					
	• Fully autoclavable body, <b>assembled</b>					
5	<b>20-200 µL SINGLE CHANNEL MICROPIPETTOR</b>	2	Unit			
	• Volume range: 20-200µL					
	• Fully visible volume display					
	• Allows selecting desired volume and locking the indicator					
	• Finger hook					
	• Fully autoclavable body, <b>assembled</b>					
6	<b>100-1000 µL SINGLE CHANNEL MICROPIPETTOR</b>	3	Unit			
	• Volume range: 100-1000µL					
	• Fully visible volume display					
	• Allows selecting desired volume and locking the indicator					
	• Finger hook					
	• Fully autoclavable body, <b>assembled</b>					
7	<b>2-20 µL MULTICHANNEL MICROPIPETTOR</b>	1	Unit			
	• 8 channels					
	• <b>Adjustable volume</b>					
	• Pipette Type: Manual air displacement					
	• Ejector type: Plastic					
	• Volume: 2-20µL					
8	<b>0.5-10 µL MULTICHANNEL MICROPIPETTOR</b>	1	Unit			
	• 8 channels					
	• <b>Adjustable volume</b>					
	• Pipette Type: Manual air displacement					
	• Ejector type: Plastic					
	• Volume: 0.5-10µL					
9	<b>20-200 µL MULTICHANNEL MICROPIPETTOR</b>	2	Unit			
	• 8 channels					
	• <b>Adjustable volume</b>					

# Section VII. Technical Specifications

**IB No. UMC 22-45 Procurement of Various Medical Equipment for Molecular Laboratory  
(Revised as indicated in the Bid Bulletin No. 01-2022)**

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	Procurement of Various Medical Equipment for Molecular Laboratory	Qty	Unit	Brand	Model	Statement of Compliance
	<ul style="list-style-type: none"> <li>• Pipette Type: Manual air displacement</li> </ul>					
	<ul style="list-style-type: none"> <li>• Ejector type: Plastic</li> </ul>					
	<ul style="list-style-type: none"> <li>• Volume: 20-200µL</li> </ul>					
10	<b>100-1000 µL MULTICHANNEL MICROPIPETTOR</b>	1	Unit			
	<ul style="list-style-type: none"> <li>• 8 channels</li> </ul>					
	<ul style="list-style-type: none"> <li>• <b>Adjustable volume</b></li> </ul>					
	<ul style="list-style-type: none"> <li>• Pipette Type: Manual air displacement</li> </ul>					
	<ul style="list-style-type: none"> <li>• Ejector type: Plastic</li> </ul>					
	<ul style="list-style-type: none"> <li>• Volume: 100-1000µL</li> </ul>					
11	<b>VORTEX MIXER</b>	2	Unit			
	<ul style="list-style-type: none"> <li>• Orbit range: 4mm</li> </ul>					
	<ul style="list-style-type: none"> <li>• Speed range: 0-3000 rpm</li> </ul>					
	<ul style="list-style-type: none"> <li>• Temperature Resistance: 5-40°C</li> </ul>					
	<ul style="list-style-type: none"> <li>• Wattage: 30W</li> </ul>					
	<ul style="list-style-type: none"> <li>• Power: 220V</li> </ul>					
	<ul style="list-style-type: none"> <li>• Includes universal rack</li> </ul>					
12	<b>MINIFUGE / STRIPFUGE / PLATEFUGE</b>	2	Unit			
	<ul style="list-style-type: none"> <li>• Speed at rated voltage:                             <ul style="list-style-type: none"> <li>○ Approx. 6400 RPM (under empty load)</li> <li>○ Approx. 4400 RPM (under full load)</li> </ul> </li> </ul>					
	<ul style="list-style-type: none"> <li>• Relative centrifuge force (at rated voltage)                             <ul style="list-style-type: none"> <li>○ Approx. 2046 g</li> </ul> </li> </ul>					
	<ul style="list-style-type: none"> <li>• Start/stop button switch</li> </ul>					
	<ul style="list-style-type: none"> <li>• Safely interlock switch</li> </ul>					
	<ul style="list-style-type: none"> <li>• Indicator light: Red LED indicator</li> </ul>					
	<ul style="list-style-type: none"> <li>• Rotors: For 8 x 1.5/2.0 ml tubes &amp; 36 x 0.2mL Strip tubes</li> </ul>					
	<ul style="list-style-type: none"> <li>• Power: DC 12V, 1.5A</li> </ul>					
	<ul style="list-style-type: none"> <li>• Operating condition:                             <ul style="list-style-type: none"> <li>○ Temperature: 2-40 C</li> </ul> </li> </ul>					

# Section VII. Technical Specifications

**IB No. UMC 22-45 Procurement of Various Medical Equipment for Molecular Laboratory  
(Revised as indicated in the Bid Bulletin No. 01-2022)**

Item No.	Specification	Statement of Compliance				
		<i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i>				
	Procurement of Various Medical Equipment for Molecular Laboratory	Qty	Unit	Brand	Model	Statement of Compliance
	o Humidity: 80% max					
	o Altitude: up to 2000m					
	• Includes 0.5 and 0.2 mL Adapters					
	• Designed for quick spin down from the wall and cap of micro-centrifuge tubes					
	• Supplied with interchangeable angle rotor for 8 capacity of 1.5/2.0mL tubes and strip rotor for 36 x 0.2 mL strip tubes					
	<b>After Sales / Warranty: (provide certification)</b>					
	• Service engineer can respond to call within 24 hours					
	• 2 years on parts and services, including Preventive Maintenance and Calibration					
	---nothing follows---					

Name of Bidder: \_\_\_\_\_

Submitted by: \_\_\_\_\_  
(Printed Name and Signature)

Designation: \_\_\_\_\_

Date: \_\_\_\_\_