



SUPPLEMENTAL/BID BULLETIN

Project Reference No. **IB No. 22-46**
 Name of the Project **Procurement of Various Laboratory Supplies for Molecular Laboratory**
 Location of the Project **WVSU Medical Center**

ADDENDUM NO. 01 - 2022

October 24, 2022

This Addendum No. 01 is issued to modify or amend items in the Bidding Documents. Attached is the **Revised Section VII. Technical Specifications (pages 18-20)**. These shall form an integral part of the Bidding Documents.

A. Section VIII. Checklist of Technical and Financial Documents

Class "B" Documents – if applicable, a duly signed joint venture agreement (JVA) in case the joint venture is already in existence; or duly notarized statements from all the potential joint venture partners stating that they will enter into and abide by the provisions of the JVA in the instance that the bid is successful. **If NOT applicable, please provide a certification or statement duly signed by the authorized representative that JVA is NOT APPLICABLE.**

B. Section VII. Technical Specifications

Item No.	Item Description	Qty	Unit	Corrections/Modifications/Amendments
1	RNASE DECONTAMINATION SOLUTION	50	Bot	
	<ul style="list-style-type: none"> Volume: at least 300mL 			<ul style="list-style-type: none"> Volume: at least 300mL or 15,000mL total volume
	<ul style="list-style-type: none"> Composition (please indicate) 			<ul style="list-style-type: none"> Composition (please indicate) (delete)
	<ul style="list-style-type: none"> pH range (please indicate) 			<ul style="list-style-type: none"> pH range (please indicate) (delete)
	<ul style="list-style-type: none"> Eliminates nuclease and nucleic acid contamination 			
	<ul style="list-style-type: none"> Safe for apparatus, bench-tops, glassware and plasticware 			
	<ul style="list-style-type: none"> No mixing, ready to use 			
	<ul style="list-style-type: none"> Excellent trigger sprayer with superfine mist 			
	<ul style="list-style-type: none"> Leak-proof 			
	<ul style="list-style-type: none"> Storage at room temperature 			
3	REAGENT RESERVOIR	1500	Pc	
	<ul style="list-style-type: none"> 50 mL Capacity 			
	<ul style="list-style-type: none"> Sterile 			<ul style="list-style-type: none"> Autoclavable
	<ul style="list-style-type: none"> V-bottom 			
	<ul style="list-style-type: none"> Disposable 			<ul style="list-style-type: none"> Disposable (delete)
	<ul style="list-style-type: none"> Polystyrene material 			
	<ul style="list-style-type: none"> RNase / DNase Free 			
	<ul style="list-style-type: none"> Storage at room temperature 			
9	PAPER TOWELS	2000	Pack	
	<ul style="list-style-type: none"> 175 pulls/pack 			<ul style="list-style-type: none"> at least 175 pulls/pack, 1 ply
	<ul style="list-style-type: none"> Interfolded 			
	<ul style="list-style-type: none"> 100% virgin pulp 			<ul style="list-style-type: none"> 100% virgin pulp, white



WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City

"PhilHealth Accredited Health Care Provider"

Tel No.: (033) 320 2431 | Fax No.: (033) 3202623 | Email Address: medcenter@wvsu.edu.ph



SUPPLEMENTAL/BID BULLETIN

Project Reference No. **IB No. 22-46**
 Name of the Project **Procurement of Various Laboratory Supplies for Molecular Laboratory**
 Location of the Project **WVSU Medical Center**

ADDENDUM NO. 01 - 2022

October 24, 2022

B. Section VII. Technical Specifications

Item No.	Item Description	Qty	Unit	Corrections/Modifications/Amendments
11	MEGABOX PLASTIC STORAGE	10	Pc	
	• 95 L			
	• Durable			
	• For storage			
	• Snap lock storage box			
	• Stackable			
				• Color: white

For guidance and information of all concerned.

MARY EUGENE C. ORILLO
Chairperson
Bids and Awards Committee



Received by the Bidder:

Printed Name & Signature

Name of Bidder/Company

Date: _____

Section VII. Technical Specifications

**IB No. UMC 22-46 Procurement of Various Laboratory Supplies for Molecular Laboratory
(Revised as indicated in the Bid Bulletin No. 01-2022)**

Item No.	Specification	Statement of Compliance			
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]			
	Procurement of Various Laboratory Supplies for Molecular Laboratory	Qty	Unit	Brand	Statement of Compliance
1	RNASE DECONTAMINATION SOLUTION	50	Bot		
	<ul style="list-style-type: none"> • Volume: at least 300mL or 15,000mL total volume • Eliminates nuclease and nucleic acid contamination • Safe for apparatus, bench-tops, glassware and plasticware • No mixing, ready to use • Excellent trigger sprayer with superfine mist • Leak-proof • Storage at room temperature 				
2	DNASE/RNASE-FREE DISTILLED WATER	50	Bot		
	<ul style="list-style-type: none"> • Grade: DNase-Free, Molecular Biology Grade, Protease-Free, RNase-Free • pH: 6 to 8 • Volume: 500mL • Membrane-filtered • Storage at room temperature 				
3	REAGENT RESERVOIR	1500	Pc		
	<ul style="list-style-type: none"> • 50 mL Capacity • Autoclavable • V-bottom • Polystyrene material • RNase / DNase Free • Storage at room temperature 				
4	LOW-LINT, ANTI-STATIC WIPES	51	Pack		
	<ul style="list-style-type: none"> • Absorbent, low lint • 280 pulls • Non-Static • Easily picks up liquid and dust without scratching delicate surfaces • Made of 100% virgin wood fiber 				

Section VII. Technical Specifications

**IB No. UMC 22-46 Procurement of Various Laboratory Supplies for Molecular Laboratory
(Revised as indicated in the Bid Bulletin No. 01-2022)**

Item No.	Specification	Statement of Compliance			
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]			
Procurement of Various Laboratory Supplies for Molecular Laboratory		Qty	Unit	Brand	Statement of Compliance
5	TEST TUBE RACK	20	Pc		
	<ul style="list-style-type: none"> • 18mm diameter slots • 50 slots per rack 				
6	STERILE GRADUATED CONICAL TUBES	1500	Pc		
	<ul style="list-style-type: none"> • 50 mL Capacity with graduation • Sterile with screw cap lid • RNase / DNase Free 				
7	CRYOGENIC BOX	200	Pc		
	<ul style="list-style-type: none"> • 81 slots (9x9) • Can fit 2.0 mL cryogenic tubes • Plastic material 				
8	ACCESSORY SET, COMPATIBLE WITH KINGFISHER	28	Set		
	<ul style="list-style-type: none"> • Package unit: 1 set • Handling: magnetic separation • Sterile; RNase / DNase Free • Scope of delivery: 4 pcs Deep-well plates for KingFisher, 1 pc Elution plate, 1 pc Tip comb for KingFisher 				
	Note: provide sample for validation				
9	PAPER TOWELS	2000	Pack		
	<ul style="list-style-type: none"> • at least 175 pulls/pack, 1 ply • Interfolded • 100% virgin pulp, white 				
10	70% ETHYL ALCOHOL WITH MOISTURIZER	50	Gal		
	<ul style="list-style-type: none"> • 1 gallon • 70% ethyl alcohol • With moisturizer • FDA-approved 				

Section VII. Technical Specifications

**IB No. UMC 22-46 Procurement of Various Laboratory Supplies for Molecular Laboratory
(Revised as indicated in the Bid Bulletin No. 01-2022)**

Item No.	Specification	Statement of Compliance			
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]			
	Procurement of Various Laboratory Supplies for Molecular Laboratory	Qty	Unit	Brand	Statement of Compliance
11	MEGABOX PLASTIC STORAGE	10	Pc		
	• 95 L				
	• Durable				
	• For storage				
	• Snap lock storage box				
	• Stackable				
	• Color: white				
12	MINIMUM / MAXIMUM DIGITAL THERMOMETER WITH HUMIDITY	20	Unit		
	• Large LCD Display				
	• Temperature and Humidity Gauge Display				
	• With Certificate of Calibration				
13	UV BOX	2	Unit		
	• Sterilization box				
	• Portable				
	• Programmable timer				
	• 8 Watts germicidal lamp with ozone				
	• 220 Vac outlet				
	• 200nm to 280nm wavelength				
	• Full aluminum interior				
	• At least 10 Liters				
	---nothing follows---				

Name of Bidder: _____

Submitted by: _____
(Printed Name and Signature)

Designation: _____

Date: _____