



SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 23-37
Name of the Project Procurement of Brand New Low Temperature Sterilization Machine
Location of the Project WVSU Medical Center

ADDENDUM NO. 01 - 2023

July 3, 2023

This Addendum No. 01 is issued to modify or amend items in the Bidding Documents. Attached is the Revised Section VII. Technical Specifications (pages 18-19). These shall form an integral part of the Bidding Documents.

A. Section VII. Technical Specifications

Table with 8 columns: Item No., Item Description, Qty., Unit, Item No., Item Description, Qty., Unit. It details technical specifications for a Low Temperature Sterilization Machine, including requirements for certification from tertiary accredited hospitals.

For guidance and information of all concerned.

PATRICIA P. TUPAS, D.D.M.
Vice Chairperson
Bids and Awards Committee

Received by the Bidder:

Printed Name & Signature

Name of Bidder/Company

Date:

# Section VII. Technical Specifications

**IB No. UMC 23-37 Procurement of Brand New Low Temperature Sterilization Machine  
(Revised as indicated in the Bid Bulletin No. 01-2023)**

Item No.	Specification	Statement of Compliance				
		<p><i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>				
	Procurement of Brand New Low Temperature Sterilization Machine	Qty	Unit	Brand	Model	Statement of Compliance
1	<b>Low Temperature Sterilization Machine, BRAND NEW</b>	1	unit			
	<b>Technical Specifications:</b>					
	▪ Must be either hydrogen peroxide or formaldehyde sterilization, 20-25 cycles/bottle or its equivalent					
	▪ Chamber type: Single or Double door system, rectangular, made of stainless steel (SUS) material or its equivalent					
	▪ Chamber Dimensions: 310mm-450mm(W) x 310mm-420mm(H) x 750mm-950mm(D); Over-all dimensions: 670mm-920mm(W) x 1550mm-1890mm(H) x 920mm-1200mm(D)					
	▪ Total Volume: 120L – 150L; Usable: 117L– 118L (410mm-425mm(W) x 390mm-405mm(H) x 715mm-725mm(D)					
	▪ 3 cycle time phases: Quick- 30-35±5 minutes, Standard- 40-45±5 minutes, Special- 50-55±5 minutes					
	▪ Weight: 290kg- 600kg					
	▪ Electrical/Power supply: 220-230V, 60Hz, 1Phase					
	▪ Cycle temperature: 50-55±5°C					
	▪ Rapid warmup & Dry System, Convenient and Safe Sterilizing Agent					
	▪ 7-10" full touch LCD, Easy monitoring the cycle information					
	▪ Plug and play, USB history memory, Built in Thermal Printer, Printout with actual cycle information					
	▪ Monitoring system; Auto Interlock & Open System.					
	▪ Installation Requirements: Front: 95-100cm, Rear: 95-100cm, left side, Right side: 8-10cm, Built-in Wheels to provide mobility					
	▪ Sterility Assurance Level: 10-6					
	▪ Control: Microprocessor & Windows CE Embedded					
	▪ Cycle Information: USB, Ethernet (optional)					
	▪ PM Cycle: Automatic Alarm & Setup					

# Section VII. Technical Specifications

**IB No. UMC 23-37 Procurement of Brand New Low Temperature Sterilization Machine  
(Revised as indicated in the Bid Bulletin No. 01-2023)**

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	<b>Procurement of Brand New Low Temperature Sterilization Machine</b>	<b>Qty</b>	<b>Unit</b>	<b>Brand</b>	<b>Model</b>	<b>Statement of Compliance</b>
	▪ Emergency Stop (Front), Operator ID Login, Self-diagnosis, Multi-Language, RFID Coding system					
	▪ Directive/Standard: ISO 13485:2016; Medical Devices Directive 93/42/EEC EC Declaration of conformity					
	▪ <b>Accessories:</b>					
	✓ Sterilizing agent compatible with machine					
	✓ Chemical indicator strip					
	✓ Chemical indicator tape					
	✓ Super rapid biological indicator of at least 15-30 minutes					
	✓ Rapid incubator machine of at least 15-30 minutes reading time					
	✓ Sterilization pouch, printer paper and sealing machine					
	✓ AVR of at least 5KVA					
	<b>Requirements - Please provide certification for the following:</b>					
	▪ Must be at least 2 years of free services on preventive maintenance and calibration, warranty on parts and services after the acceptance by the end-user.					
	▪ With available technical assistance from medical device specialist in the locality (Western Visayas Region) that can respond within 24-48 hours upon request.					
	▪ <b>Must provide certification of good quality and performance from at least three (3) tertiary accredited hospitals in the Philippines of the same unit.</b>					
	---nothing follows---					

Name of Bidder: \_\_\_\_\_

Submitted by: \_\_\_\_\_  
(Printed Name and Signature)

Designation: \_\_\_\_\_

Date: \_\_\_\_\_