



**SUPPLEMENTAL/BID BULLETIN**

Project Reference No. **IB No. 23-44**  
Name of the Project **Procurement of Hospital Bed with Bedside Table and Overbed Table**  
Location of the Project **WVSU Medical Center**

**ADDENDUM NO. 01 - 2023**

July 25, 2023

This Addendum No. 01 is issued to modify or amend items in the Bidding Documents. Attached is the **Revised Section VII. Technical Specifications (pages 18-19)**. These shall form an integral part of the Bidding Documents.

**A. Delivery Period**

As Specified in the Bidding Documents	Corrections/Modifications/Amendments
Forty-Five (45) Calendar Days	<b>Sixty (60) Calendar Days</b>

**B. Section VII. Technical Specifications**

Item No.	Item Description	Qty	Unit	Corrections/Modifications/Amendments
1	<b>HOSPITAL BED WITH BEDSIDE TABLE AND OVERBED TABLE</b>	58	set	
	- Working load: 150 kg. – 225 kg.			- Working load: 150 kg. – <b>250 kg.</b>
	- Plastic molded headboard and footboard construction			- <b>ABS material</b> headboard and footboard construction; <b>single piece with button to lay down, six grade aluminum side rail</b>
	- With IV Pole			- With IV Pole, <b>four IV pole holes, detachable telescopic IV pole with foldable four hooks</b>
	- Bed Length: 85"-90", bed height low position 15"-20", high position 25"-32" underbed clearances 5"-5", head articulation angle 0 degree – 45 degree			- Bed Length: <b>82"-90"</b> , bed height low position 15"-20", high position 25"-32" underbed clearances <b>10"-15"</b> , head articulation <b>angle 0 degree – 90 degree</b>
	- Bedside Table size: L460-600mm x W470-600mm x H700-800mm			- Bedside <b>Cabinet</b> size: L460-600mm x W <b>380-800</b> mm x H700-800mm
	- With 4 lockable caster wheel 1.5"-5"			- With 4 <b>wheels 1.5"-5"</b> , <b>at least 2 lockable caster wheels</b>
	- Made of premium ABS material			- Made of premium ABS material, <b>with at least one drawer and one shelf with cover</b>
	- Overbed Tabletop size: L700-750mm x W400-450mm with adjustable height			- Overbed Tabletop size: L700- <b>820</b> mm x W <b>390-450</b> mm with adjustable height
	- Steel construction powder-coated paint finish, fine welded			- Steel construction powder-coated paint finish, fine welded/ <b>ABS material</b>

For guidance and information of all concerned.

*Cesar Q. Tronco*  
CESAR Q. TRONCO, R.N. *TRONCO*  
Chairperson  
Bids and Awards Committee

Received by the Bidder:

\_\_\_\_\_  
Printed Name & Signature

\_\_\_\_\_  
Name of Bidder/Company

Date: \_\_\_\_\_



# Section VII. Technical Specifications

**IB No. UMC 23-44 Procurement of Hospital Bed with Bedside Table and Overbed Table  
(Revised as indicated in the Bid Bulletin No. 01-2023)**

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	Procurement of Hospital Bed with Bedside Table and Overbed Table	Qty	Unit	Brand	Model	Statement of Compliance
1	<b>HOSPITAL BED WITH BEDSIDE TABLE AND OVERBED TABLE</b>	58	set			
	- Brand new					
	- Manual/ Mechanical Type					
	- Working load: 150 kg. – 250 kg.					
	- Steel construction powder-coated paint finish, fine welded/ stainless steel bed frame					
	- <b>ABS material</b> headboard and footboard construction; <b>single piece with button to lay down, six grade aluminum side rail</b>					
	- 3 crank handle control					
	- Polyurethane foam with leatherette cover mattress 6-8 inches thick					
	- With IV Pole, <b>four IV pole holes, detachable telescopic IV pole with foldable four hooks</b>					
	- Bed Length: <b>82"-90"</b> , bed height low position 15"-20", high position 25"-32" underbed clearances <b>10"-15"</b> , head articulation <b>angle 0 degree – 90 degree</b>					
	- Lockable caster wheel 4.5" – 5"					
	- <b>Bedside Cabinet size: L460-600mm x W380-800mm x H700-800mm</b>					
	- With 4 <b>wheels 1.5"-5"</b> , <b>at least 2 lockable caster wheels</b>					
	- Made of premium ABS material, <b>with at least one drawer and one shelf with cover</b>					
	- <b>Overbed Tabletop size: L700-820mm x W390-450mm</b> with adjustable height					
	- Steel construction powder-coated paint finish, fine welded/ <b>ABS material</b>					
	- Each unit shall be assembled upon delivery					

# Section VII. Technical Specifications

**IB No. UMC 23-44 Procurement of Hospital Bed with Bedside Table and Overbed Table  
(Revised as indicated in the Bid Bulletin No. 01-2023)**

Item No.	Specification	Statement of Compliance				
		<p><i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>				
	<b>Procurement of Hospital Bed with Bedside Table and Overbed Table</b>	<b>Qty</b>	<b>Unit</b>	<b>Brand</b>	<b>Model</b>	<b>Statement of Compliance</b>
	<b>After sales requirements. Please provide certification for the following:</b>					
	1. 72 hours response time for technicians for troubleshooting purpose					
	2. Available equipment for use if the unit needs to be pulled out while being repaired					
	3. Supplier ensures availability and immediate supply of spare parts for the duration of the warranty period					
	4. Provide preventive maintenance agreement after the warranty					
	5. Quality Assurance (ISO, TUV, etc.): ISO 13485					
	6. Warranty: 1 – 2 years warranty					
	---nothing follows---					

Name of Bidder: \_\_\_\_\_

Submitted by: \_\_\_\_\_  
(Printed Name and Signature)

Designation: \_\_\_\_\_

Date: \_\_\_\_\_