



SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 23-47
Name of the Project Procurement of Various Medical Equipment
Location of the Project WVSU Medical Center

ADDENDUM NO. 01 - 2023

August 3, 2023

This Addendum No. 01 is issued to modify or amend items in the Bidding Documents. Attached is the Revised Section VII. Technical Specifications (pages 18-27). These shall form an integral part of the Bidding Documents.

A. Section VII. Technical Specifications

Table with 5 columns: Item No., Item Description, Qty, Unit, Corrections/Modifications/Amendments. Row 1: DELIVERY TABLE (ELECTRIC), 1 unit. Includes detailed specifications like power source, dimensions, and accessories.



WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City

"PhilHealth Accredited Health Care Provider"

Tel No.: (033) 320 2431 | Fax No.: (033) 3202623 | Email Address: medcenter@wvsu.edu.ph



SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 23-47
 Name of the Project Procurement of Various Medical Equipment
 Location of the Project WVSU Medical Center

ADDENDUM NO. 01 - 2023

August 3, 2023

A. Section VII. Technical Specifications

Item No.	Item Description	Qty	Unit	Corrections/Modifications/Amendments
2	DELIVERY TABLE (MANUAL)	1	unit	
	Specifications:			
	Brand new			
	Stainless steel frame			
	Up and down motion is hydraulic; with foot control			
	Mobile, with 4 - 5 casters with brake			
	Tilting Capabilities for:			
	o Normal Tilting			
	o Back Section (up & down)			
	o Trendelenburg			
	Load Capacity: 150kg-250kg			
	Dimension:			
	o Length: 1250mm-1800mm			
	o Width: 500mm - 650mm			
	o Height: 700mm-1000mm			
	Accessories:			
	o Arm Support (Padded)			
	o Hand Grips			
	o Detachable Padded Stirrups			
	o Detachable Padded Foot Stirrups			
	o Infusion Rods (optional)			
	o Mattress: detachable; at least 4in; high density foam, with leatherette cover, water proof			
	o Revolving adjustable stool			
	o Pail: stainless; at least 10L capacity			
	Warranty: 2 years			<i>(delete)</i> Warranty: 2 years
	Regulatory Standard Compliant			
	ISO Standards Compliant: (ISO 13485:2016) or its equivalent.			ISO Standards Compliant: (ISO 13485:2016) or its equivalent. <i>(please provide copy of certificate)</i>
	AFTER-SALES REQUIREMENTS:			
	(Provide certification for each)			
	Technicians should comply with the required 72 hours response time for trouble shooting purposes.			
	If the unit needs to be pulled-out while being repaired, company should have an available back-up equipment for use at unit.			
	Must provide Preventive Maintenance schedule within the warranty period.			
				<i>(add)</i> Warranty: At least 2 years on parts and service



SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 23-47
Name of the Project Procurement of Various Medical Equipment
Location of the Project WVSU Medical Center

ADDENDUM NO. 01 - 2023

August 3, 2023

A. Section VII. Technical Specifications

Table with 5 columns: Item No., Item Description, Qty, Unit, Corrections/Modifications/Amendments. Row 1: 3, STRETCHER, 20, units. Includes detailed specifications like 'Brand new', 'Heavy Duty', 'Length: 1900-2000 mm', etc.



WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City

"PhilHealth Accredited Health Care Provider"

Tel No.: (033) 320 2431 | Fax No.: (033) 3202623 | Email Address: medcenter@wvsu.edu.ph



SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 23-47
 Name of the Project Procurement of Various Medical Equipment
 Location of the Project WVSU Medical Center

ADDENDUM NO. 01 - 2023

August 3, 2023

A. Section VII. Technical Specifications

Item No.	Item Description	Qty	Unit	Corrections/Modifications/Amendments
4	PEDIATRIC BED / COMFY BED	15	units	
	- Brand new			
	- Heavy Duty			
	- Manual			
	- Construction - Combined between steel pipes welding process.			
	- Mattress Platform - Made of steel sheet with phosphate coating perforated with emboss style and equipped with mattress stopper at sides which made of steel rod			
	- Head Panel and Foot Panel - Made of steel pipes assembled with the welding process.			
	- With Polyurethane foam with leatherette cover mattress 2-4 inches thick			
	- Safety Side Guard - Sliding type and lockable, and made of steel welding process			
	- Position - Flat with Back Raise			
	- I.V. Pole - Made of stainless steel and adjustable height with locking screw, can be positioned in four different corners of the bed			
	- Castor wheel: four 5" swivel castor (brakes & swivel locks)			- Castor wheel: four 5" swivel castor (brakes & swivel locks)
	- Coating - Epoxy powder coating finish			
	- Max. Load - 120 Kg.			
	- Color: Blue			
				(add) - Quality Assurance (ISO, TUV, etc.): ISO 13485 (please provide copy of certificate)
	After-sales requirements. Please provide certification for the following:			
	- 72 hours response time for technicians for troubleshooting purpose			
	- Available equipment for use if the unit needs to be pulled out while being repaired			
	- Supplier ensures availability and immediate supply of spare parts for the duration of the warranty period			
	- Provide preventive maintenance agreement after the warranty			
	- Quality Assurance (ISO, TUV, etc.): ISO 13485:			(delete) - Quality Assurance (ISO, TUV, etc.): ISO 13485:
	- Warranty: 1 – 2 years warranty			



WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City
"PhilHealth Accredited Health Care Provider"

Tel No.: (033) 320 2431 | Fax No.: (033) 3202623 | Email Address: medcenter@wvsu.edu.ph



SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 23-47
Name of the Project Procurement of Various Medical Equipment
Location of the Project WVSU Medical Center

ADDENDUM NO. 01 - 2023

August 3, 2023

A. Section VII. Technical Specifications

Item No.	Item Description	Qty	Unit	Corrections/Modifications/Amendments
5	DEFIBRILLATOR W/ CARDIAC MONITOR	2	units	
	Specifications:			
	· Brand New Unit			· Brand New Unit <i>with trolley</i>
	· Defibrillator with specifications for pediatric and adult patients.			
	· NIBP capable with accessories (BP cuff to accommodate pediatric patients)			
	· SpO2 monitoring with sensor			· SpO2 monitoring with sensor (<i>adult and pedia</i>)
	· ECG monitoring, recording and printing			
	· Different modes available:			
	· Manual Mode, AED Mode and synchronous defibrillation (with infant/child capability)			
	· Convertible external paddles with function buttons for both Pediatric and Adult, charging time 5-8 seconds			
	· Sensor detection for pad/lead placement			
	· Supply Unit: Power Supply Main Line: 220 V/240V			· Supply Unit: Power Supply Main Line: 220 V/240V <i>or Autovolt</i>
				(<i>add</i>) · Regulatory Standard Compliance ISO-14385-2003:2016; CE123 (please provide copy of certificate)
	· Terms and Conditions:			· Terms and Conditions (Please provide certification):
	· Covers a 12-month warranty			
	· Regulatory Standard Compliance ISO-14385-2003:2016; CE123			(<i>delete</i>) · Regulatory Standard Compliance ISO-14385-2003:2016; CE123
	· Free preventive maintenance: Should have a quarterly preventive maintenance during the warranty period.			
	· Should have an Authorized service center in the Visayas			
	· Technicians should comply with the required 72hours response time for trouble shooting purposes			
	· If the unit needs to be pulled out while being repaired, company should have an available back-up equipment of similar specifications for use			
	· Replacement of unit if defective.			



WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City

"PhilHealth Accredited Health Care Provider"

Tel No.: (033) 320 2431 | Fax No.: (033) 3202623 | Email Address: medcenter@wvsu.edu.ph



SUPPLEMENTAL/BID BULLETIN

Project Reference No. **IB No. 23-47**
 Name of the Project **Procurement of Various Medical Equipment**
 Location of the Project **WVSU Medical Center**

ADDENDUM NO. 01 - 2023

August 3, 2023

A. Section VII. Technical Specifications

Item No.	Item Description	Qty	Unit	Corrections/Modifications/Amendments
6	PHOTOTHERAPY WITH BASSINET	2	units	
	Specifications:			
	· Brand New Unit			
	· Blue light with top and bottom light source, lamp duration at least 60,000 hours			
	· Height adjustable (non-sectional)			
	· Provision of Eye shade protector			
	· Timer with digital Display			
	· Bassinet: Stable, clear scratch free frame, with comfortable hand hold around the rim, removable			· Bassinet: Stable, clear scratch free frame, with comfortable hand hold around the rim, removable
	· Supply Unit: Power Supply Main Line: 220 V/240V			
				(add) · Regulatory Standard Compliance ISO-14385-2003:2016; CE123 (please provide copy of certificate)
	· Terms and Conditions:			· Terms and Conditions (Please provide certification):
	· Covers a 12-month warranty			
	· Regulatory Standard Compliance ISO-14385-2003:2016; CE123			(delete) · Regulatory Standard Compliance ISO-14385-2003:2016; CE123
	· Free preventive maintenance: Should have a quarterly preventive maintenance during the warranty period.			
	· Should have an Authorized service center in the Visayas			
	· Technicians should comply with the required 72hours response time for trouble shooting purposes			
	· If the unit needs to be pulled out while being repaired, company should have an available back-up equipment of similar specifications for use			
	· Replacement of unit if defective.			



WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City

"PhilHealth Accredited Health Care Provider"

Tel No.: (033) 320 2431 | Fax No.: (033) 3202623 | Email Address: medcenter@wvsu.edu.ph



SUPPLEMENTAL/BID BULLETIN

Project Reference No. **IB No. 23-47**
 Name of the Project **Procurement of Various Medical Equipment**
 Location of the Project **WVSU Medical Center**

ADDENDUM NO. 01 - 2023

August 3, 2023

A. Section VII. Technical Specifications

Item No.	Item Description	Qty	Unit	Corrections/Modifications/Amendments
7	INCUBATOR	2	units	
	Specifications:			
	· Brand new unit			
	· Skin and Air Temperature Monitoring			
	· Manual and automatic thermoregulation and humidity settings			
	· Mattress with external control knob for adjustment of mattress height and tilt, provision for x-ray cassette tray			· Mattress with external/electric (provided with back-up battery of at least 12-24 hrs running time) control knob for adjustment of mattress height and tilt, provision for x-ray cassette tray
	· Oxygen Control			
	· Utility Shelf and IV pole			
	· Audio and Visual Alarm System			
	· Digital Weighing scale			
	· Two-sided hinged access panel			
	· LCD/LED Setting Display show skin/air temperature and humidity			
	· Data storage and Trend monitoring of weight gain and loss for 7 days			
	· Storage Cabinets			
	· Condensation Management System			
	· Kangaroo Care Mode			· Kangaroo Care Mode or its equivalent
	· 4 casters with foot brakes			
	· Supply Unit: Power Supply Main Line: 220 V/240V			
				(add) · Regulatory Standard Compliance ISO-14385-2003:2016; CE123 (please provide copy of certificate)
	· Terms and Conditions:			· Terms and Conditions (Please provide certification):
	· Covers a 12-month warranty			
	· Regulatory Standard Compliance ISO-14385-2003:2016; CE123			(delete) · Regulatory Standard Compliance ISO-14385-2003:2016; CE123
	· Free preventive maintenance: Should have a quarterly preventive maintenance during the warranty period.			
	· Should have an Authorized service center in the Visayas			
	· Technicians should comply with the required 72hours response time for trouble shooting purposes			
	· If the unit needs to be pulled out while being repaired, company should have an available back-up equipment of similar specifications for use			
	· Replacement of unit if defective.			

For guidance and information of all concerned.

CESAR Q. TRONCO, R.N.

Chairperson

Bids and Awards Committee

Received by the Bidder:

Printed Name & Signature

Name of Bidder/Company

Date: _____

Section VII. Technical Specifications

**IB No. UMC 23-47 Procurement of Various Medical Equipment
(Revised as indicated in the Bid Bulletin No. 01-2023)**

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	Procurement of Various Medical Equipment	Qty	Unit	Brand	Model	Statement of Compliance
1	DELIVERY TABLE (ELECTRIC)	1	unit			
	Specifications:					
	Brand new					
	Power Source: 220-240V, 50/60Hz					
	Stainless steel frame and columns					
	Up and down motion is hydraulic; with hand/foot control					
	Mobile, with 4 - 5 casters with brake					
	Tilting Capabilities for:					
	o Normal Tilting					
	o Back Section (up & down)					
	o Trendelenburg					
	Load Capacity: 150kg-250kg					
	Built-in battery for unhampered operation					
	Dimension:					
	o Length: 1250mm-1800mm					
	o Width: 500mm - 650mm					
	o Height: 700mm-1000mm					
	Accessories:					
	o Arm Support (Padded)					
	o Hand Grips					
	o Detachable Padded Stirrups					
	o Detachable Padded Foot Stirrups					
	o Infusion Rods (optional)					
	o Mattress: at least 3in; polyurethane foam, with leatherette, water-proof cover					
	o Revolving adjustable stool					
	o Pail/Basin: stainless; at least 8L capacity					
	Regulatory Standard Compliant					
	ISO Standards Compliant: (ISO 13485:2016) or its equivalent. (please provide copy of certificate)					

Section VII. Technical Specifications

**IB No. UMC 23-47 Procurement of Various Medical Equipment
(Revised as indicated in the Bid Bulletin No. 01-2023)**

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	Procurement of Various Medical Equipment	Qty	Unit	Brand	Model	Statement of Compliance
	Compliant with IEC (60601) for Electrical Safety and EN Standard for Equipment.					
	AFTER-SALES REQUIREMENTS:					
	(Provide certification for each.)					
	Supplier must provide UPS with built-in AVR.					
	Technicians should comply with the required 72 hours response time for trouble shooting purposes.					
	If the unit needs to be pulled-out while being repaired, company should have an available back-up equipment for use at unit.					
	Must provide Preventive Maintenance & Calibration schedule within the warranty period.					
	Warranty: At least 2 years on parts and service					
2	DELIVERY TABLE (MANUAL)	1	unit			
	Specifications:					
	Brand new					
	Stainless steel frame					
	Up and down motion is hydraulic; with foot control					
	Mobile, with 4 - 5 casters with brake					
	Tilting Capabilities for:					
	o Normal Tilting					
	o Back Section (up & down)					
	o Trendelenburg					
	Load Capacity: 150kg-250kg					
	Dimension:					
	o Length: 1250mm-1800mm					
	o Width: 500mm - 650mm					
	o Height: 700mm-1000mm					
	Accessories:					
	o Arm Support (Padded)					
	o Hand Grips					
	o Detachable Padded Stirrups					
	o Detachable Padded Foot Stirrups					

Section VII. Technical Specifications

**IB No. UMC 23-47 Procurement of Various Medical Equipment
(Revised as indicated in the Bid Bulletin No. 01-2023)**

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	Procurement of Various Medical Equipment	Qty	Unit	Brand	Model	Statement of Compliance
	o Infusion Rods (optional)					
	o Mattress: detachable; at least 4in; high density foam, with leatherette cover, water proof					
	o Revolving adjustable stool					
	o Pail: stainless; at least 10L capacity					
	Regulatory Standard Compliant					
	ISO Standards Compliant: (ISO 13485:2016) or its equivalent. (please provide copy of certificate)					
	AFTER-SALES REQUIREMENTS:					
	(Provide certification for each)					
	Technicians should comply with the required 72 hours response time for trouble shooting purposes.					
	If the unit needs to be pulled-out while being repaired, company should have an available back-up equipment for use at unit.					
	Must provide Preventive Maintenance schedule within the warranty period.					
	Warranty: At least 2 years on parts and service					
3	STRETCHER	20	units			
	- Brand new					
	- Heavy Duty					
	- Size:					
	- Length: 1900-2000 mm					
	- Width: 645 – 670 mm					
	- Adjustment Range:					
	- Backrest: 0-80°					
	- High low: 500-900 mm					
	- Materials:					
	- Bed board: integrated molding of polypropylene (PP) resin					
	- Frame: Aluminum Alloy pressing and high-quality steel					
	- Two long Dropdown protective guard rails					
	- With transferring mattress surface material					

Section VII. Technical Specifications

**IB No. UMC 23-47 Procurement of Various Medical Equipment
(Revised as indicated in the Bid Bulletin No. 01-2023)**

Item No.	Specification	Statement of Compliance				
		<p><i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>				
	Procurement of Various Medical Equipment	Qty	Unit	Brand	Model	Statement of Compliance
	- With double-sided caster lockable wheel and 5th wheel system control the direction of moving					
	- Safe working load: 170 – 200 kgs.					
	- With oxygen portable cylinder holder					
	- With 1 pc. Telescopic IV pole, easy to use and store.					
	- With Polyurethane foam with leatherette cover mattress 3-5 inches thick					
	- Quality Assurance (ISO, TUV, etc.): ISO 13485 <i>(please provide copy of certificate)</i>					
	After-sales requirements. Please provide certification for the following:					
	- 72 hours response time for technicians for troubleshooting purpose					
	- Available equipment for use if the unit needs to be pulled out while being repaired					
	- Supplier ensures availability and immediate supply of spare parts for the duration of the warranty period					
	- Provide preventive maintenance agreement after the warranty					
	- Warranty: 1 – 2 years warranty					
4	PEDIATRIC BED / COMFY BED	15	units			
	- Brand new					
	- Heavy Duty					
	- Manual					
	- Construction - Combined between steel pipes welding process.					
	- Mattress Platform - Made of steel sheet with phosphate coating perforated with emboss style and equipped with mattress stopper at sides which made of steel rod					
	- Head Panel and Foot Panel - Made of steel pipes assembled with the welding process.					
	- With Polyurethane foam with leatherette cover mattress 2-4 inches thick					
	- Safety Side Guard - Sliding type and lockable, and made of steel welding process					
	- Position - Flat with Back Raise					

Section VII. Technical Specifications

**IB No. UMC 23-47 Procurement of Various Medical Equipment
(Revised as indicated in the Bid Bulletin No. 01-2023)**

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	Procurement of Various Medical Equipment	Qty	Unit	Brand	Model	Statement of Compliance
	- I.V. Pole - Made of stainless steel and adjustable height with locking screw, can be positioned in four different corners of the bed					
	- Caster wheel: four 5" swivel castor (brakes & swivel locks)					
	- Coating - Epoxy powder coating finish					
	- Max. Load - 120 Kg.					
	- Color : Blue					
	- Quality Assurance (ISO, TUV, etc.): ISO 13485 (please provide copy of certificate)					
	After-sales requirements. Please provide certification for the following:					
	- 72 hours response time for technicians for troubleshooting purpose					
	- Available equipment for use if the unit needs to be pulled out while being repaired					
	- Supplier ensures availability and immediate supply of spare parts for the duration of the warranty period					
	- Provide preventive maintenance agreement after the warranty					
	- Warranty : 1 – 2 years warranty					
5	DEFIBRILLATOR W/ CARDIAC MONITOR	2	units			
	Specifications:					
	· Brand New Unit with trolley					
	· Defibrillator with specifications for pediatric and adult patients.					
	· NIBP capable with accessories (BP cuff to accommodate pediatric patients)					
	· SpO2 monitoring with sensor (adult and pedia)					
	· ECG monitoring, recording and printing					
	· Different modes available:					
	· Manual Mode, AED Mode and synchronous defibrillation (with infant/child capability)					
	· Convertible external paddles with function buttons for both Pediatric and Adult, charging time 5-8 seconds					
	· Sensor detection for pad/lead placement					
	· Supply Unit: Power Supply Main Line: 220 V/240V or Autovolt					
	· Regulatory Standard Compliance ISO-14385-2003:2016; CE123 (please provide copy of certificate)					

Section VII. Technical Specifications

**IB No. UMC 23-47 Procurement of Various Medical Equipment
(Revised as indicated in the Bid Bulletin No. 01-2023)**

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	Procurement of Various Medical Equipment	Qty	Unit	Brand	Model	Statement of Compliance
	Terms and Conditions (Please provide certification):					
	Covers a 12-month warranty					
	Free preventive maintenance: Should have a quarterly preventive maintenance during the warranty period.					
	Should have an Authorized service center in the Visayas					
	Technicians should comply with the required 72hours response time for trouble shooting purposes					
	If the unit needs to be pulled out while being repaired, company should have an available back-up equipment of similar specifications for use					
	Replacement of unit if defective.					
6	PHOTOTHERAPY WITH BASSINET	2	units			
	Specifications:					
	Brand New Unit					
	Blue light with top and bottom light source, lamp duration at least 60,000 hours					
	Height adjustable (non-sectional)					
	Provision of Eye shade protector					
	Timer with digital Display					
	Bassinets: Stable, clear scratch free frame, with comfortable hand hold around the rim, removable					
	Supply Unit: Power Supply Main Line: 220 V/240V					
	Regulatory Standard Compliance ISO-14385-2003:2016; CE123 (please provide copy of certificate)					
	Terms and Conditions (Please provide certification):					
	Covers a 12-month warranty					
	Free preventive maintenance: Should have a quarterly preventive maintenance during the warranty period.					
	Should have an Authorized service center in the Visayas					

Section VII. Technical Specifications

**IB No. UMC 23-47 Procurement of Various Medical Equipment
(Revised as indicated in the Bid Bulletin No. 01-2023)**

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	Procurement of Various Medical Equipment	Qty	Unit	Brand	Model	Statement of Compliance
	Technicians should comply with the required 72hours response time for trouble shooting purposes					
	If the unit needs to be pulled out while being repaired, company should have an available back-up equipment of similar specifications for use					
	Replacement of unit if defective.					
7	INCUBATOR	2	units			
	Specifications:					
	Brand new unit					
	Skin and Air Temperature Monitoring					
	Manual and automatic thermoregulation and humidity settings					
	Mattress with external/electric (provided with back-up battery of at least 12-24 hrs running time) control knob for adjustment of mattress height and tilt, provision for x-ray cassette tray					
	Oxygen Control					
	Utility Shelf and IV pole					
	Audio and Visual Alarm System					
	Digital Weighing scale					
	Two-sided hinged access panel					
	LCD/LED Setting Display show skin/air temperature and humidity					
	Data storage and Trend monitoring of weight gain and loss for 7 days					
	Storage Cabinets					
	Condensation Management System					
	Kangaroo Care Mode <i>or its equivalent</i>					
	4 casters with foot brakes					
	Supply Unit: Power Supply Main Line: 220 V/240V					
	Regulatory Standard Compliance ISO-14385-2003:2016; CE123 (please provide copy of certificate)					
	Terms and Conditions (Please provide certification):					
	Covers a 12-month warranty					

Section VII. Technical Specifications

**IB No. UMC 23-47 Procurement of Various Medical Equipment
(Revised as indicated in the Bid Bulletin No. 01-2023)**

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-annotated sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	Procurement of Various Medical Equipment	Qty	Unit	Brand	Model	Statement of Compliance
	Free preventive maintenance: Should have a quarterly preventive maintenance during the warranty period.					
	Should have an Authorized service center in the Visayas					
	Technicians should comply with the required 72hours response time for trouble shooting purposes					
	If the unit needs to be pulled out while being repaired, company should have an available back-up equipment of similar specifications for use					
	Replacement of unit if defective.					
8	LEAD APRON SHIELD	5	units			
	• Made of vinyl or nylon					
	• Wrapped around type					
	• 0.5mm front and 0.25mm back lead equivalent					
	• Size: 600mm x 900mm					
	• Not more than 4.3 kilogram					
	• With thyroid shield attached					
	Note:					
	1. The bidder must be a recognized supplier of radiation shielding material in the Philippines (submit certification or documents as proof)					
	2. Warranty of one (1) year after the delivery period					
9	EMERGENCY TROLLEY CART	2	units			
	• Size: L 745 mm or higher: W 470 mm or higher: H 920 mm					
	• Must be made of aluminum alloy column with ABS engineering construction					
	• The guard rail and bumper must be made of stainless steel					
	• Must have 5 drawers (2 small, 2 middle and 1 large)					
	• The small and middle drawers must have a medicine divider for easy organization (vials, syringes, slides, alcohol and test tube)					
	• Must have a pull out writing drawer table					

Section VII. Technical Specifications

**IB No. UMC 23-47 Procurement of Various Medical Equipment
(Revised as indicated in the Bid Bulletin No. 01-2023)**

Item No.	Specification	Statement of Compliance				
		<p><i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>				
	Procurement of Various Medical Equipment	Qty	Unit	Brand	Model	Statement of Compliance
	• Must have a push handle for easy and efficient design					
	• Must have 4 or 5 inches lockable caster wheels 2 of which with brakes					
	• Must have the following emergency accessories:					
	1. IV pole					
	2. Waste bin					
	3. File cassette					
	4. Needle disposal container					
	5. Defibrillator shelf					
	6. Lateral side with sliding working shelf					
	7. Oxygen tank holder					
	8. Power outlet and hooks					
10	PATIENT TRANSPORT TROLLEY (STRETCHER)	2	units			
	• Material must be made of electrostatic spray steel frame and bed flat form					
	• The bed must be combined by four parts (crank type)					
	• Backrest and leg section controlled by hydraulic auxiliary gas spring for easy operation					
	• Size: L 1925 mm or higher: W 590 mm					
	• Height must be adjustable by crank between 520 - 650mm / 810 - 890mm					
	• The side rails must be foldable with double locking system					
	• The main frame must be casted with high quality cold rolled steel and can be automatically pressed					
	• The surface and side rail must be rounded coated with hard plastic material					
	• The bed surface should be composed of two parts; the upper part must be controlled by a gas spring for easy operation and the lower part is fixed					
	• There should be a transfer mattress made of polyurethane foam synthetic leather cover					
	• Must have a central control braking system of about 120 - 130 mm in diameter that is stable and reliable in providing directions during transport					

Section VII. Technical Specifications

**IB No. UMC 23-47 Procurement of Various Medical Equipment
(Revised as indicated in the Bid Bulletin No. 01-2023)**

Item No.	Specification	Statement of Compliance				
		<p>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</p>				
	Procurement of Various Medical Equipment	Qty	Unit	Brand	Model	Statement of Compliance
	• Must have a 5th center wheel system for easy pushing and directional control					
	• Must have a standard oxygen holder mounted horizontally					
	• Must have a rail and IV hanger rod					
	• Must have a detachable thoracic bottle holder with rail hook					
	• Safe working load of at least 215 kilogram					
	• Must be ISO 9001, 13485 and FDA approved					
11	IMAGING PLATE	3	pcs			
	Technical Specifications					
	• Must include the holder and the phosphor material					
	• 14x17 inches size					
	• 6 seconds transfer time					
	• GOS scintillator using photo stimulated luminescence (PSL) phosphor					
	• Must be compatible with existing Image Digitizer, FUJI Prima T2 Brand					
	• Warranty of 1 year					
	---nothing follows---					

Name of Bidder: _____

Submitted by: _____
(Printed Name and Signature)

Designation: _____

Date: _____