



SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 23-58
Name of the Project Procurement of Various Medical Equipment (Re-bidding)
Location of the Project WVSU Medical Center

ADDENDUM NO. 01 - 2023

October 6, 2023

This Addendum No. 01 is issued to modify or amend items in the Bidding Documents. Attached is the Revised Section VII. Technical Specifications (pages 18-21). These shall form an integral part of the Bidding Documents.

A. Section VII. Technical Specifications

Table with 5 columns: Item No., Item Description, Qty, Unit, Corrections/Modifications/Amendments. Row 1: DELIVERY TABLE (ELECTRIC), 1 unit. Includes specifications like Brand new, Power Source: 220-240V, 50/60Hz, and modifications such as (delete) o Normal Tilting and o Back Section (up & down - up to 90 degrees).



SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 23-58
Name of the Project Procurement of Various Medical Equipment (Re-bidding)
Location of the Project WVSU Medical Center

ADDENDUM NO. 01 - 2023

October 6, 2023

A. Section VII. Technical Specifications

Table with 5 columns: Item No., Item Description, Qty, Unit, Corrections/Modifications/Amendments. Row 1: 2, DELIVERY TABLE (MANUAL), 1, unit. Includes specifications like 'Brand new', 'Stainless steel frame', and 'Load Capacity: 150kg-250kg'. Includes amendments like 'Length: 1250mm-2000mm'.



SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 23-58
Name of the Project Procurement of Various Medical Equipment (Re-bidding)
Location of the Project WVSU Medical Center

ADDENDUM NO. 01 - 2023

October 6, 2023

A. Section VII. Technical Specifications

Table with 5 columns: Item No., Item Description, Qty, Unit, Corrections/Modifications/Amendments. Row 3: PEDIATRIC BED / COMFY BED, 15 units. Includes detailed technical specifications and amendments such as 'Head Panel and Foot Panel - Made of detachable ABS plastic' and 'I.V. Pole - Made of stainless steel and adjustable height'.

For guidance and information of all concerned.

Cesar Q. Tronco, R.N.
Chairperson
Bids and Awards Committee

Received by the Bidder:

Printed Name & Signature

Name of Bidder/Company

Date:

Section VII. Technical Specifications

IB No. UMC 23-58 Procurement of Various Medical Equipment (Re-bidding)
(Revised as indicated in the Bid Bulletin No. 01-2023)

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	Procurement of Various Medical Equipment (Re-bidding)	Qty	Unit	Brand	Model	Statement of Compliance
1	DELIVERY TABLE (ELECTRIC)	1	unit			
	Specifications:					
	Brand new					
	Power Source: 220-240V, 50/60Hz					
	Stainless steel frame and columns					
	Up and down motion is hydraulic; with hand/foot control					
	Mobile, with 4 - 5 casters with brake					
	Tilting Capabilities for:					
	o Back Section (<i>up & down - up to 90 degrees</i>)					
	o Trendelenburg					
	Load Capacity: 150kg-250kg					
	Built-in battery for unhampered operation					
	Dimension:					
	o Length: 1250mm-2000mm					
	o Width: 500mm - 650mm					
	o Height: 700mm-1000mm					
	Accessories:					
	o Arm Support (Padded)					
	o Hand Grips					
	o Detachable Padded Stirrups					
	o Detachable Padded Foot Stirrups					
	o Infusion Rods (optional)					
	o Mattress: at least 3in; polyurethane foam, with leatherette, water-proof cover					
	o Revolving adjustable stool					
	o Pail/Basin: stainless; at least 8L capacity					
	Regulatory Standard Compliant					
	ISO Standards Compliant: (ISO 13485:2016) or its equivalent. (please provide copy of certificate)					

Section VII. Technical Specifications

IB No. UMC 23-58 Procurement of Various Medical Equipment (Re-bidding)

(Revised as indicated in the Bid Bulletin No. 01-2023)

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	Procurement of Various Medical Equipment (Re-bidding)	Qty	Unit	Brand	Model	Statement of Compliance
	Compliant with IEC (60601) for Electrical Safety and EN Standard for Equipment.					
	AFTER-SALES REQUIREMENTS:					
	(Provide certification for each.)					
	Supplier must provide UPS with built-in AVR.					
	Technicians should comply with the required 72 hours response time for trouble shooting purposes.					
	If the unit needs to be pulled-out while being repaired, company should have an available back-up equipment for use at unit.					
	Must provide Preventive Maintenance & Calibration schedule within the warranty period.					
	Warranty: At least 2 years on parts and service					
2	DELIVERY TABLE (MANUAL)	1	unit			
	Specifications:					
	Brand new					
	Stainless steel frame					
	Up and down motion is hydraulic; with foot control					
	Mobile, with 4 - 5 casters with brake					
	Tilting Capabilities for:					
	o Back Section (<i>up & down - up to 90 degrees</i>)					
	o Trendelenburg					
	Load Capacity: 150kg-250kg					
	Dimension:					
	o Length: 1250mm-2000mm					
	o Width: 500mm - 650mm					
	o Height: 700mm-1000mm					
	Accessories:					
	o Arm Support (Padded)					
	o Hand Grips					
	o Detachable Padded Stirrups					
	o Detachable Padded Foot Stirrups					

Section VII. Technical Specifications

IB No. UMC 23-58 Procurement of Various Medical Equipment (Re-bidding)

(Revised as indicated in the Bid Bulletin No. 01-2023)

Item No.	Specification	Statement of Compliance				
		<p><i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>				
	Procurement of Various Medical Equipment (Re-bidding)	Qty	Unit	Brand	Model	Statement of Compliance
	o Infusion Rods (optional)					
	o Mattress: detachable;at least 4in;high density foam, with leatherette cover, water proof					
	o Revolving adjustable stool					
	o Pail: stainless; at least 10L capacity					
	Regulatory Standard Compliant					
	ISO Standards Compliant: (ISO 13485:2016) or its equivalent. (please provide copy of certificate)					
	AFTER-SALES REQUIREMENTS:					
	(Provide certification for each)					
	Technicians should comply with the required 72 hours response time for trouble shooting purposes.					
	If the unit needs to be pulled-out while being repaired, company should have an available back-up equipment for use at unit.					
	Must provide Preventive Maintenance schedule within the warranty period.					
	Warranty: At least 2 years on parts and service					
3	PEDIATRIC BED / COMFY BED	15	units			
	- Brand new					
	- Heavy Duty					
	- Manual					
	- Construction - Combined between steel pipes welding process.					
	- Mattress Platform - Made of steel sheet with phosphate coating perforated with emboss style and equipped with mattress stopper at sides which made of steel rod					
	- Head Panel and Foot Panel - Made of detachable ABS plastic					
	- With Polyurethane foam with leatherette cover mattress 2-4 inches thick					
	- Safety Side Guard - Sliding type and lockable, and made of steel welding process					
	- Position - Flat with Back Raise					
	- I.V. Pole - Made of stainless steel and adjustable height with locking screw, can be positioned in four different corners of the bed -Telescopic with four pig tail					

Section VII. Technical Specifications

IB No. UMC 23-58 Procurement of Various Medical Equipment (Re-bidding)

(Revised as indicated in the Bid Bulletin No. 01-2023)

Item No.	Specification	Statement of Compliance				
		<p><i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>				
	Procurement of Various Medical Equipment (Re-bidding)	Qty	Unit	Brand	Model	Statement of Compliance
	- Caster wheel: four 5" swivel castor (brakes & swivel locks)					
	- Coating - Epoxy powder coating finish					
	- Max. Load - 120 Kg.					
	- Quality Assurance (ISO, TUV, etc.): ISO 13485 (please provide copy of certificate)					
	After-sales requirements. Please provide certification for the following:					
	- 72 hours response time for technicians for troubleshooting purpose					
	- Available equipment for use if the unit needs to be pulled out while being repaired					
	- Supplier ensures availability and immediate supply of spare parts for the duration of the warranty period					
	- Provide preventive maintenance agreement after the warranty					
	- Warranty: 1 – 2 years warranty					
4	IMAGING PLATE	3	pcs			
	Technical Specifications					
	• Must include the holder and the phosphor material					
	• 14x17 inches size					
	• 6 seconds transfer time					
	• GOS scintillator using photo stimulated luminescence (PSL) phosphor					
	• Must be compatible with existing Image Digitizer; FUJI Prima T2 Brand					
	• Warranty of 1 year					
	---nothing follows---					

Name of Bidder: _____

Submitted by: _____
(Printed Name and Signature)

Designation: _____

Date: _____