



SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 23-60
Name of the Project Procurement of Various Laboratory Equipment
Location of the Project WVSU Medical Center

ADDENDUM NO. 01 - 2023
October 9, 2023

This Addendum No. 01 is issued to modify or amend items in the Bidding Documents. Attached are the Revised Section VII. Technical Specifications (pages 18-21). These shall form an integral part of the Bidding Documents.

A. Invitation to Bid

Table with 2 columns: As indicated in the Bidding Documents, Corrections/Modifications/Amendments (in bold & italicized words). Rows 2, 7, 9 detailing delivery and bid opening changes.

B. Section VII. Technical Specifications

Table with 5 columns: Item No., Item Description, Qty, Unit, Corrections/Modifications/Amendments (in bold & italicized words). Rows 1, 2, 4 detailing technical specifications for Rotary Microtome, Pharmaceutical Refrigerator, and Platelet Agitator.



WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City

"PhilHealth Accredited Health Care Provider"

Tel No.: (033) 320 2431 | Fax No.: (033) 3202623 | Email Address: medcenter@wvsu.edu.ph



SUPPLEMENTAL/BID BULLETIN

Project Reference No. **IB No. 23-60**
 Name of the Project **Procurement of Various Laboratory Equipment**
 Location of the Project **WVSU Medical Center**

ADDENDUM NO. 01 - 2023

October 9, 2023

Item No.	Item Description	Qty	Unit	Corrections/Modifications/Amendments (in bold & italicized words)
	2. Provide certification of installation of the unit being offered with good product performance and after sales from at least 3 tertiary hospitals			2. Provide certification of installation of the unit being offered with good product performance and after sales from at least 3 tertiary hospitals
	3. Free training of staff upon installation of unit			3. Free training of staff upon installation of unit
	4. 24-48 hour on site response time by the Service Engineer in case of machine breakdown/malfunction			4. 24-48 hour on site response time by the Service Engineer in case of machine breakdown/malfunction
	5. Certification signifying the availability of spare parts for 5 years			5. Certification signifying the availability of spare parts for 5 years
	6. With Line-Interactive Uninterruptible Power Supply (UPS) with Automatic Voltage Regulator (AVR) & Surge Protection feature compatible to the equipment			6. With Line-Interactive Uninterruptible Power Supply (UPS) with Automatic Voltage Regulator (AVR) & Surge Protection feature compatible to the equipment
	7. ISO/IEC/UL/JIS/BG compliant			7. ISO/IEC/UL/JIS/BG compliant
5	Blood Bank Refrigerator	1	unit	Blood Bank Refrigerator
	Other Requirements:			Other Requirements: (Please provide certificate):
	1. Two years warranty on parts and services, including Calibration and Preventive Maintenance			1. Two years warranty on parts and services, including Calibration and Preventive Maintenance
	2. Provide certification of installation of the unit being offered with good product performance and after sales from at least 3 tertiary hospitals			2. Provide certification of installation of the same brand of unit being offered with good product performance and after sales from at least 3 tertiary hospitals
	3. Free training of staff upon installation of unit			3. Free training of staff upon installation of unit
	4. 24-48 hour on site response time by the Service Engineer in case of machine breakdown/malfunction			4. 24-48 hour on site response time by the Service Engineer in case of machine breakdown/malfunction
	5. Certification signifying the availability of spare parts for 5 years			5. Certification signifying the availability of spare parts for 5 years
	6. With Line-Interactive Uninterruptible Power Supply (UPS) with Automatic Voltage Regulator (AVR) & Surge Protection feature compatible to the equipment			6. With Line-Interactive Uninterruptible Power Supply (UPS) with Automatic Voltage Regulator (AVR) & Surge Protection feature compatible to the equipment
	7. ISO/IEC/UL/JIS/BG compliant			7. ISO/IEC/UL/JIS/BG compliant

For guidance and information of all concerned.


CESAR Q. TRONCO
 Chairperson
 Bids and Awards Committee

Received by the Bidder:

Printed Name & Signature

Name of Bidder/Company

Date: _____

Section VII. Technical Specifications

**IB No. UMC 23-60 Procurement of Various Laboratory Equipment
(Revised as indicated in the Bid Bulletin No. 01-2023)**

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	Procurement of Various Laboratory Equipment	Qty	Unit	Brand	Model	Statement of Compliance
1	Rotary Microtome	1	unit			
	Manual Rotary Microtome					
	Section Thickness Range: 0.5 - 100 µm					
	Trimming Thickness Range: 10 – 60 µm					
	Horizontal Feed Range: 24 – 30 mm.					
	Vertical Stroke length: 58 – 70 mm.					
	Coarse Feed Wheel Turn Direction: Clockwise or counterclockwise					
	Blade holder for high and/or low profile blades					
	Universal cassette clamp with alternatively standard object orientation unit					
	Handwheel lockable					
	With <i>anti-static/removable</i> section waste tray					
	Other Requirements: (Please provide certificate)					
	1. Provide certification with a "Very Satisfactory" product performance and after sales of the same product being offered from at least five (5) tertiary hospitals					
	2. Valid ISO Certificate 13485:2016					
	3. Valid CE Certificate					
	4. Certification signifying the availability of spare parts for five (5) years from purchase date.					
	5. 24 – 48 hour on site response time by the Service Engineer in case of machine breakdown/ malfunction.					
	6. Two-years warranty on parts and services.					
	7. Provide and perform quarterly scheduled preventive maintenance and annual calibration of equipment with certification for two (2) years after installation without cost to the hospital.					
	8. Provide free training of staff upon installation.					
2	Pharmaceutical Refrigerator	1	unit			
	Cabinet Type: Upright, double door					
	Capacity: 1200 liters or more					
	Temperature Range: 2 - 8 °C					
	Ambient Temperature: 16-32 °C					
	Controller: Microprocessor					
	Display: Digital Display					

Section VII. Technical Specifications

**IB No. UMC 23-60 Procurement of Various Laboratory Equipment
(Revised as indicated in the Bid Bulletin No. 01-2023)**

Item No.	Specification	Statement of Compliance			
		Qty	Unit	Brand	Model
	Procurement of Various Laboratory Equipment				Statement of Compliance
	Cooling type: Air cooling				
	Defrost Mode: Automatic				
	Refrigerant: Non- HCFC				
	Construction:				
	Adjustable shelves				
	Lighting: LED				
	Door Lock with Key				
	Casters with break				
	Data logging/ Recording Time				
	Alarm:				
	Temperature: High/low temperature				
	High Ambient Temperature				
	Condenser overheating				
	Electrical: Power failure, Low battery				
	System: Sensor failure, Door Ajar				
	Power Supply (V/Hz): 220V, 60Hz, 1 phase				
	Additional Requirements: (Please provide certificate)				
	Provision of Compatible Automatic Voltage Regulator				
	After Sales Services - Service Engineer can respond to call within 24 hours				
	Warranty - 2 years free Preventive Maintenance & Calibration				
	Warranty - 2 years in parts and services				
	Has a valid ISO 13485 Certification of the Manufacturer or equivalent certification				
	Has a certificate of Authorized Distributorship from the Principal				
3	Agglutination Viewer	2	unit		
	Designed for use in blood typing, cross matching, prothrombin time, Kahn and other agglutinations, flocculation, and serodiagnosis tests.				
	Illuminates test tube contents through the use of a 5-watt bulb from above and a magnifying mirror from below.				
	Easy to use side-mounted ON/OFF switch.				
	Flexible neck allows for easy manipulation of light source.				
	Input: 220/60 (V/Hz). 18w, Single phase				
	Output: 12 VAC 500mA				

Section VII. Technical Specifications

*IB No. UMC 23-60 Procurement of Various Laboratory Equipment
(Revised as indicated in the Bid Bulletin No. 01-2023)*

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	Procurement of Various Laboratory Equipment	Qty	Unit	Brand	Model	Statement of Compliance
4	Platelet Agitator	1	unit			
	Capacity: at least 42 bags					
	Power: 220 V 60Hz					
	Rotational Speed: 70 rpm					
	Weight: at least 25 kg					
	Size: approximately 450*340*360 mm					
	Shelves: 5-7					
	Stainless steel shelves					
	Other Requirements: (Please provide certificate)					
	1. Two years warranty on parts and services, including Calibration and Preventive Maintenance					
	2. Provide certification of installation of the unit being offered with good product performance and after sales from at least 3 tertiary hospitals					
	3. Free training of staff upon installation of unit					
	4. 24-48 hour on site response time by the Service Engineer in case of machine breakdown/malfunction					
	5. Certification signifying the availability of spare parts for 5 years					
	6. With Line-Interactive Uninterruptible Power Supply (UPS) with Automatic Voltage Regulator (AVR) & Surge Protection feature compatible to the equipment					
	7. ISO/IEC/UL/JIS/BG compliant					
5	Blood Bank Refrigerator	1	unit			
	Temp Control: Micro Processor Controller					
	Temperature Range: 2 to 6 °C					
	Temperature Accuracy: 0.1°C					
	Temperature Uniformity: 2°C at 4°C					
	Capacity: at least 290 L					
	Door: Single Door, Tempered Triple Safety Glass					
	Display Type: LED Display					
	Circulation Fan: Direct Cooling, ancillary fan					
	Safety Device: Exclusive Low & High Temp.					
	Protector, Alarm for power failure					
	Power: 220 V 60Hz					
	Shelf type with basket or tray					

Section VII. Technical Specifications

*IB No. UMC 23-60 Procurement of Various Laboratory Equipment
(Revised as indicated in the Bid Bulletin No. 01-2023)*

Item No.	Specification	Statement of Compliance				
		<p><i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>				
	Procurement of Various Laboratory Equipment	Qty	Unit	Brand	Model	Statement of Compliance
	Other Requirements: (Please provide certificate)					
	1. Two years warranty on parts and services, including Calibration and Preventive Maintenance					
	2. Provide certification of installation of the same brand of unit being offered with good product performance and after sales from at least 3 tertiary hospitals					
	3. Free training of staff upon installation of unit					
	4. 24-48 hour on site response time by the Service Engineer in case of machine breakdown/malfunction					
	5. Certification signifying the availability of spare parts for 5 years					
	6. With Line-Interactive Uninterruptible Power Supply (UPS) with Automatic Voltage Regulator (AVR) & Surge Protection feature compatible to the equipment					
	7. ISO/IEC/UL/JIS/BG compliant					
	---Nothing Follows---					

Name of Bidder: _____

Submitted by: _____
(Printed Name and Signature)

Designation: _____

Date: _____