



WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City

"PhilHealth Accredited Health Care Provider"

Tel No.: (033) 320 2431 | Fax No.: (033) 3202623 | Email Address: medcenter@wvsu.edu.ph



SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 23-70

Name of the Project Procurement of Ambulatory Blood Pressure Monitoring with Analysis and Licensed Work Station with 3 units Adult Cuff

Location of the Project WVSU Medical Center

ADDENDUM NO. 01 - 2023

December 11, 2023

This Addendum No. 01 is issued to modify or amend items in the Bidding Documents. Attached are the Revised Section III. Bid Data Sheet (page 12), and Section VII. Technical Specifications (pages 18-20). These shall form an integral part of the Bidding Documents.

A. Instruction to Bidders

As Specified in the Bidding Documents	Corrections/Modifications/Amendments
10.2. The Bidder's SLCC as indicated in ITB Clause 5.3 should have been completed within <u>2 years</u> prior to the deadline for the submission and receipt of bids.	10.2. The Bidder's SLCC as indicated in ITB Clause 5.3 should have been completed within <u>3 years</u> prior to the deadline for the submission and receipt of bids.

B. Bid Data Sheet

As Specified in the Bidding Documents	Corrections/Modifications/Amendments
ITB Clause	
5.3	b. Completed within <u>3 years</u> prior to the deadline for the submission and receipt of bids.

C. Section VII. Technical Specifications

As Specified in the Bidding Documents	Corrections/Modifications/Amendments
Item Description	Item Description
AMBULATORY BLOOD PRESSURE MONITORING WITH ANALYSIS AND LICENSED WORK STATION WITH 3 UNITS ADULT CUFF	
TECHNICAL PARAMETERS:	
Maximum recording time: 72 hours	
Measurement method: oscillometric, automatic null balancing	
Measuring range of heart rate: 35-220 beats/min	
Maximum number of measurements: 500 to 600	
Systolic blood pressure: 70 to 270 mmHg	
Diastolic blood pressure: 40-155 mmHg	
Maximum cuff pressure: 300 mmHg	
Duration of measurement: 30 seconds	
Standard measurement intervals: Day phase every 15 minutes, night phase every 30 minutes, if required creation of additional phases Intervals adjustable between 5 and 90 minutes setting of individual profiles	
3 cuff sizes : Standard 24 x 32 cm. XL 32-40 cm. XXL 38-50 cm.	
Data transfer : via optical cable/infrared data transfer	



SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 23-70
Name of the Project Procurement of Ambulatory Blood Pressure Monitoring with Analysis and Licensed Work Station with 3 units Adult Cuff
Location of the Project WVSU Medical Center

ADDENDUM NO. 01 - 2023
December 11, 2023

C. Section VII. Technical Specifications

Table with 2 columns: As Specified in the Bidding Documents and Corrections/Modifications/Amendments. It lists technical specifications for blood pressure monitoring equipment, including item descriptions, dimensions, weight, standards, and additional requirements like warranty and training.



SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 23-70
Name of the Project Procurement of Ambulatory Blood Pressure Monitoring with Analysis and Licensed Work Station with 3 units Adult Cuff
Location of the Project WVSU Medical Center

ADDENDUM NO. 01 - 2023
December 11, 2023

C. Section VII. Technical Specifications

Table with 2 columns: As Specified in the Bidding Documents and Corrections/Modifications/Amendments. Rows include Regulatory Standards, Workstation Requirements, and Inclusions.



WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City

"PhilHealth Accredited Health Care Provider"

Tel No.: (033) 320 2431 | Fax No.: (033) 3202623 | Email Address: medcenter@wvsu.edu.ph



SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 23-70
 Name of the Project Procurement of Ambulatory Blood Pressure Monitoring with Analysis and Licensed Work Station with 3 units Adult Cuff
 Location of the Project WVSU Medical Center

ADDENDUM NO. 01 - 2023

December 11, 2023

C. Section VII. Technical Specifications

As Specified in the Bidding Documents	Corrections/Modifications/Amendments
Item Description	Item Description
	(add) ADDITIONAL REQUIREMENTS:
	(add) 1. Provide certification for the inclusions:
	(add) a. 1 unit isolation transformer
	(add) b. Uninterrupted power supply
	(add) c. 1 unit computer table with swivel chair
	(add) d. USB mouse and keyboard
	(add) e. Power cord
	(add) f. VGA/HDMI/DVI cable, manual, driver software
	(add) 2. Warranty for workstation: at least 1 year free on parts and labor

For guidance and information of all concerned.


CESAR Q. TRONCO, R.N.
 Chairperson
 Bids and Awards Committee

Received by the Bidder:

Printed Name & Signature

Name of Bidder/Company

Date: _____

Section III. Bid Data Sheet

(Revised as indicated in the Bid Bulletin No. 01-2023)

ITB Clause																
5.3	For this purpose, contracts similar to the Project shall be: <ol style="list-style-type: none"> a. <u>Procurement of Ambulatory Blood Pressure Monitoring.</u> b. Completed within <u>3 years</u> prior to the deadline for the submission and receipt of bids. 															
7.1	N/A															
12	The price of the Goods shall be quoted DDP <u>in Philippines</u> or the applicable International Commercial Terms (INCOTERMS) for this Project.															
14.1	The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts: <ol style="list-style-type: none"> a. The amount of not less than <u>PhP 20,000.00</u> [Indicate the amount equivalent to two percent (2%) of ABC], if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or b. The amount of not less than <u>PhP 50,000.00</u> [Indicate the amount equivalent to five percent (5%) of ABC] if bid security is in Surety Bond. 															
19.3	<p><i>Procurement of Ambulatory Blood Pressure Monitoring with Analysis and Licensed Work Station with 3 units Adult Cuff - P 1,000,000.00</i></p> <table border="1" style="width: 100%; border-collapse: collapse; margin-top: 10px;"> <thead> <tr> <th style="width: 8%;">Item No.</th> <th style="width: 42%;">Description</th> <th style="width: 10%;">Qty.</th> <th style="width: 10%;">Unit</th> <th style="width: 30%;">ABC</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">1</td> <td>AMBULATORY BLOOD PRESSURE MONITORING WITH ANALYSIS AND LICENSED WORK STATION WITH 3 UNITS ADULT CUFF</td> <td style="text-align: center;">1</td> <td style="text-align: center;">set</td> <td style="text-align: right;">1,000,000.00</td> </tr> <tr> <td></td> <td style="text-align: center;">TOTAL ABC FOR THE PROJECT</td> <td></td> <td></td> <td style="text-align: right;">1,000,000.00</td> </tr> </tbody> </table>	Item No.	Description	Qty.	Unit	ABC	1	AMBULATORY BLOOD PRESSURE MONITORING WITH ANALYSIS AND LICENSED WORK STATION WITH 3 UNITS ADULT CUFF	1	set	1,000,000.00		TOTAL ABC FOR THE PROJECT			1,000,000.00
Item No.	Description	Qty.	Unit	ABC												
1	AMBULATORY BLOOD PRESSURE MONITORING WITH ANALYSIS AND LICENSED WORK STATION WITH 3 UNITS ADULT CUFF	1	set	1,000,000.00												
	TOTAL ABC FOR THE PROJECT			1,000,000.00												
20.2	<i>No additional Requirements.</i>															
21.2	<i>No additional contract documents</i>															

Section VII. Technical Specifications

IB No. UMC 23-70 Procurement of Ambulatory Blood Pressure Monitoring with Analysis and Licensed Work Station with 3 units Adult Cuff

(Revised as indicated in the Bid Bulletin No. 01-2023)

Item No.	Specification	Statement of Compliance				
		<p><i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>				
	Procurement of Ambulatory Blood Pressure Monitoring with Analysis and Licensed Work Station with 3 units Adult Cuff	Qty	Unit	Brand	Model	Statement of Compliance
1	AMBULATORY BLOOD PRESSURE MONITORING WITH ANALYSIS AND LICENSED WORK STATION WITH 3 UNITS ADULT CUFF	1	set			
	TECHNICAL PARAMETERS:					
	Maximum recording time: 72 hours					
	Measurement method: oscillometric, automatic null balancing					
	Measuring range of heart rate: 35-220 beats/min					
	Maximum number of measurements: 500 to 600					
	Systolic blood pressure: 70 to 270 mmHg					
	Diastolic blood pressure: 40-155 mmHg					
	Maximum cuff pressure: 300 mmHg					
	Duration of measurement: 30 seconds					
	Standard measurement intervals: Day phase every 15 minutes, night phase every 30 minutes, if required creation of additional phases Intervals adjustable between 5 and 90 minutes setting of individual profiles					
	3 cuff sizes : Standard 24 x 32 cm. XL 32-40 cm. XXL 38-50 cm.					
	Data transfer : via optical cable/infrared data transfer					
	Voltage supply: 3 mignon 1.5 volt, type AA or 3 rechargeable batteries, Ni-MH, 1.2 volt.min. 1500 mAh					
	Dimensions: approx. 98-100mm(L), 66-69mm(W), 26-29mm(H)					
	Weight: 170 grams – 190 grams including Batteries					
	Standards: BHS (British Hypertension Society) AAMI (Association for the Advancement of Medical Instrumentation)					
	ADDITIONAL REQUIREMENTS:					
	1. Provide certification for the inclusions:					
	a. Basic unit					
	b. Pouch					
	c. Belt					

Section VII. Technical Specifications

IB No. UMC 23-70 Procurement of Ambulatory Blood Pressure Monitoring with Analysis and Licensed Work Station with 3 units Adult Cuff

(Revised as indicated in the Bid Bulletin No. 01-2023)

Item No.	Specification	Statement of Compliance				
		<p><i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>				
	Procurement of Ambulatory Blood Pressure Monitoring with Analysis and Licensed Work Station with 3 units Adult Cuff	Qty	Unit	Brand	Model	Statement of Compliance
	d. Optical/USB interface					
	e. 3 cuff sizes: standard, XL and XXL					
	f. Carry case					
	g. Set of accumulators					
	h. Accumulator charger					
	i. Licensed software					
	j. User manual					
	2. Certificate of Distributorship					
	3. Response time for repair – within 48 hours after receipt of complaint					
	4. Warranty for equipment – 2 years on parts and labor including quarterly preventive maintenance and annual calibration					
	5. Delivery and good product certification from at least 1 tertiary hospital (same equipment being offered)					
	6. Certificate to provide 1 week basic training for Cardio Technicians					
	7. Regulatory Standards					
	a. Compliant with IEC (60601) for Electrical safety and EN Standard for equipment					
	b. CE and ISO approved or its equivalent					
	WORKSTATION REQUIREMENTS:					
	1 unit branded desktop computer with printer, all-in-one tank, black and colored, wi-fi ready					
	Technical Requirements:					
	CPU					
	Processor: 4 cores, 6MB Cache, 4.0 GHZ clock speed or higher					
	Memory: 4 GB DDR4 or higher					
	Storage: 1 TB GB					
	Optical drive: DVD writer					

Section VII. Technical Specifications

IB No. UMC 23-70 Procurement of Ambulatory Blood Pressure Monitoring with Analysis and Licensed Work Station with 3 units Adult Cuff

(Revised as indicated in the Bid Bulletin No. 01-2023)

Item No.	Specification	Statement of Compliance				
		<p>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-animated sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</p>				
	Procurement of Ambulatory Blood Pressure Monitoring with Analysis and Licensed Work Station with 3 units Adult Cuff	Qty	Unit	Brand	Model	Statement of Compliance
	Graphics: 2GB or higher					
	Audio: Integrated high-definition					
	Networking: gigabit Ethernet, wireless LAN and Bluetooth					
	Standard I/O ports: USB ports, HDMI, Ethernet port, audio-in/out jacks, card reader					
	Monitor/Screen: 21 to 27 inches LED monitor					
	Operating system: Pre-loaded Licensed windows 10 (professional edition)					
	ADDITIONAL REQUIREMENTS:					
	1. Provide certification for the inclusions:					
	a. 1 unit isolation transformer					
	b. Uninterrupted power supply					
	c. 1 unit computer table with swivel chair					
	d. USB mouse and keyboard					
	e. Power cord					
	f. VGA/HDMI/DVI cable, manual, driver software					
	2. Warranty for workstation: at least 1 year free on parts and labor					

Name of Bidder: _____

Submitted by: _____
(Printed Name and Signature)

Designation: _____

Date: _____