



WEST VISAYAS STATE UNIVERSITY MEDICAL CENTER

E. Lopez St., Jaro, Iloilo City

"PhilHealth Accredited Health Care Provider"

Tel No.: (033) 320 2431 | Fax No.: (033) 3202623 | Email Address: medcenter@wvsu.edu.ph



SUPPLEMENTAL/BID BULLETIN

Project Reference No. **IB No. 24-21**
 Name of the Project **Supply and Delivery of Medical Oxygen Flowmeter and Other Medical Gas Accessories**
 Location of the Project **WVSU Medical Center**

ADDENDUM NO. 01 - 2024

May 15, 2024

This Addendum No. 01 is issued to modify or amend items in the Bidding Documents. Attached are the **Revised Section III. Bid Data Sheet (page 12), Section VI. Schedule of Requirements (page 17), Section VII. Technical Specifications (page 18) and Price Schedule for Goods (page 22)**. These shall form an integral part of the Bidding Documents.

A. Section VII. Technical Specifications

As Specified in the Bidding Documents		Corrections/Modifications/Amendments
Item No.	Item Description	Item Description
1	Medical Oxygen Flowmeter with Humidifier. 0 - 15 LPM capacity, wall type, DISS Connection. Calibrated at 0.35 Mpa, 50 Psi, 20 C. US Standard.	Medical Oxygen Flowmeter with Humidifier. 0 - 15 LPM capacity, wall type, DISS Connection, Calibrated @50-60psi, @20°C. US Standard. 360° swivel inlet tube. For Adult
2	Medical Vacuum Pump, Single Stage, Single Acting, 220V/440V, Single Phase.	Medical Vacuum Pump: 5.0 hp, 220V-240V, 60Hz, 3-phase. 2-stage, single acting, Oil-less, 3-reciprocating air piston Compressor. Maximum flow rate 100-115cfm
3	Medical Oxygen Flowmeter with Humidifier. 0.2 - 10 LPM capacity, wall type, DISS Connection. Calibrated at 0.35 Mpa, 50 Psi, 20 C. US Standard. For Infant use	Medical Oxygen Flowmeter with Humidifier. 0.2 - 10 LPM capacity, wall type, DISS Connection, Calibrated @50-60psi, @20°C. US Standard. 360° swivel inlet tube. For Infant
4	Medical Vacuum Pump Oil, SAE 40, Heavy Duty, Diesel Engine Oil.	Medical Vacuum Pump Oil, SAE 15W40 , Heavy Duty, Diesel Engine Oil.
5	Medical Vacuum Pump Belt. V-Belt, Size; B-65, Oil and Heat Resistant.	
6	Magnetic Contactor for Medical Vacuum Pump 5Hp Electric Motor	Magnetic Contactor for Medical Vacuum Pump 5Hp Electric Motor. 3 pole magnetic contactor 220V Coil/60Hz, AC3, 15Kw/240V, 2-NC Auxiliary Contact, 2-NO Auxiliary Contact, AC1 Rating: 80 amperes
7	Starting and Running Capacitor for Medical Vacuum Pump 5Hp Electric Motor	Starting and Running Capacitor for Medical Vacuum Pump 5Hp Electric Motor, 450µf-500µf, 450/500VAC, 50/60Hz, and 40µf ±5%, 450/500VAC, 50/60Hz.
		(Add)
		CONDITIONS:
		Item # 2
		1. Must be compatible with the existing unit of WVSUMC
		2. One year warranty on Pumps and Motor.
		Item #1 and #3
		1. Check actual mounted fitting for compatibility prior to Bid.
		2. Brand offered must be fit and compatible with the existing outlet.
		3. Conduct On-site calibration of the unit upon delivery.
		4. Submit calibration certificate of oxygen flowmeter w/ humidifier.
		5. One year warranty
		Item #6
		1. Six month warranty

For guidance and information of all concerned.


CESAR Q. TRONCO, R.N.
 Chairperson
 Bids and Awards Committee

Received by the Bidder:

Printed Name & Signature

Name of Bidder/Company

Date: _____

Section III. Bid Data Sheet

(Revised as indicated in the Bid Bulletin No. 01-2024)

ITB Clause																																														
5.3	For this purpose, contracts similar to the Project shall be: <ol style="list-style-type: none"> a. <u>Supply and Delivery of Medical Oxygen Flowmeter and Other Medical Gas Accessories.</u> b. Completed within 2 years prior to the deadline for the submission and receipt of bids. 																																													
7.1	N/A																																													
12	The price of the Goods shall be quoted DDP in Philippines or the applicable International Commercial Terms (INCOTERMS) for this Project.																																													
14.1	The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts: <ol style="list-style-type: none"> a. The amount of not less than PhP 41,392.00 [Indicate the amount equivalent to two percent (2%) of ABC], if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or b. The amount of not less than PhP 103,480.00 [Indicate the amount equivalent to five percent (5%) of ABC] if bid security is in Surety Bond. 																																													
19.3	<p>Supply and Delivery of Medical Oxygen Flowmeter and Other Medical Gas Accessories - P 2,069,600.00</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 8%;">Item No.</th> <th style="width: 45%;">Description</th> <th style="width: 10%;">Qty.</th> <th style="width: 10%;">Unit</th> <th style="width: 27%;">ABC</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">1</td> <td>Medical Oxygen Flowmeter with Humidifier. 0 - 15 LPM capacity, wall type, DISS Connection, Calibrated @50-60psi, @20°C. US Standard. 360° swivel inlet tube. For Adult</td> <td style="text-align: center;">350</td> <td style="text-align: center;">Assy</td> <td style="text-align: right;">1,400,000.00</td> </tr> <tr> <td style="text-align: center;">2</td> <td>Medical Vacuum Pump: 5.0 hp, 220V-240V, 60Hz, 3-phase. 2-stage, single acting, Oil-less, 3-reciprocating air piston Compressor. Maximum flow rate 100-115cfm</td> <td style="text-align: center;">2</td> <td style="text-align: center;">Set</td> <td style="text-align: right;">480,000.00</td> </tr> <tr> <td style="text-align: center;">3</td> <td>Medical Oxygen Flowmeter with Humidifier. 0.2 - 10 LPM capacity, wall type, DISS Connection, Calibrated @50-60psi, @20°C. US Standard. 360° swivel inlet tube. For Infant</td> <td style="text-align: center;">32</td> <td style="text-align: center;">Assy</td> <td style="text-align: right;">128,000.00</td> </tr> <tr> <td style="text-align: center;">4</td> <td>Medical Vacuum Pump Oil, SAE 15W40, Heavy Duty, Diesel Engine Oil.</td> <td style="text-align: center;">4</td> <td style="text-align: center;">Gallon</td> <td style="text-align: right;">3,600.00</td> </tr> <tr> <td style="text-align: center;">5</td> <td>Medical Vacuum Pump Belt. V-Belt, Size; B-65, Oil and Heat Resistant.</td> <td style="text-align: center;">8</td> <td style="text-align: center;">Piece</td> <td style="text-align: right;">6,000.00</td> </tr> <tr> <td style="text-align: center;">6</td> <td>Magnetic Contactor for Medical Vacuum Pump 5Hp Electric Motor. 3 pole magnetic contactor 220V Coil/60Hz, AC3, 15Kw/240V, 2-NC Auxiliary Contact, 2-NO Auxiliary Contact, AC1 Rating: 80 amperes</td> <td style="text-align: center;">2</td> <td style="text-align: center;">Unit</td> <td style="text-align: right;">40,000.00</td> </tr> <tr> <td style="text-align: center;">7</td> <td>Starting and Running Capacitor for Medical Vacuum Pump 5Hp Electric Motor, 450µf-500µf, 450/500VAC, 50/60Hz, and 40µf ±5%,450/500VAC, 50/60Hz.</td> <td style="text-align: center;">4</td> <td style="text-align: center;">Set</td> <td style="text-align: right;">12,000.00</td> </tr> <tr> <td colspan="4" style="text-align: center;">TOTAL ABC FOR THE PROJECT</td> <td style="text-align: right;">2,069,600.00</td> </tr> </tbody> </table>	Item No.	Description	Qty.	Unit	ABC	1	Medical Oxygen Flowmeter with Humidifier. 0 - 15 LPM capacity, wall type, DISS Connection, Calibrated @50-60psi, @20°C. US Standard. 360° swivel inlet tube. For Adult	350	Assy	1,400,000.00	2	Medical Vacuum Pump: 5.0 hp, 220V-240V, 60Hz, 3-phase. 2-stage, single acting, Oil-less, 3-reciprocating air piston Compressor. Maximum flow rate 100-115cfm	2	Set	480,000.00	3	Medical Oxygen Flowmeter with Humidifier. 0.2 - 10 LPM capacity, wall type, DISS Connection, Calibrated @50-60psi, @20°C. US Standard. 360° swivel inlet tube. For Infant	32	Assy	128,000.00	4	Medical Vacuum Pump Oil, SAE 15W40, Heavy Duty, Diesel Engine Oil.	4	Gallon	3,600.00	5	Medical Vacuum Pump Belt. V-Belt, Size; B-65, Oil and Heat Resistant.	8	Piece	6,000.00	6	Magnetic Contactor for Medical Vacuum Pump 5Hp Electric Motor. 3 pole magnetic contactor 220V Coil/60Hz, AC3, 15Kw/240V, 2-NC Auxiliary Contact, 2-NO Auxiliary Contact, AC1 Rating: 80 amperes	2	Unit	40,000.00	7	Starting and Running Capacitor for Medical Vacuum Pump 5Hp Electric Motor, 450µf-500µf, 450/500VAC, 50/60Hz, and 40µf ±5%,450/500VAC, 50/60Hz.	4	Set	12,000.00	TOTAL ABC FOR THE PROJECT				2,069,600.00
Item No.	Description	Qty.	Unit	ABC																																										
1	Medical Oxygen Flowmeter with Humidifier. 0 - 15 LPM capacity, wall type, DISS Connection, Calibrated @50-60psi, @20°C. US Standard. 360° swivel inlet tube. For Adult	350	Assy	1,400,000.00																																										
2	Medical Vacuum Pump: 5.0 hp, 220V-240V, 60Hz, 3-phase. 2-stage, single acting, Oil-less, 3-reciprocating air piston Compressor. Maximum flow rate 100-115cfm	2	Set	480,000.00																																										
3	Medical Oxygen Flowmeter with Humidifier. 0.2 - 10 LPM capacity, wall type, DISS Connection, Calibrated @50-60psi, @20°C. US Standard. 360° swivel inlet tube. For Infant	32	Assy	128,000.00																																										
4	Medical Vacuum Pump Oil, SAE 15W40, Heavy Duty, Diesel Engine Oil.	4	Gallon	3,600.00																																										
5	Medical Vacuum Pump Belt. V-Belt, Size; B-65, Oil and Heat Resistant.	8	Piece	6,000.00																																										
6	Magnetic Contactor for Medical Vacuum Pump 5Hp Electric Motor. 3 pole magnetic contactor 220V Coil/60Hz, AC3, 15Kw/240V, 2-NC Auxiliary Contact, 2-NO Auxiliary Contact, AC1 Rating: 80 amperes	2	Unit	40,000.00																																										
7	Starting and Running Capacitor for Medical Vacuum Pump 5Hp Electric Motor, 450µf-500µf, 450/500VAC, 50/60Hz, and 40µf ±5%,450/500VAC, 50/60Hz.	4	Set	12,000.00																																										
TOTAL ABC FOR THE PROJECT				2,069,600.00																																										
20.2	No additional Requirements.																																													
21.2	No additional contract documents																																													

Section VI. Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

IB No. UMC 24-21 Supply and Delivery of Medical Oxygen Flowmeter and Other Medical Gas Accessories (Revised as indicated in the Bid Bulletin No. 01-2024)

Item No.	Description	Quantity		Total	Delivered, Weeks/Months
1	Medical Oxygen Flowmeter with Humidifier. 0 - 15 LPM capacity, wall type, DISS Connection, Calibrated @50-60psi, @20°C. US Standard. 360° swivel inlet tube. For Adult	350	Assy	350 Assy	
2	Medical Vacuum Pump: 5.0 hp, 220V-240V, 60Hz, 3-phase. 2-stage, single acting, Oil-less, 3-reciprocating air piston Compressor. Maximum flow rate 100-115cfm	2	Set	2 Set	
3	Medical Oxygen Flowmeter with Humidifier. 0.2 - 10 LPM capacity, wall type, DISS Connection, Calibrated @50-60psi, @20°C. US Standard. 360° swivel inlet tube. For Infant	32	Assy	32 Assy	
4	Medical Vacuum Pump Oil, SAE 15W40, Heavy Duty, Diesel Engine Oil.	4	Gallon	4 Gallon	
5	Medical Vacuum Pump Belt. V-Belt, Size; B-65, Oil and Heat Resistant.	8	Piece	8 Piece	
6	Magnetic Contactor for Medical Vacuum Pump 5Hp Electric Motor. 3 pole magnetic contactor 220V Coil/60Hz, AC3, 15Kw/240V, 2-NC Auxiliary Contact, 2-NO Auxiliary Contact, AC1 Rating: 80 amperes	2	Unit	2 Unit	
7	Starting and Running Capacitor for Medical Vacuum Pump 5Hp Electric Motor, 450µf-500µf, 450/500VAC, 50/60Hz, and 40µf ±5%,450/500VAC, 50/60Hz.	4	Set	4 Set	
	---nothing follows---				

Name of Bidder: _____

Submitted by: _____
(Printed Name and Signature)

Designation: _____

Date: _____

Section VII. Technical Specifications

IB No. UMC 24-21 Supply and Delivery of Medical Oxygen Flowmeter and Other Medical Gas Accessories
(Revised as indicated in the Bid Bulletin No. 01-2024)

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	Supply and Delivery of Medical Oxygen Flowmeter and Other Medical Gas Accessories	Qty	Unit	Brand	Model	Statement of Compliance
1	Medical Oxygen Flowmeter with Humidifier. 0 - 15 LPM capacity, wall type, DISS Connection, Calibrated @50-60psi, @20°C. US Standard. 360° swivel inlet tube. For Adult	350	Assy			
2	Medical Vacuum Pump: 5.0 hp, 220V-240V, 60Hz, 3-phase. 2-stage, single acting, Oil-less, 3-reciprocating air piston Compressor. Maximum flow rate 100-115cfm	2	Set			
3	Medical Oxygen Flowmeter with Humidifier. 0.2 - 10 LPM capacity, wall type, DISS Connection, Calibrated @50-60psi, @20°C. US Standard. 360° swivel inlet tube. For Infant	32	Assy			
4	Medical Vacuum Pump Oil, SAE 15W40, Heavy Duty, Diesel Engine Oil.	4	Gallon			
5	Medical Vacuum Pump Belt. V-Belt, Size; B-65, Oil and Heat Resistant.	8	Piece			
6	Magnetic Contactor for Medical Vacuum Pump 5Hp Electric Motor. 3 pole magnetic contactor 220V Coil/60Hz, AC3, 15Kw/240V, 2-NC Auxiliary Contact, 2-NO Auxiliary Contact, AC1 Rating: 80 amperes	2	Unit			
7	Starting and Running Capacitor for Medical Vacuum Pump 5Hp Electric Motor, 450µf-500µf, 450/500VAC, 50/60Hz, and 40µf ±5%,450/500VAC, 50/60Hz.	4	Set			
	CONDITIONS:					
	Item #2					
	1. Must be compatible with the existing unit of WVSUMC					
	2. One year warranty on Pumps and Motor.					
	Item #1 and #3					
	1. Check actual mounted fitting for compatibility prior to Bid.					
	2. Brand offered must be fit and compatible with the existing outlet.					
	3. Conduct On-site calibration of the unit upon delivery.					
	4. Submit calibration certificate of oxygen flowmeter w/ humidifier.					
	5. One year warranty					
	Item #6					
	1. Six month warranty					
	---nothing follows---					

Name of Bidder: _____

Submitted by: _____
(Printed Name and Signature)

Designation: _____

Date: _____

Price Schedule for Goods Offered from Within the Philippines

[shall be submitted with the Bid if bidder is offering goods from within the Philippines]

For Goods Offered from Within the Philippines

Name of Bidder _____

Project ID No. UMC 24-21 Page 1 of 1

(Revised as indicated in the Bid Bulletin No. 01-2024)

1	2	3	4	5	6	7	8	9	10
Item	Description	Country of origin	Quantity	Unit price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if Contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 9) x (col 4)
1	Medical Oxygen Flowmeter with Humidifier. 0 - 15 LPM capacity, wall type, DISS Connection, Calibrated @50-60psi, @20°C. US Standard. 360° swivel inlet tube. For Adult		350 Assy						
2	Medical Vacuum Pump: 5.0 hp, 220V-240V, 60Hz, 3-phase. 2-stage, single acting, Oil-less, 3-reciprocating air piston Compressor. Maximum flow rate 100-115cfm		2 Set						
3	Medical Oxygen Flowmeter with Humidifier. 0.2 - 10 LPM capacity, wall type, DISS Connection, Calibrated @50-60psi, @20°C. US Standard. 360° swivel inlet tube. For Infant		32 Assy						
4	Medical Vacuum Pump Oil, SAE 15W40, Heavy Duty, Diesel Engine Oil.		4 Gallon						
5	Medical Vacuum Pump Belt. V-Belt, Size; B-65, Oil and Heat Resistant.		8 Piece						
6	Magnetic Contactor for Medical Vacuum Pump 5Hp Electric Motor. 3 pole magnetic contactor 220V Coil/60Hz, AC3, 15Kw/240V, 2-NC Auxiliary Contact, 2-NO Auxiliary Contact, AC1 Rating: 80 amperes		2 Unit						
7	Starting and Running Capacitor for Medical Vacuum Pump 5Hp Electric Motor, 450µf-500µf, 450/500VAC, 50/60Hz, and 40µf ±5%, 450/500VAC, 50/60Hz.		4 Set						
	TOTAL BID								

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____