



SUPPLEMENTAL/BID BULLETIN

Project Reference No. IB No. 24-27
Name of the Project Procurement of Brand New Neuro Surgical Drill
Location of the Project WVSU Medical Center

ADDENDUM NO. 01 - 2024
June 13, 2024

This Addendum No. 01 is issued to modify or amend items in the Bidding Documents. Attached are the Revised Section VII. Technical Specifications (pages 18-19). These shall form an integral part of the Bidding Documents.

A. Section VII. Technical Specifications

Table with 2 columns: ORIGINAL AS INDICATED IN THE BIDDING DOCUMENTS and Corrections/Modifications/Amendments (in bold & italicized words). Rows include NeuroSurgical Drill, Control Unit, various handpieces (E, F, G, H, I), and requirements.

For guidance and information of all concerned.

CESAR Q. TRONCO, R.N.
Chairperson
Bids and Awards Committee

For the Chairperson:

PATRICIA P. TUPAS, D.D.M.
Vice Chairperson, Bids and Awards Committee

Received by the Bidder:

Printed Name & Signature

Name of Bidder/Company

Date:

Section VII. Technical Specifications

IB No. UMC 24-27 Procurement of Brand New Neuro Surgical Drill
(Revised as indicated in the Bid Bulletin No. 01-2024)

Item No.	Specification	Statement of Compliance				
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]				
	Procurement of Brand New Neuro Surgical Drill	Qty	Unit	Brand	Model	Statement of Compliance
1	NeuroSurgical Drill, Brand New	1	unit			
	A. Control Unit					
	1. The control unit must have at least 6.5 inch LCD touch screen, a simple design that makes it user-friendly					
	2. Two motors can be connected to the unit simultaneously, allowing for quick and easy switching between motors.					
	3. Power supply: AC 220-230V 50/60Hz					
	4. Dimension: 275- 330mm (W)x265- 300mm (D)x158- 170mm (H)					
	5. Weight: at least 6.9kg.					
	6. Integrated High-performance Irrigation Pump. Adjustable up to maximum of 75ml/min.					
	7. Multi-Selectable panel: Speed Control, Irrigation Control, Reverse Rotation Control, Priming Control					
	8. Easy configuration settings. Automatic recognition for motor handpieces.					
	B. MOTOR HANDPIECE					
	1. Accelerator adjustment (min. 10% - max. 100%)					
	2. Brake adjustment (min. 10% - max. 100%)					
	3. Torque adjustment (min 30% - max. 100%)					
	4. Oscillation adjustment (1-5 times)					
	C. FOOT CONTROL					
	Single-function foot control: 1 pedal with 1 function button, or multi-functional foot control: 1 pedal with 3 functional buttons					
	Speed control: adjustable rotational speed; min. 30% - default – max. 100%. Flush (continuous 100% irrigation when pressed)					
	D. SLIM MOTOR HANDPIECE					
	Speed: max. 80,000 min ⁻¹					
	With spray adaptor					
	E. CRANIOTOME HANDPIECE: must be compatible with Ø3.17mm bur					
	F. MEDIUM DURAGUARD- Bur: Medium Craniotome bur, Medium Router Bur (length:15.4mm, bur: 3-4mm)					
	G. PERFORATOR HANDPIECE for trephination. It delivers stable torque and achieves smooth and powerful cutting performance. Compatible chuck: Hudson Chuck (length:65mm dia: 5.5mm)					
	H. STANDARD ANGLED HANDPIECE- Medium type for bur: Ø2.35mm with irrigation nozzle					
	I. STANDARD STRAIGHT HANDPIECE- Medium type for bur: Ø2.35mm with irrigation nozzle					

Section VII. Technical Specifications

**IB No. UMC 24-27 Procurement of Brand New Neuro Surgical Drill
(Revised as indicated in the Bid Bulletin No. 01-2024)**

Item No.	Specification	Statement of Compliance				
		Qty	Unit	Brand	Model	Statement of Compliance
	Procurement of Brand New Neuro Surgical Drill					
	J. ACCESSORIES/ CONSUMABLES:					
	1. Cleaning and lubricator agent (alcohol-contained lubricant) for handpieces to ensure handpiece performance and durability					
	2. Craniotome blades: at least 50pcs					
	3. Round fluted burrs, Medium (any size compatible for medium handpiece): at least 20pcs					
	4. Diamond Burr, Medium (any size compatible for medium handpiece): at least 10pcs					
	5. Neuro Burr, Medium (any size compatible for medium handpiece): at least 20pcs					
	6. Perforator chuck: atleast 30pcs					
	7. Surgical irrigating tubes, Sterilization case with Metal Lid and sterilization divider and Silicone Mat					
	K. Regulatory Standard Compliant:					
	ISO Certification from Medical Devices- ISO 13485					
	TUV/ DIN EN/ISO 9001:2015					
	REQUIREMENTS:					
	1. Must be at least one (1) year of free services and preventive maintenance and calibration, warranty on parts and services after the acceptance by the end-user.					
	2. With available technical assistance from medical device specialist that can respond within 24 to 48 hrs upon request					
	3. Provide certificate of good quality and product performance from at least 3 tertiary hospitals in the Philippines of the same unit or equivalent					

Name of Bidder: _____

Submitted by: _____
(Printed Name and Signature)

Designation: _____

Date: _____