



West Visayas State University

(Formerly Iloilo Normal School)

Bids and Awards Committee Secretariat Office/ Procurement Section

Luna St., La Paz, Iloilo City 5000

Iloilo, Philippines

* Trunkline: (063) (033) 320-0870 loc1103/1104 * Telefax No.: (033) 320-0879

* Website: www.wvsu.edu.ph * Email Address: bac@wvsu.edu.ph



PHI-18939-2-QM

SUPPLEMENTAL/BID BULLETIN

Project Reference No.	IB No. 20-10-14
Name of the Project	Procurement of Anatomical Models and Hospital Bed
Location of the Project	WVSU - Main Campus

ADDENDUM NO. 2020-04

October 28, 2020

This **Addendum No. 2020-04** is issued to modify the **Section I. Invitation to Bid, Section V. Special Conditions of Contract and Section VII. Technical Specifications and Price Schedule for Goods offered from Abroad.** This shall form an integral part of the Bidding Documents.

I. Section I. Invitation to Bid:

Item No.	Description	Correction/ Modification/ Instruction
9.	<i>Bid opening shall be on <u>November 04, 2020; 2:00 P.M.</u> at the given <u>address below or through electronic submission using a two-factor security procedure consisting of an archive format compression and password protection.</u> Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.</i>	<i>Bid opening shall be on <u>November 04, 2020; 2:00 P.M.</u> at the given <u>address below or through electronic submission using a two-factor security procedure consisting of an archive format compression and password protection with separate password for technical and financial component envelope.</u> Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.</i>
10.	<i>Requiring the Bidders to submit their bids using a two-factor security procedure consisting of an archive format compression and password protection and disclose the password for accessing their respective bid submission only during the actual bid opening.</i>	<i>Requiring the Bidders to submit their bids using a two-factor security procedure consisting of an archive format compression and password protection <u>with separate password for technical and financial component envelope</u> and disclose the password for accessing their respective bid submission only during the actual bid opening.</i>

II. Section V. Special Conditions of Contract:

Item No.	Description	Correction/ Modification/ Instruction
4	The Inspections and tests that will be conducted are: <u>Performance Test</u>	The Inspections and tests that will be conducted are: <u>Performance and Functionality Tests.</u>

III. Section VII. Technical Specifications:

Item No.	Description					Correction/ Modification/ Instruction
VII.	Technical Specifications					Technical Specifications
1,2 and 3	Qty	Unit	Unit Cost	Brand	Model	Omission of columns for Qty, Unit, Unit Cost, Brand and Model
1.	Adult Multipurpose Simulator					Adult Multipurpose Simulator
						Qty: 1
						Unit: unit
						Brand (pls. indicate brand)
						Model (pls. specify model)
	<i>Manikins that can produce physical signs (tachypnea/tachycardia), wheezing, crackles, seizures and feed, physiological signs to monitors (ventricular tachycardia, flutter) or anisocoria etc. They can talk, breath, blink and respond either automatically or manually to physical and pharmacological interventions. This machine may be manipulated by the users depending on the clinical scenarios. Monitors are capable of recording the findings like ECG tracing & breathing patterns (capnographs).</i>					<i>Manikins that can produce physical signs (tachypnea/tachycardia), wheezing, crackles, seizures and feed, physiological signs to monitors (ventricular tachycardia, flutter) or anisocoria etc. They can talk, breath, blink and respond either automatically or manually to physical and pharmacological interventions. This machine may be manipulated by the users depending on the clinical scenarios. Monitors are capable of recording the findings like ECG tracing & breathing patterns (capnographs).</i>



West Visayas State University

(Formerly Iloilo Normal School)

Bids and Awards Committee Secretariat Office/ Procurement Section

Luna St., La Paz, Iloilo City 5000

Iloilo, Philippines

* Trunkline: (063) (033) 320-0870 loc1103/1104 * Telefax No.: (033) 320-0879

* Website: www.wvsu.edu.ph * Email Address: bac@wvsu.edu.ph



PHI-18939-2-QM

Item No.	Description	Correction/ Modification/ Instruction
1.	Adult Multipurpose Simulator	Adult Multipurpose Simulator
	<ul style="list-style-type: none"> • Able to provide auscultation training. • Airway Management. • Provides values for cardiac and pulmonary monitors and simulate critical care setting. This includes PCWP, Arterial line monitor and CVP values. • Basic and advance cardiac life support. • Spirometry and ECG tracings. • Normal and abnormal heart sounds. • Arterial blood gas values. • Vital signs measurement. • Ventilator Settings. 	<ul style="list-style-type: none"> • Able to provide auscultation training. • Airway Management. • Provides values for cardiac and pulmonary monitors and simulate critical care setting. This includes PCWP, Arterial line monitor and CVP values. • Basic and advance cardiac life support. • Spirometry and ECG tracings. • Normal and abnormal heart sounds. • Arterial blood gas values. • Vital signs measurement. • Ventilator Settings.
		Others:
		Bidders shall indicate in its submission a video presentation showing features of the model, including demonstration of the various required features indicated in the Section VII.
		Bidders should include/ submit brochures indicating brand and model offered.
2.	Digital Anatomical Table	Digital Anatomical Table
		Qty: 2
		Unit: unit
		Brand (pls. indicate brand)
		Model (pls. specify model)
		Others:
		Bidders shall indicate in its submission a video presentation showing features of the model, including demonstration of the various required features indicated in the Section VII.
		Bidders should include/ submit brochures indicating brand and model offered.
3.	Hospital Bed	Hospital Bed
		Qty: 2
		Unit: unit
		Brand (pls. indicate brand)
		Model (pls. specify model)
		Others:
		Bidders shall indicate in its submission a video presentation showing features of the model, including demonstration of the various required features indicated in the Section VII.
		Bidders should include/ submit brochures indicating brand and model offered.

IV. Price Schedule for Goods Offered from Abroad

Attachment: as stated.

For guidance and information of all concerned.

JULIUS B. UNДАР

Chairperson, Bids and Awards Committee



West Visayas State University

(Formerly Iloilo Normal School)

Bids and Awards Committee Secretariat Office/ Procurement Section

Luna St., La Paz, Iloilo City 5000

Iloilo, Philippines

* Trunkline: (063) (033) 320-0870 loc1103/1104 * Telefax No.: (033) 320-0879

* Website: www.wvsu.edu.ph * Email Address: bac@wvsu.edu.ph



PHI-18939-2-QM

	<p>Final Destination Gross weight Any special lifting instructions Any special handling instructions Any relevant HAZCHEM classifications</p>
	<p>A packaging list identifying the contents and quantities of the package is to be placed on an accessible point of the outer packaging if practical. If not practical the packaging list is to be placed inside the outer packaging but outside the secondary packaging.</p> <p>Transportation –</p> <p>Where the Supplier is required under Contract to deliver the Goods CIF, CIP, or DDP, transport of the Goods to the port of destination or such other named place of destination in the Philippines, as shall be specified in this Contract, shall be arranged and paid for by the Supplier, and the cost thereof shall be included in the Contract Price.</p> <p>Where the Supplier is required under this Contract to transport the Goods to a specified place of destination within the Philippines, defined as the Project Site, transport to such place of destination in the Philippines, including insurance and storage, as shall be specified in this Contract, shall be arranged by the Supplier, and related costs shall be included in the contract price.</p>
	<p>Where the Supplier is required under Contract to deliver the Goods CIF, CIP or DDP, Goods are to be transported on carriers of Philippine registry. In the event that no carrier of Philippine registry is available, Goods may be shipped by a carrier which is not of Philippine registry provided that the Supplier obtains and presents to the Procuring Entity certification to this effect from the nearest Philippine consulate to the port of dispatch. In the event that carriers of Philippine registry are available but their schedule delays the Supplier in its performance of this Contract the period from when the Goods were first ready for shipment and the actual date of shipment the period of delay will be considered force majeure.</p> <p>The Procuring Entity accepts no liability for the damage of Goods during transit other than those prescribed by INCOTERMS for DDP deliveries. In the case of Goods supplied from within the Philippines or supplied by domestic Suppliers risk and title will not be deemed to have passed to the Procuring Entity until their receipt and final acceptance at the final destination.</p> <p>Intellectual Property Rights –</p> <p>The Supplier shall indemnify the Procuring Entity against all third-party claims of infringement of patent, trademark, or industrial design rights arising from use of the Goods or any part thereof.</p>
2.2	<p>[If partial payment is allowed, state] “The terms of payment shall be as follows: <u>not applicable.</u>”</p>
4	<p>The inspections and tests that will be conducted are: <u>Performance and Functionality Tests.</u></p>



West Visayas State University

(Formerly Iloilo Normal School)

Bids and Awards Committee Secretariat Office/ Procurement Section

Luna St., La Paz, Iloilo City 5000

Iloilo, Philippines

* Trunkline: (063) (033) 320-0870 loc1103/1104 * Telefax No.: (033) 320-0879

* Website: www.wvsu.edu.ph * Email Address: bac@wvsu.edu.ph



PHI-18939-2-QM

Section VII. Technical Specifications

Item No.	Specification	Statement of Compliance <i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i>
PROCUREMENT OF ANATOMICAL MODELS AND HOSPITAL BED		Statement of Compliance
1.	Adult Multipurpose Simulator	
	<i>Qty: 1</i>	
	<i>Unit: unit</i>	
	<i>Brand (pls. indicate brand)</i>	
	<i>Model (pls. specify model)</i>	
	<p>A simulator is defined as device that enables the operator to reproduce or represent under test conditions phenomena likely to occur in actual performance. Fidelity is the extent to which the appearance and behavior of the simulation matches the appearance and behavior of the simulated system. High Fidelity simulators combine part or whole body manikins to carry the intervention with computers that drive the manikins to produce physical signs and physiological signs to monitors. They are usually designed to resemble the reality. They can talk, breath, blink, and respond either automatically or manually to physical and pharmacological interventions.</p>	
	<p>Manikins that can produce physical signs (tachypnea/tachycardia), wheezing, crackles, seizures and physiological signs to monitors (ventricular tachycardia, flutter) or anisocoria etc. They can talk, breath, blink and respond either automatically or manually to physical and pharmacological interventions. This machine may be manipulated by the users depending on the clinical scenarios. Monitors are capable of recording the findings like ECG tracing & breathing patterns (capnographs).</p>	
	<p>They are able to provide the following:</p>	
	<ul style="list-style-type: none"> • Able to provide auscultation training. • Airway Management. • Provides values for cardiac and pulmonary monitors and simulate critical care setting. This includes PCWP, Arterial line monitor and CVP values. • Basic and advance cardiac life support. • ECG tracings. • Normal and abnormal heart sounds. • Arterial blood gas values. • Vital signs measurement. • Ventilator Settings. 	



Section VII. Technical Specifications

Item No.	Specification	Statement of Compliance
		<p><i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>
	PROCUREMENT OF ANATOMICAL MODELS AND HOSPITAL BED	Statement of Compliance
	Feature Highlights	
	<p>General</p> <ul style="list-style-type: none"> • Available in Ethnic Skin tones • Powered from internal rechargeable battery or wall outlet • Simulator receives commands from a wireless tablet PC and Operates at distances up to 300 meters • Use preprogrammed scenarios, modify them, or create your own quickly and easily • Installation and training worldwide 	
	<p>Airway</p> <ul style="list-style-type: none"> • Programmed airway: tongue edema, laryngospasm, and pharyngeal swelling • Multiple upper airway sound synchronized with breathing • Right mainstream intubation • Sensor detect depth of intubation • Placement of conventional airway adjuncts • Endotracheal intubation • Retrograde intubation • View vocal cord with Sellick maneuver • Realistic surgical trachea allows tracheostomy or needle cricothyrotomy 	
	<p>Breathing</p> <ul style="list-style-type: none"> • Control rate and depth of respiration and observe chest rise • Select Independent lung sounds: upper right front and back; upper left front and back; lower right front and back; lower left front and back • Chest rise and lung sounds are synchronized with selectable breathing patterns • CO2 on exhalation (4 levels) using replaceable cartridge mounted inside the simulator • Attach to real mechanical ventilators • Bilateral chest rise and fall • Unilateral chest rise simulates pneumothorax 	



West Visayas State University

(Formerly Iloilo Normal School)

Bids and Awards Committee Secretariat Office/ Procurement Section

Luna St., La Paz, Iloilo City 5000

Iloilo, Philippines

* Trunkline: (063) (033) 320-0870 loc1103/1104 * Telefax No.: (033) 320-0879

* Website: www.wvsu.edu.ph * Email Address: bac@wvsu.edu.ph



PHI-18939-2-QM

Section VII. Technical Specifications

Item No.	Specification	Statement of Compliance <i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i>
PROCUREMENT OF ANATOMICAL MODELS AND HOSPITAL BED		Statement of Compliance
	<ul style="list-style-type: none"> • Anterior and posterior auscultation sites • Bilateral needle decompression at second intercostal • Dynamic airway and Lung compliance/resistance <ul style="list-style-type: none"> ➢ Ten level of static compliance, 15-50 ml/cm H₂O ➢ Ten levels of airway resistance ➢ Hold PEEP from 5 to 20 cmH₂O ➢ Exhales real and measurable CO₂ ➢ Vary lung mechanics throughout the scenario ➢ Receive real-time feedback from real mechanical ventilators ➢ Capable of assisting the ventilator at variable respiratory rates ➢ Compliance and resistance can be varied while connected to the ventilator 	
	<p>Cardiac</p> <ul style="list-style-type: none"> • ECGs are generated in real time with physiologic variations never repeating textbook patterns • Heart sounds may be auscultated and are synchronized with ECG • eCPR sensors-chest compressions are measured and logged • 12-Lead ECG with integrated MI model 	
	<p>Circulation</p> <ul style="list-style-type: none"> • Measure blood pressure by palpation or auscultation • Use real BP cuff rather than a "virtual" cuff to measure blood pressure • Korotkoff sounds audible between systolic and diastolic pressures • Oxygen saturation detected using real monitors rather than a "virtual" value 	



Section VII. Technical Specifications

Item No.	Specification	Statement of Compliance
		<p style="text-align: center;">Statement of Compliance</p> <p><i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>
	PROCUREMENT OF ANATOMICAL MODELS AND HOSPITAL BED	Statement of Compliance
	<ul style="list-style-type: none"> • Pulse sites synchronized with BP and heart rate • Bilateral IV arms with fill/drain sites • SubQ and IM Injection sites • Intraosseous access at tibia • ECG Monitoring using real devices • Defibrillate, cardiovert, and pace using real devices • Multiple heart sounds, rates, and intensities • ECG rhythms are generated in real time • Bilateral carotid, radial, brachial, femoral, popliteal and pedal pulses synchronized with ECG • Pulses vary with blood pressure and are continuous and synchronized with the ECG even during a paced rhythm 	
	<p>Instructor or Automatic Mode</p> <ul style="list-style-type: none"> • Vital sign are generated in real time • Drug library with medications • Use of medications change conditions in real time, mimicking real clinical situations 	
	<p>Drug Recognition System</p> <ul style="list-style-type: none"> • Identifies drug type and volume injected into veins of the right hand and forearm • Supplied with 20 wireless tag syringes • Use drugs from the library or choose to model other drugs using software template 	
	<p>Neural Responses</p> <ul style="list-style-type: none"> • Eyes are controlled automatically by the physiologic model or directly by the instructor • Select pupillary response to light 	
	<p>Speech</p> <ul style="list-style-type: none"> • Wireless streaming audio • Create and store vocal response in any language 	



Section VII. Technical Specifications

Item No.	Specification	Statement of Compliance
	PROCUREMENT OF ANATOMICAL MODELS AND HOSPITAL BED	Statement of Compliance
	Vital Sign Monitor <ul style="list-style-type: none"> Controlled via wireless tablet PC Use selected configuration or create your own configuration to mimic the real monitors used in your facility Share images such as ultrasounds, CT scans, and lab result Touchscreen Control Monitor can be configured by the instructor to suit the scenario 	<p><i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>
	Other <ul style="list-style-type: none"> Central cyanosis Fill bladder and perform Foley catheterization Interchangeable genitalia Insert feeding tubes Auscultate bowel sounds 	
	User Interface <ul style="list-style-type: none"> Sensors track student actions Changes in condition and care provided are time stamped and logged Supplied with wireless tablet PC Includes pre-programmed scenarios which can be modified by the instructor even during the scenario Create your own scenarios (add/edit) Change the simulator's condition during the scenario 	
	Warranty: 1 year	
	Others:	
	<i>Bidders shall indicate in its submission a video presentation showing features of the model, including demonstration of the various required features indicated in the Section VII.</i>	
	<i>Bidders should include/ submit brochures indicating brand and model offered.</i>	



Section VII. Technical Specifications

Item No.	Specification	Statement of Compliance
PROCUREMENT OF ANATOMICAL MODELS AND HOSPITAL BED		Statement of Compliance
2.	Digital Anatomical Table	
	Qty: 2	
	Unit: unit	
	Brand (pls. indicate brand)	
	Model (pls. specify model)	
	Description: Digital Anatomical Table with VH Dissector Touch Software is designed for use on a large touch screen for collaborative environment.	
	Features: <ul style="list-style-type: none"> • Built on real anatomy from the National Library of Medicine's Visible Human Project Interact with 3D and cross-sectional view of over 2,000 anatomical structures. Interact through identification, dissection, assembly and rotation of the VH Dissector helps the student understand the complex three-dimensional structure of the human body. • Provide Step-by-Step Dissection Instructions • High-resolution Virtual "Bone Box" • Lab Practical Self-Assessment • Clinical Rotations • 75" Digital Multi-Touch Screen Display • Intel i7 processor Windows 10 Pro • Motorized Lift/Tilt System with Caster wheels 	
	Display Specification: <ul style="list-style-type: none"> • Android TV "75 Class (Diagonal) • RAM: 4GB • Storage: 32Gb • Resolution: 3840x 2160 (4K) • Android Ver.: Android 8.0.0 • 20 Touch Points Simultaneous 	
	Input Ports: <ul style="list-style-type: none"> • VGA/Dp • Dp • HDMI Tc • HDMI 1 • HDMI 2 • HDMI Tc • USB FL • USB • TF • WiFi Antenna • Bluetooth Antenna 	



Section VII. Technical Specifications

Item No.	Specification	Statement of Compliance
		<p align="center">Statement of Compliance</p> <p align="center"><i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>
	PROCUREMENT OF ANATOMICAL MODELS AND HOSPITAL BED	Statement of Compliance
	Input/output Ports Bottom Panel: <ul style="list-style-type: none"> • RJ45 OUT • RJ45 IN • HDMI OUT • COAX • RS232 • Earphone • AV IN • YPBPR • PC Audio • VGA 	
	Accessories: <ul style="list-style-type: none"> • IQ Share Lite Dongle • Remote for Android TV • USB Cable • USB Adaptor • Pointing Pen 2 pcs • Power Cord • Motorized and Tilttable • Lift Stand with Remote • Heavy Duty Caster • Wheels with Lock 	
	Front Ports: <ul style="list-style-type: none"> • HDMI 1 • Media USB • PC USB • Touch USB 	
	Built In PC:	
	OS	Window 10
	CP U	Intel Core i7
	RAM	8 GB
	HDD	1 TB
	Display	Built in Intel
	HD Graphics	4600
	PC Input/ Output Ports: <ul style="list-style-type: none"> • Mic Line • Speaker Out • HDMI • DP • VGA • USB x 6 port • RS232 • LAN Port • WiFi Antenna Option: 65" Display also available	
	Warranty: 1 year	



Section VII. Technical Specifications

Item No.	Specification	Statement of Compliance
		<p><i>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>
PROCUREMENT OF ANATOMICAL MODELS AND HOSPITAL BED		Statement of Compliance
	<i>Others:</i>	
	<i>Bidders shall indicate in its submission a video presentation showing features of the model, including demonstration of the various required features indicated in the Section VII.</i>	
	<i>Bidders should include/ submit brochures indicating brand and model offered.</i>	
3.	Hospital Bed	
	<i>Qty: 2</i>	
	<i>Unit: unit</i>	
	<i>Brand (pls. indicate brand)</i>	
	<i>Model (pls. specify model)</i>	
	<i>Product Size:</i> 2200*980*520/740mm	
	<i>Backrest Angle:</i> 0-75°+/-5°	
	<i>Legrest Angle:</i> 0-40°+/-5°	
	<i>Height adjustable:</i> 520-740mm +/-100mm	
	<i>Trendelenburg and reverse angle:</i> 0-12°+/-2°	
	<i>Structure:</i>	
	<i>Bed frame made of steel with powder coated</i>	
	<i>ABS Bed head/foot boards with bumpers</i>	
	<i>Bed platform metal with powder coated by mould</i>	
	<i>4*125mm wheels with central locking system based on luxury chassis</i>	
	<i>ABS foldable side rail with angle indicator</i>	
	<i>Heavy duty durable structure, with safety capacity 250kgs</i>	
	<i>Packing Size:</i> 2200*1500*550mm	
	<i>G.W:</i> 170kg	
	<i>Warranty:</i> 1 year	
	<i>Others:</i>	
	<i>Bidders shall indicate in its submission a video presentation showing features of the model, including demonstration of the various required features indicated in the Section VII.</i>	
	<i>Bidders should include/ submit brochures indicating brand and model offered.</i>	
	--Nothing Follows--	



West Visayas State University

(Formerly Iloilo Normal School)

Bids and Awards Committee Secretariat Office/ Procurement Section

Luna St., La Paz, Iloilo City 5000

Iloilo, Philippines

* Trunkline: (063) (033) 320-0870 loc1103/1104 * Telefax No.: (033) 320-0879

* Website: www.wvsu.edu.ph * Email Address: bac@wvsu.edu.ph



PHI-18939-2-QM

Price Schedule for Goods Offered from Abroad
[shall be submitted with the Bid if bidder is offering goods from Abroad]

For Goods Offered from Abroad

Name of Bidder _____ Project ID No. 20-10-14 Page 1 of 1

1	2	3	4	5	6	7	8	9
Item	Description	Country of origin	Quantity	Unit price CIF port of entry (specify port) or CIP named place (specify border point or place of destination)	Total CIF for CIP price per item (col. 4 x 5)	Unit Price Delivered Duty Unpaid (DDU)	Unit price Delivered Duty Paid (DDP)	Total Price delivered DDP (col 4 x 8)
1.	Adult Multipurpose Simulator		1					
2.	Digital Anatomical Table		2					
3.	Hospital Bed		2					

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____