



West Visayas State University

(Formerly Iloilo Normal School)

Bids and Awards Committee Secretariat Office/ Procurement Section

Luna St., La Paz, Iloilo City 5000

Iloilo, Philippines

* Trunkline: (063) (033) 320-0870 loc1103/1104 * Telefax No.: (033) 320-0879

* Website: www.wvsu.edu.ph * Email Address: bac@wvsu.edu.ph



SUPPLEMENTAL/BID BULLETIN

Project Reference No.	IB No. 20-12-35
Name of the Project	Procurement of Various Dental Equipment (Re-bidding)
Location of the Project	WVSU - Main Campus

ADDENDUM NO. 2020-18

December 16, 2020

This **Addendum No. 2020-18** is issued to modify the **Section I. Invitation to Bid, Section VI. Schedule of Requirements, Section VII. Technical Specifications, Price Schedule for Goods Offered from Within Philippines and Price Schedule for Goods Offered from Abroad.** This shall form an integral part of the Bidding Documents.

I. Section VII. Technical Specifications:

Item No.	Description	Correction/ Modification/ Instruction
1.	Stationery Dental Unit with Prosthodontics, with Heavy Duty Compressor	Stationery Dental Unit with Prosthodontics, with Heavy Duty Compressor
	<i>Qty/ Unit: 10 sets</i>	<i>Qty/ Unit: 10 sets</i>
	Brand: (pls. indicate brand)	Brand: (pls. indicate brand)
	Model: (pls. specify model)	Model: (pls. specify model)
	Technical Specifications:	Technical Specifications:
	Power Supply	220V with automatic voltage regulator
	Backrest movement	Movable from 0 to 90 degrees
	Arm rest	Movable & adjustable
	2-hole handpiece adaptor for low speed	2-hole handpiece adaptor for low speed
	2-hole handpiece adaptor for high speed	2-hole handpiece adaptor for high speed
	Foot control	Multi-function design
	2 Dental compressors:	2 Dental compressors:
	<ul style="list-style-type: none"> Silent type, oil-free compressor 5 Hp; 15 liters tank capacity and 3 pistons 220 V 	<ul style="list-style-type: none"> 5-7.5 Hp; at least 25 liters tank capacity and 3 pistons cylinders/unit 220 V With fitting for drainage
	Accessories:	Accessories:
	<ul style="list-style-type: none"> Dentist's stool small receptacle/tray adjustable arms flexible light holder Cuspidor with filter, drainage hose and attachment 	<ul style="list-style-type: none"> Dentist's stool - standard with stainless steel frame small receptacle/tray adjustable arms flexible light holder Cuspidor with filter, drainage hose and attachment
	• Receptacle Holder	• Receptacle Holder
	• Cup holder	• Cup holder
	• Saliva ejector	• Saliva ejector



West Visayas State University

(Formerly Iloilo Normal School)

Bids and Awards Committee Secretariat Office/ Procurement Section

Luna St., La Paz, Iloilo City 5000

Iloilo, Philippines

* Trunkline: (063) (033) 320-0870 loc1103/1104 * Telefax No.: (033) 320-0879

* Website: www.wvsu.edu.ph * Email Address: bac@wvsu.edu.ph



PHI-18939-2-QM

Item No.	Description	Correction/ Modification/ Instruction	
	OTHER REQUIREMENTS:	OTHER REQUIREMENTS:	
	Functional <ul style="list-style-type: none"> • A stationary dental unit is rudimentary in design. • It has a dental light to illuminate the oral cavity and a spittoon and cup holder for rinsing. A small tray with adjustable arm is attached where instruments are placed. It will be used for pre-clinical exercises. • There should be foot control for handpieces. 	Functional <ul style="list-style-type: none"> • A stationary dental unit is rudimentary in design. • It has a dental light to illuminate the oral cavity and a spittoon and cup holder for rinsing. A small tray with adjustable arm is attached where instruments are placed. It will be used for pre-clinical exercises. • There should be foot control for handpieces. 	
	OTHER REQUIREMENTS:	OTHER REQUIREMENTS:	
	Performance <ul style="list-style-type: none"> • In Prosthodontics, mouth impressions on live patient are a competency exercise. Thus, a basic and simple dental unit is needed. When making complete denture as a pre-clinical requirement, several procedures require patient appointment that needs basic dental set up. 	Performance <ul style="list-style-type: none"> • In Prosthodontics, mouth impressions on live patient are a competency exercise. Thus, a basic and simple dental unit is needed. When making complete denture as a pre-clinical requirement, several procedures require patient appointment that needs basic dental set up. 	
	Environmental Interface <ul style="list-style-type: none"> • The unit must be placed near a wash basin or sink. The disadvantage of this unit is that disinfection and waste disposal is not properly implemented. 	Environmental Interface <ul style="list-style-type: none"> • The unit must be placed near a wash basin or sink. The disadvantage of this unit is that disinfection and waste disposal is not properly implemented. 	
	Other Terms and Conditions: <ul style="list-style-type: none"> • Delivery 90 days • Warranty 3 years for parts & services • Preventive maintenance- quarterly • Free Training: Dentist: 1-2 days Tech 3 days • Supplier should include in the submission a certificate of dealership or exclusivity from manufacturer or dealer local or abroad of the item/s offered. 	Other Terms and Conditions: <ul style="list-style-type: none"> • Delivery 90 days • Warranty 3 years for parts & services • Preventive maintenance- quarterly • Free Training: Dentist: 1-2 days Tech 3 days • Supplier should include in the submission a certificate of dealership or exclusivity from manufacturer or dealer local or abroad of the item/s offered. 	
	Others:	Others:	
	Bidders should include/ submit brochures indicating brand and model offered.	Bidders should include/ submit brochures indicating brand and model offered.	
2.	Fully Motorized Dental Chair, with Full Accessories, with Heavy Duty Compressor	Fully Motorized Dental Chair, with Full Accessories, with Heavy Duty Compressor	
	Qty/ Unit: 20 sets	Qty/ Unit: 20 sets	
	Brand: (pls. indicate brand)	Brand: (pls. indicate brand)	
	Model: (pls. specify model)	Model: (pls. specify model)	
	Technical Specifications:	Technical Specifications:	
	Power Supply	220V with automatic voltage regulator	Power Supply
	Automatic movements	4 to 5 working positions	Automatic movements
	Headrest lock system	Bi-articulated	Headrest lock system
	Armrest	with lateral movements	Armrest
	Lighting	Sensor halogen wide illumination; light is sensor operated with high and low illumination	Lighting
			Bi-articulated-can be adjusted for adult & pediatric patients
			with lateral or vertical movements
			Sensor halogen wide illumination; light is sensor operated with high and low illumination



West Visayas State University

(Formerly Iloilo Normal School)

Bids and Awards Committee Secretariat Office/ Procurement Section

Luna St., La Paz, Iloilo City 5000

Iloilo, Philippines

* Trunkline: (063) (033) 320-0870 loc1103/1104 * Telefax No.: (033) 320-0879

* Website: www.wvsu.edu.ph * Email Address: bac@wvsu.edu.ph



PHI-18939-2-QM

Item No.	Description	Correction/ Modification/ Instruction
	2-hole handpiece adaptor for low speed	2-hole handpiece adaptor for low speed
	2-hole handpiece adaptor for high speed	2-hole handpiece adaptor for high speed
	Foot controls with multi-function feature for hand pieces, light activation, & seat & backrest movement, & other dental chair functions	Foot controls with multi-function feature for hand pieces, light activation, & seat & backrest movement, & other dental chair functions
	CHAIR:	CHAIR:
	Ambidextrous serves both right and left handed	Preferably Ambidextrous serves both right and left handed operator
	Fully motorized chair power	Fully motorized chair power
	Gear type motor	Gear type motor
	Fixed foot switch type for the functions of dental chair	Fixed foot switch type for the functions of dental chair
	Chair with seamless upholstery and plastic cover	Chair with seamless upholstery and plastic cover
	Fixed armrest	Fixed armrest
	Articulated Armrest	Articulated Armrest
	DELIVERY UNIT/ TRAY:	DELIVERY UNIT/ TRAY:
	Hanging Type with pneumatic arm movement air break switch for up and down adjustment	Hanging Type with pneumatic arm movement air break switch for up and down adjustment
	1 Three-way Syringe Terminal	1 Three-way Syringe Terminal
	1 Low Speed Terminal (2 Holes)	1 Low Speed Terminal (2 Holes)
	1 High Speed Terminal (2 Holes)	1 High Speed Terminal (2 Holes)
	1 X-Ray Viewer	1 X-Ray Viewer
	1 Accessory Tray	1 Accessory Tray with full adjustable vertical & horizontal movements
	1 Clean Water System Bottle	1 Clean Water System Bottle
	WATER UNIT/ CUSPIDOR:	WATER UNIT/ CUSPIDOR:
	Porcelain Bowl Switch Type	Porcelain Bowl Switch Type
	Cup Filler Switch Type	Cup Filler Switch Type
	1 Saliva Ejector	1 Saliva Ejector
	1 Hivac Suction	1 Hivac Suction
	Sensor water for cuspidor & cupfiller	Sensor water control for cuspidor & cupfiller that can be controlled by the operator from the accessory tray
	DENTAL LIGHT:	DENTAL LIGHT:
	12V Halogen Type	12V Halogen Type
	Activation through Foot Control switch of the Dental Chair	High & low illumination that can be controlled from the accessory tray
	STOOL:	STOOL:
	Standard stool upholstery with plastic cover	Standard stool upholstery with plastic cover; ; stainless frame with rollers
	ACCESSORIES:	ACCESSORIES:
	<ul style="list-style-type: none"> • Receptacle holder • Automatic activation of water flow to cuspidor & cup holder • 220-mm diameter round ceramic or porcelain cuspidor with attachments to drainage • 90 degrees movement of cuspidor 	<ul style="list-style-type: none"> • Receptacle holder • Automatic activation of water flow to cuspidor & cup holder • 220-mm diameter round ceramic or porcelain cuspidor with attachments to drainage • 90 degrees movement of cuspidor



West Visayas State University

(Formerly Iloilo Normal School)

Bids and Awards Committee Secretariat Office/ Procurement Section

Luna St., La Paz, Iloilo City 5000

Iloilo, Philippines

* Trunkline: (063) (033) 320-0870 loc1103/1104 * Telefax No.: (033) 320-0879

* Website: www.wvsu.edu.ph * Email Address: bac@wvsu.edu.ph



PHI-18939-2-QM

Item No.	Description	Correction/ Modification/ Instruction
	<p>1 Dental compressor:</p> <ul style="list-style-type: none"> • Silent type, oil-less compressor • 15 Hp with a tank capacity of 360 liters • 220 V 	<p>Dental compressor:</p> <ul style="list-style-type: none"> • 2 units 7.5 Hp with a tank capacity of at least 300 liters; 3 pistons cylinders/unit; 220 V Or • 3 units 5 hp; at least 25 liters tank capacity and 3 pistons cylinders/unit; 220 V • With fitting for drainage
	OTHER REQUIREMENTS:	OTHER REQUIREMENTS:
	<p>Design Comes in different colors; it is robust and easy to clean. It is ambidextrous with equal ease when used by a right or left-handed clinician or clinical instructor.</p>	<p>Design Comes in different colors; it is robust and easy to clean. It is ambidextrous with equal ease when used by a right or left-handed clinician or clinical instructor.</p>
	<p>Functional</p> <ul style="list-style-type: none"> • A motorized cushioned steel chair that will allow patients to sit comfortably and clinicians to work efficiently during clinical treatment. • The automatic movements, and sensor halogen lighting technology provides greater durability and visibility. 	<p>Functional</p> <ul style="list-style-type: none"> • A motorized cushioned steel chair that will allow patients to sit comfortably and clinicians to work efficiently during clinical treatment. • The automatic movements, and sensor halogen lighting technology provides greater durability and visibility.
	<p>Performance</p> <ul style="list-style-type: none"> • The chair can carry up to 200kg patient weight, seat him comfortably. • All clinicians are trained to use the chair in different positions to optimize accessibility and visibility of the working area (inside the mouth) while the patient is comfortably seated. 	<p>Performance</p> <ul style="list-style-type: none"> • The chair can carry up to 200kg patient weight, seat him comfortably. • All clinicians are trained to use the chair in different positions to optimize accessibility and visibility of the working area (inside the mouth) while the patient is comfortably seated.
	<p>Environmental Interface</p> <ul style="list-style-type: none"> • The chair is expected to be in full function with regular cleaning and oiling by a trained dental technician. • Its superior quality and function may be affected by inappropriate use of a disinfectant that can cause oxidation and alter its surface 	<p>Environmental Interface</p> <ul style="list-style-type: none"> • The chair is expected to be in full function with regular cleaning and oiling by a trained dental technician. • Its superior quality and function may be affected by inappropriate use of a disinfectant that can cause oxidation and alter its surface
	<p>Other Terms and Conditions</p> <ul style="list-style-type: none"> • Delivery 90 days • Warranty 3 years for parts & services • Preventive maintenance- quarterly • Free training: Dentist: 1-2 days Technician: 3 days <p>Supplier should include in the submission a certificate of dealership or exclusivity from manufacturer or dealer local or abroad of the item/s offered.</p>	<p>Other Terms and Conditions</p> <ul style="list-style-type: none"> • Delivery 90 days • Warranty 3 years for parts & services • Preventive maintenance- quarterly • Free training: Dentist: 1-2 days Technician: 3 days <p>Supplier should include in the submission a certificate of dealership or exclusivity from manufacturer or dealer local or abroad of the item/s offered.</p>
	Others:	Others:
	Bidders should include/ submit brochures indicating brand and model offered.	Bidders should include/ submit brochures indicating brand and model offered.



West Visayas State University

(Formerly Iloilo Normal School)

Bids and Awards Committee Secretariat Office/ Procurement Section

Luna St., La Paz, Iloilo City 5000

Iloilo, Philippines

* Trunkline: (063) (033) 320-0870 loc1103/1104 * Telefax No.: (033) 320-0879

* Website: www.wvsu.edu.ph * Email Address: bac@wvsu.edu.ph



PHI-18939-2-QM

Item No.	Description		Correction/ Modification/ Instruction	
	3. Dental X-Ray Machine, Periapical		3. Dental X-Ray Machine, Periapical	
	Qty/ Unit: 1 unit		Qty/ Unit: 1 unit	
	Brand: (pls. indicate brand)		Brand: (pls. indicate brand)	
	Model: (pls. specify model)		Model: (pls. specify model)	
	Technical Specifications:		Technical Specifications:	
	Features:	Can be adjusted to short and long cone techniques	Features:	Preferably digital with soft touch-controlled panel & remote controlled power button Can be adjusted to short and long cone techniques
	Head angle (turning)	360 level turning angle of head	Head angle (turning)	360 level turning angle of head
	Head angle (rounding)	At least 270 rounding angle of head	Head angle (rounding)	At least 270 rounding angle of head
	Exposure time:	0.2 to 4 sec.	Exposure time:	0.2 to 4 sec.
	Mobile scissor arm:	With 180 rotation, light & precise movements	Mobile scissor arm:	With 180 rotation, light & precise movements
	Arms:	Able to stand stable in the air, can spin & turn	Arms:	Able to stand stable in the air, can spin & turn
	Accessories:		Accessories:	
	<ul style="list-style-type: none"> With patient's stool With developing box & lead apron With automatic voltage regulator (AVR) With high voltage generator: 70kV Pantograph Extensor Timer Life span of at least 6 years 		<ul style="list-style-type: none"> With patient's stool, stainless steel frame With developing box & lead apron With automatic voltage regulator (AVR) With high voltage generator: 70kV Timer located at the control panel Life span of at least 6 years 	
	OTHER REQUIREMENTS:		OTHER REQUIREMENTS:	
	Design <ul style="list-style-type: none"> The machine is so designed to produce images that will aid in evaluation of detailed tooth parts and supporting structures, in order to come up with an accurate diagnosis after clinical examination and history taking. Easy to manipulate and stored accordingly. Light material with high degree of insulation 		Design <ul style="list-style-type: none"> The machine is so designed to produce images that will aid in evaluation of detailed tooth parts and supporting structures, in order to come up with an accurate diagnosis after clinical examination and history taking. Easy to manipulate and stored accordingly. Light material with high degree of insulation 	
	Functional <ul style="list-style-type: none"> The machine consists of the X-ray tube, which is made of the anode and cathode, the power supply and timer. It is used to produce and demonstrate detailed images of the tooth and supporting structures by using an x-ray film (of different sizes) that is developed and fixed. It teaches the student on the various bisecting and paralleling techniques and angulation in order to assess a particular area in the dentition and facial region 		Functional <ul style="list-style-type: none"> The machine consists of the X-ray tube, which is made of the anode and cathode, the power supply and timer. It is used to produce and demonstrate detailed images of the tooth and supporting structures by using an x-ray film (of different sizes) that is developed and fixed. It teaches the student on the various bisecting and paralleling techniques and angulation in order to assess a particular area in the dentition and facial region 	



West Visayas State University

(Formerly Iloilo Normal School)

Bids and Awards Committee Secretariat Office/ Procurement Section

Luna St., La Paz, Iloilo City 5000

Iloilo, Philippines

* Trunkline: (063) (033) 320-0870 loc1103/1104 * Telefax No.: (033) 320-0879

* Website: www.wvsu.edu.ph * Email Address: bac@wvsu.edu.ph



PHI-18939-2-QM

Item No.	Description	Correction/ Modification/ Instruction
	<p>Performance</p> <ul style="list-style-type: none"> • The machine is primarily used by preclinical students in Roentology to as an introduction to the use of dental imaging in dentistry. • It is used during initial evaluation in the following disciplines: Oral Surgery, Periodontology, Endodontics, Dental Implantology, Prosthodontics, Orthodontic-Pediatric). 	<p>Performance</p> <ul style="list-style-type: none"> • The machine is primarily used by preclinical students in Roentology to as an introduction to the use of dental imaging in dentistry. • It is used during initial evaluation in the following disciplines: Oral Surgery, Periodontology, Endodontics, Dental Implantology, Prosthodontics, Orthodontic-Pediatric).
	<p>Environmental Interface</p> <ul style="list-style-type: none"> • The machine emits hazardous radiation thus, it is enclosed in a room lined with lead or with 4 inches thick of solid walls. • Appropriate disinfectant is used to clean the machine. • Dose specifications must be followed to prevent unnecessary exposure of the bone marrow, thyroid, gonad etc. • Exposure and dose can be reduced by considering factors like patient selection, use of intraoral high-speed image film, intensifying screens and calibrating the focal spot-to-film distance, and use of leaded aprons and collars. • The machine walls and door are routinely checked by a trained dental technician for leakage 	<p>Environmental Interface</p> <ul style="list-style-type: none"> • The machine emits hazardous radiation thus, it is enclosed in a room lined with lead or with 4 inches thick of solid walls. • Appropriate disinfectant is used to clean the machine. • Dose specifications must be followed to prevent unnecessary exposure of the bone marrow, thyroid, gonad etc. • Exposure and dose can be reduced by considering factors like patient selection, use of intraoral high-speed image film, intensifying screens and calibrating the focal spot-to-film distance, and use of leaded aprons and collars. • The machine walls and door are routinely checked by a trained dental technician for leakage
	<p>Other Term and Conditions</p> <ul style="list-style-type: none"> • Delivery 90 days • Warranty 3 years for parts & services • Preventive maintenance- quarterly • Free training: Dentist: 1-2 days Technician: 3 days <p>Supplier should include in the submission a certificate of dealership or exclusivity from manufacturer or dealer local or abroad of the item/s offered.</p>	<p>Other Term and Conditions</p> <ul style="list-style-type: none"> • Delivery 90 days • Warranty 3 years for parts & services • Preventive maintenance- quarterly • Free training: Dentist: 1-2 days Technician: 3 days <p>Supplier should include in the submission a certificate of dealership or exclusivity from manufacturer or dealer local or abroad of the item/s offered.</p>
	<p>Others:</p>	<p>Others:</p>
	<p>Bidders should include/ submit brochures indicating brand and model offered.</p>	<p>Bidders should include/ submit brochures indicating brand and model offered.</p>



West Visayas State University

(Formerly Iloilo Normal School)

Bids and Awards Committee Secretariat Office/ Procurement Section

Luna St., La Paz, Iloilo City 5000

Iloilo, Philippines

* Trunkline: (063) (033) 320-0870 loc1103/1104 * Telefax No.: (033) 320-0879

* Website: www.wvsu.edu.ph * Email Address: bac@wvsu.edu.ph



PHI-18939-2-QM

Item No.	Description	Correction/ Modification/ Instruction
4.	Digital Dental Panoramic-Celaphometric-TMJ X-ray Machine	Digital Dental Panoramic- <i>Cephalometric</i> - TMJ X-ray Machine
	<i>Qty/ Unit: 1 unit</i>	<i>Qty/ Unit: 1 unit</i>
	Brand: (pls. indicate brand)	Brand: (pls. indicate brand)
	Model: (pls. specify model)	Model: (pls. specify model)
	Technical Specifications:	Technical Specifications:
	Power supply: 220 V Automatic voltage regulator included	Power supply: 220 V Automatic voltage regulator included
	Focal spot: 0.5 mm	Focal spot: 0.5 mm
	Gray scale: 14 bit	Gray scale: 14 bit
	Scan time: average 10.1 to 13.5 seconds for Pano and 12.9 seconds for Cephalometric	Scan time: average 10.1 to 13.5 seconds for Pano and 12.9 seconds for Cephalometric
	Sensor type: CMOS	Sensor type: CMOS
	Generator: Voltage: 50-90 kVp / Current: 4-10 mA	Generator: Voltage: 50-90 kVp / Current: 4-10 mA
	<ul style="list-style-type: none"> • With at least 2 built-in sensor for panoramic and cephalometric 	<ul style="list-style-type: none"> • With at least 2 built-in sensor for panoramic and cephalometric
	Inclusion:	Inclusion:
	<ul style="list-style-type: none"> • PC with specifications to run both 2D softwares 	<ul style="list-style-type: none"> • PC with specifications to run both 2D softwares
	<ul style="list-style-type: none"> • 22" LED Monitor and Printer 	<ul style="list-style-type: none"> • 22" LED Monitor and Printer
	<ul style="list-style-type: none"> • 3,000 watts automatic voltage regulator with time delay 	<ul style="list-style-type: none"> • 3,000 watts automatic voltage regulator with time delay
	<ul style="list-style-type: none"> • One (1) piece Lead Apron 	<ul style="list-style-type: none"> • Two (2) pieces Lead Apron
	<ul style="list-style-type: none"> • 40 pieces special photopapers 	<ul style="list-style-type: none"> • 100 pieces special photopapers
	<ul style="list-style-type: none"> • 5 pieces special acetate for CEPH tracing 	<ul style="list-style-type: none"> • 100 pieces special acetate for CEPH tracing
	<ul style="list-style-type: none"> • licensed software that Dentists can educate patients on why certain procedures are necessary and how these procedures will be done (e.g. Implants, Prosthodontics, Endodontics, etc.); free upgrades and no expirations 	<ul style="list-style-type: none"> • licensed software that Dentists can educate patients on why certain procedures are necessary and how these procedures will be done (e.g. Implants, Prosthodontics, Endodontics, etc.); free upgrades and no expirations
	<ul style="list-style-type: none"> • With free two (2) new softwares for animation.. 	<ul style="list-style-type: none"> • With free two (2) new softwares for animation..
	OTHER REQUIREMENTS:	OTHER REQUIREMENTS:
	Functional <ul style="list-style-type: none"> • It is used as an adjunct to clinical examination and history taking to come up with an accurate assessment and diagnosis (all disciplines in dentistry). • It can be used in patients for more effective and vivid consultation, with the function demonstrated before and after treatment. • Scans cephalometric and must be ideal for orthodontic diagnosis. • Can perform panoramic, cephalometric, TMJ radiograph on left or right or both, hand and palm radiograph 	Functional <ul style="list-style-type: none"> • It is used as an adjunct to clinical examination and history taking to come up with an accurate assessment and diagnosis (all disciplines in dentistry). • It can be used in patients for more effective and vivid consultation, with the function demonstrated before and after treatment. • Scans cephalometric and must be ideal for orthodontic diagnosis. • Can perform panoramic, cephalometric, TMJ radiograph on left or right or both, hand and palm radiograph
	Performance <ul style="list-style-type: none"> • It is a fully digital imaging system with different modes that captures the dentition, jaw, maxilla, temporomandibular joint and sinuses based on the diagnostic purpose and requirement. 	Performance <ul style="list-style-type: none"> • It is a fully digital imaging system with different modes that captures the dentition, jaw, maxilla, temporomandibular joint and sinuses based on the diagnostic purpose and requirement.



West Visayas State University

(Formerly Iloilo Normal School)

Bids and Awards Committee Secretariat Office/ Procurement Section

Luna St., La Paz, Iloilo City 5000

Iloilo, Philippines

* Trunkline: (063) (033) 320-0870 loc1103/1104 * Telefax No.: (033) 320-0879

* Website: www.wvsu.edu.ph * Email Address: bac@wvsu.edu.ph



PHI-18939-2-QM

Item No.	Description	Correction/ Modification/ Instruction
	<p>Digital imaging:</p> <ul style="list-style-type: none"> • panoramic, • cephalometric • TMJ • bitewing, • orthogonal exposures/view • can take hand and palm radiograph • Can take sinus radiograph both PA and lateral. <p>Stored images can be viewed or manipulated anywhere within the practice network.</p>	<p>Digital imaging:</p> <ul style="list-style-type: none"> • panoramic, • cephalometric • TMJ • bitewing, • orthogonal exposures/view • can take hand and palm radiograph • Can take sinus radiograph both PA and lateral. <p>Stored images can be viewed or manipulated anywhere within the practice network.</p>
	<p>Environmental Interface</p> <ul style="list-style-type: none"> • The machine emits hazardous radiation thus, it is enclosed in a room lined with lead or with 4 inches thick of solid walls. • Allows easy access for Pano/Ceph functions without dropping sensor having two dedicated sensors can also prolong the lifespan of the CMOS sensors. • The machine should be in full function with regular maintenance and calibration by a trained x-ray technician. It may react or malfunction with incorrect wiring, accidents like fall and may break 	<p>Environmental Interface</p> <ul style="list-style-type: none"> • The machine emits hazardous radiation thus, it is enclosed in a room lined with lead or with 4 inches thick of solid walls. • Allows easy access for Pano/Ceph functions without dropping sensor having two dedicated sensors can also prolong the lifespan of the CMOS sensors. • The machine should be in full function with regular maintenance and calibration by a trained x-ray technician. It may react or malfunction with incorrect wiring, accidents like fall and may break
	<p>Other Terms and Conditions</p> <ul style="list-style-type: none"> • Delivery 90 days • Warranty 3 years for parts & services • Preventive • Free training on basic use, trouble shooting, & software manipulation <p style="padding-left: 40px;">Dentist: 1-2 days Technician: 3 days</p> <p>Supplier should include in the submission a certificate of dealership or exclusivity from manufacturer or dealer local or abroad of the item/s offered.</p>	<p>Other Terms and Conditions</p> <ul style="list-style-type: none"> • Delivery 90 days • Warranty 3 years for parts & services • Preventive • Free training on basic use, trouble shooting, & software manipulation <p style="padding-left: 40px;">Dentist: 1-2 days Technician: 3 days</p> <p>Supplier should include in the submission a certificate of dealership or exclusivity from manufacturer or dealer local or abroad of the item/s offered.</p>
	Others:	Others:
	Bidders should include/ submit brochures indicating brand and model offered.	Bidders should include/ submit brochures indicating brand and model offered.
	--Nothing Follows--	--Nothing Follows--

II. Section I. Invitation to Bid, Section VI. Schedule of Requirements , Price Schedule for Goods Offered from Within Philippines and Price Schedule for Goods Offered from Abroad

Item No.	Description	Correction/ Modification/ Instruction
4.	Digital Dental Panoramic-Celaphometric-TMJ X-ray Machine	Digital Dental Panoramic-Cephalometric-TMJ X-ray Machine
	--Nothing Follows--	--Nothing Follows--

Attachment: as stated.

For guidance and information of all concerned.

JULIUS B. UNDAR

Chairperson, Bids and Awards Committee